

CALIBRATION VERIFICATION SUMMARY

Contract: NOBI03

Lab Code: CHEM Case No.: P5306 SAS No.: P5306 SDG NO.: P5306

GC Column: RTX-CLP ID: 0.32 (mm) Initi. Calib. Date(s): 12/23/2024 12/23/2024

Client Sample No.: CCAL05 Date Analyzed: 12/26/2024

Lab Sample No.: HSTDCCC750 Data File : PS028825.D Time Analyzed: 18:40

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
2,4,5-T	9.491	9.397	9.597	752.840	712.500	5.7
2,4,5-TP(Silvex)	9.199	9.106	9.306	763.510	712.500	7.2
2,4-D	8.326	8.232	8.432	724.320	705.000	2.7
2,4-DB	10.061	9.969	10.169	723.660	712.500	1.6
2,4-DCAA	7.206	7.110	7.310	765.790	750.000	2.1
Dalapon	2.620	2.520	2.720	720.310	682.500	5.5
DICAMBA	7.391	7.297	7.497	754.170	705.000	7.0
DICHLORPROP	8.096	8.002	8.202	744.200	705.000	5.6
Dinoseb	11.268	11.176	11.376	743.400	705.000	5.4

CALIBRATION VERIFICATION SUMMARY

Contract: NOBI03

Lab Code: CHEM Case No.: P5306 SAS No.: P5306 SDG NO.: P5306

GC Column: RTX-CLP2 ID: 0.32 (mm) Initi. Calib. Date(s): 12/23/2024 12/23/2024

Client Sample No.: CCAL05 Date Analyzed: 12/26/2024

Lab Sample No.: HSTDCCC750 Data File : PS028825.D Time Analyzed: 18:40

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
2,4,5-T	10.253	10.160	10.360	695.020	712.500	-2.5
2,4,5-TP(Silvex)	9.835	9.742	9.942	702.440	712.500	-1.4
2,4-D	8.934	8.840	9.040	684.480	705.000	-2.9
2,4-DB	10.818	10.726	10.926	700.080	712.500	-1.7
2,4-DCAA	7.695	7.600	7.800	723.350	750.000	-3.6
Dalapon	2.673	2.574	2.774	665.660	682.500	-2.5
DICAMBA	7.893	7.798	7.998	695.150	705.000	-1.4
DICHLORPROP	8.606	8.512	8.712	698.120	705.000	-1.0
Dinoseb	11.196	11.104	11.304	677.580	705.000	-3.9