

CALIBRATION VERIFICATION SUMMARY

Contract: NOBI03

Lab Code: CHEM Case No.: P5306 SAS No.: P5306 SDG NO.: P5306

GC Column: RTX-CLP ID: 0.32 (mm) Initi. Calib. Date(s): 12/23/2024 12/23/2024

Client Sample No.: CCAL06 Date Analyzed: 12/30/2024

Lab Sample No.: HSTDCCC750 Data File : PS028849.D Time Analyzed: 10:25

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
2,4,5-T	9.495	9.397	9.597	750.040	712.500	5.3
2,4,5-TP(Silvex)	9.204	9.106	9.306	757.550	712.500	6.3
2,4-D	8.330	8.232	8.432	734.230	705.000	4.1
2,4-DB	10.066	9.969	10.169	716.190	712.500	0.5
2,4-DCAA	7.210	7.110	7.310	785.200	750.000	4.7
Dalapon	2.622	2.520	2.720	700.190	682.500	2.6
DICAMBA	7.395	7.297	7.497	750.320	705.000	6.4
DICHLORPROP	8.100	8.002	8.202	734.990	705.000	4.3
Dinoseb	11.272	11.176	11.376	744.240	705.000	5.6

CALIBRATION VERIFICATION SUMMARY

Contract: NOBI03

Lab Code: CHEM Case No.: P5306 SAS No.: P5306 SDG NO.: P5306

GC Column: RTX-CLP2 ID: 0.32 (mm) Initi. Calib. Date(s): 12/23/2024 12/23/2024

Client Sample No.: CCAL06 Date Analyzed: 12/30/2024

Lab Sample No.: HSTDCCC750 Data File : PS028849.D Time Analyzed: 10:25

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
2,4,5-T	10.250	10.160	10.360	647.660	712.500	-9.1
2,4,5-TP(Silvex)	9.833	9.742	9.942	659.120	712.500	-7.5
2,4-D	8.932	8.840	9.040	632.820	705.000	-10.2
2,4-DB	10.816	10.726	10.926	651.870	712.500	-8.5
2,4-DCAA	7.694	7.600	7.800	675.500	750.000	-9.9
Dalapon	2.673	2.574	2.774	630.140	682.500	-7.7
DICAMBA	7.891	7.798	7.998	649.660	705.000	-7.8
DICHLORPROP	8.604	8.512	8.712	644.640	705.000	-8.6
Dinoseb	11.193	11.104	11.304	635.000	705.000	-9.9