

CALIBRATION VERIFICATION SUMMARY

Contract: NOBI03

Lab Code: CHEM Case No.: P5306 SAS No.: P5306 SDG NO.: P5306

GC Column: RTX-CLP ID: 0.32 (mm) Initi. Calib. Date(s): 12/23/2024 12/23/2024

Client Sample No.: CCAL07 Date Analyzed: 12/30/2024

Lab Sample No.: HSTDCCC750 Data File : PS028853.D Time Analyzed: 12:02

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
2,4,5-T	9.494	9.397	9.597	746.270	712.500	4.7
2,4,5-TP(Silvex)	9.203	9.106	9.306	753.330	712.500	5.7
2,4-D	8.329	8.232	8.432	732.860	705.000	4.0
2,4-DB	10.065	9.969	10.169	713.990	712.500	0.2
2,4-DCAA	7.209	7.110	7.310	778.690	750.000	3.8
Dalapon	2.621	2.520	2.720	691.330	682.500	1.3
DICAMBA	7.395	7.297	7.497	746.480	705.000	5.9
DICHLORPROP	8.099	8.002	8.202	734.360	705.000	4.2
Dinoseb	11.271	11.176	11.376	739.130	705.000	4.8

CALIBRATION VERIFICATION SUMMARY

Contract: NOBI03

Lab Code: CHEM Case No.: P5306 SAS No.: P5306 SDG NO.: P5306

GC Column: RTX-CLP2 ID: 0.32 (mm) Initi. Calib. Date(s): 12/23/2024 12/23/2024

Client Sample No.: CCAL07 Date Analyzed: 12/30/2024

Lab Sample No.: HSTDCCC750 Data File : PS028853.D Time Analyzed: 12:02

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
2,4,5-T	10.251	10.160	10.360	652.470	712.500	-8.4
2,4,5-TP(Silvex)	9.833	9.742	9.942	662.660	712.500	-7.0
2,4-D	8.932	8.840	9.040	630.920	705.000	-10.5
2,4-DB	10.816	10.726	10.926	653.250	712.500	-8.3
2,4-DCAA	7.694	7.600	7.800	679.980	750.000	-9.3
Dalapon	2.675	2.574	2.774	626.290	682.500	-8.2
DICAMBA	7.891	7.798	7.998	653.140	705.000	-7.4
DICHLORPROP	8.604	8.512	8.712	653.060	705.000	-7.4
Dinoseb	11.194	11.104	11.304	636.520	705.000	-9.7