

Surrogate Summary

SDG No.: P5306

Client: Nobis Group

Analytical Method: 8151A

Lab Sample ID	Client ID	Parameter	Column	Spike	Result	Rec	Qual	Limits	
								Low	High
I.BLK-PS028786.D	PIBLK-PS028786.D	2,4-DCAA	1	500	469	94		32	138
		2,4-DCAA	2	500	482	96		32	138
I.BLK-PS028793.D	PIBLK-PS028793.D	2,4-DCAA	1	500	510	102		32	138
		2,4-DCAA	2	500	488	98		32	138
PB165776BL	PB165776BL	2,4-DCAA	1	500	558	112		27	122
		2,4-DCAA	2	500	511	102		27	122
PB165776BS	PB165776BS	2,4-DCAA	1	500	547	109		27	122
		2,4-DCAA	2	500	513	103		27	122
I.BLK-PS028799.D	PIBLK-PS028799.D	2,4-DCAA	1	500	525	105		32	138
		2,4-DCAA	2	500	494	99		32	138
P5306-01	OU4-VSL-07-121224	2,4-DCAA	1	500	229	46		27	122
		2,4-DCAA	2	500	196	39		27	122
P5306-03	OU4-VSL-08-121224	2,4-DCAA	1	500	396	79		27	122
		2,4-DCAA	2	500	347	69		27	122
P5306-05	OU4-VSL-09-121224	2,4-DCAA	1	500	351	70		27	122
		2,4-DCAA	2	500	313	63		27	122
P5306-07	OU4-VSL-10-121224	2,4-DCAA	1	500	311	62		27	122
		2,4-DCAA	2	500	274	55		27	122
P5306-09	OU4-VSL-11-121224	2,4-DCAA	1	500	394	79		27	122
		2,4-DCAA	2	500	342	68		27	122
P5306-11	OU4-VSL-12-121224	2,4-DCAA	1	500	373	75		27	122
		2,4-DCAA	2	500	335	67		27	122
I.BLK-PS028810.D	PIBLK-PS028810.D	2,4-DCAA	1	500	536	107		32	138
		2,4-DCAA	2	500	500	100		32	138
I.BLK-PS028819.D	PIBLK-PS028819.D	2,4-DCAA	1	500	566	113		32	138
		2,4-DCAA	2	500	504	101		32	138
P5306-13	OU4-VSL-13-121224	2,4-DCAA	1	500	295	59		27	122
		2,4-DCAA	2	500	245	49		27	122
P5306-15MS	OU4-VSL-14-121224MS	2,4-DCAA	1	500	506	101		27	122
		2,4-DCAA	2	500	414	83		27	122
P5306-15MSD	OU4-VSL-14-121224MSD	2,4-DCAA	1	500	507	101		27	122
		2,4-DCAA	2	500	413	83		27	122
I.BLK-PS028824.D	PIBLK-PS028824.D	2,4-DCAA	1	500	559	112		32	138
		2,4-DCAA	2	500	504	101		32	138
I.BLK-PS028848.D	PIBLK-PS028848.D	2,4-DCAA	1	500	599	120		32	138
		2,4-DCAA	2	500	512	102		32	138
P5306-15	OU4-VSL-14-121224	2,4-DCAA	1	500	512	102		27	122
		2,4-DCAA	2	500	398	80		27	122
I.BLK-PS028852.D	PIBLK-PS028852.D	2,4-DCAA	1	500	602	120		32	138
		2,4-DCAA	2	500	515	103		32	138