

CALIBRATION VERIFICATION SUMMARY

Contract: NOBI03

Lab Code: CHEM Case No.: P5306 SAS No.: P5306 SDG NO.: P5306

Continuing Calib Date: 12/18/2024 Initial Calibration Date(s): 12/06/2024 12/06/2024

Continuing Calib Time: 20:22 Initial Calibration Time(s): 14:19 22:34

GC Column: ZB-MR1 ID: 0.32 (mm)

COMPOUND	CCAL RT	AVG RT	RT WINDOW		DIFF RT
			FROM	TO	
Aroclor-1242-1 (1)	4.80	4.81	4.71	4.91	0.01
Aroclor-1242-2 (2)	4.82	4.83	4.73	4.93	0.01
Aroclor-1242-3 (3)	4.88	4.88	4.78	4.98	0.00
Aroclor-1242-4 (4)	5.00	5.01	4.91	5.11	0.01
Aroclor-1242-5 (5)	5.65	5.66	5.56	5.76	0.01
Tetrachloro-m-xylene	3.71	3.71	3.61	3.81	0.00
Decachlorobiphenyl	8.78	8.79	8.69	8.89	0.02

CALIBRATION VERIFICATION SUMMARY

Contract: NOBI03

Lab Code: CHEM Case No.: P5306 SAS No.: P5306 SDG NO.: P5306

Continuing Calib Date: 12/18/2024 Initial Calibration Date(s): 12/06/2024 12/06/2024

Continuing Calib Time: 20:22 Initial Calibration Time(s): 14:19 22:34

GC Column: ZB-MR2 ID: 0.32 (mm)

COMPOUND	CCAL RT	AVG RT	RT WINDOW		DIFF RT
			FROM	TO	
Aroclor-1242-1 (1)	4.79	4.80	4.70	4.90	0.01
Aroclor-1242-2 (2)	4.81	4.82	4.72	4.92	0.01
Aroclor-1242-3 (3)	4.99	4.99	4.89	5.09	0.00
Aroclor-1242-4 (4)	5.07	5.08	4.98	5.18	0.01
Aroclor-1242-5 (5)	5.59	5.60	5.50	5.70	0.01
Tetrachloro-m-xylene	3.70	3.71	3.61	3.81	0.01
Decachlorobiphenyl	8.73	8.74	8.64	8.84	0.01