

CALIBRATION VERIFICATION SUMMARY

Contract: NOBI03

Lab Code: CHEM Case No.: P5306 SAS No.: P5306 SDG NO.: P5306

GC Column: ZB-MR1 ID: 0.32 (mm) Initi. Calib. Date(s): 12/06/2024 12/06/2024

Client Sample No.: CCAL01 Date Analyzed: 12/18/2024

Lab Sample No.: AR1660CCC500 Data File : PO108620.D Time Analyzed: 14:46

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	4.802	4.709	4.909	474.580	500.000	-5.1
Aroclor-1016-2	4.822	4.729	4.929	475.440	500.000	-4.9
Aroclor-1016-3	4.878	4.785	4.985	475.580	500.000	-4.9
Aroclor-1016-4	4.999	4.907	5.107	474.750	500.000	-5.1
Aroclor-1016-5	5.257	5.165	5.365	473.130	500.000	-5.4
Aroclor-1260-1	6.300	6.210	6.410	455.930	500.000	-8.8
Aroclor-1260-2	6.489	6.398	6.598	462.940	500.000	-7.4
Aroclor-1260-3	6.859	6.769	6.969	466.590	500.000	-6.7
Aroclor-1260-4	7.119	7.029	7.229	461.520	500.000	-7.7
Aroclor-1260-5	7.360	7.270	7.470	462.190	500.000	-7.6
Decachlorobiphenyl	8.775	8.691	8.891	44.110	50.000	-11.8
Tetrachloro-m-xylene	3.706	3.610	3.810	48.140	50.000	-3.7

CALIBRATION VERIFICATION SUMMARY

Contract: NOBI03

Lab Code: CHEM Case No.: P5306 SAS No.: P5306 SDG NO.: P5306

GC Column: ZB-MR2 ID: 0.32 (mm) Initi. Calib. Date(s): 12/06/2024 12/06/2024

Client Sample No.: CCAL01 Date Analyzed: 12/18/2024

Lab Sample No.: AR1660CCC500 Data File : PO108620.D Time Analyzed: 14:46

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
Aroclor-1016-1	4.790	4.698	4.898	530.530	500.000	6.1
Aroclor-1016-2	4.810	4.718	4.918	527.700	500.000	5.5
Aroclor-1016-3	4.986	4.894	5.094	521.780	500.000	4.4
Aroclor-1016-4	5.027	4.935	5.135	514.400	500.000	2.9
Aroclor-1016-5	5.241	5.150	5.350	524.590	500.000	4.9
Aroclor-1260-1	6.276	6.186	6.386	524.770	500.000	5.0
Aroclor-1260-2	6.462	6.373	6.573	525.150	500.000	5.0
Aroclor-1260-3	6.616	6.527	6.727	523.700	500.000	4.7
Aroclor-1260-4	7.089	7.000	7.200	541.420	500.000	8.3
Aroclor-1260-5	7.328	7.239	7.439	546.360	500.000	9.3
Decachlorobiphenyl	8.724	8.641	8.841	50.970	50.000	1.9
Tetrachloro-m-xylene	3.703	3.608	3.808	52.590	50.000	5.2