

INITIAL CALIBRATION OF MULTICOMPONENT ANALYTES

Contract: NOBI03

Lab Code: CHEM **Case No.:** P5306 **SAS No.:** P5306 **SDG NO.:** P5306

Instrument ID: ECD_L **Date(s) Analyzed:** 11/25/2024 11/25/2024

GC Column: ZB-MR2 **ID:** 0.32 (mm)

COMPOUND	AMOUNT (ng)	PEAK	RT	RT WINDOW		CALIBRATION FACTOR
				FROM	TO	
Toxaphene	500	1	6.23	6.13	6.33	24617300
		2	6.44	6.34	6.54	14127800
		3	7.06	6.96	7.16	77602600
		4	7.15	7.05	7.25	57658100
		5	7.93	7.83	8.03	42510000

INITIAL CALIBRATION OF MULTICOMPONENT ANALYTES

Contract: NOBI03

Lab Code: CHEM **Case No.:** P5306 **SAS No.:** P5306 **SDG NO.:** P5306

Instrument ID: ECD_L **Date(s) Analyzed:** 11/25/2024 11/25/2024

GC Column: ZB-MR1 **ID:** 0.32 (mm)

COMPOUND	AMOUNT (ng)	PEAK	RT	RT WINDOW		CALIBRATION FACTOR
				FROM	TO	
Toxaphene	500	1	5.00	4.90	5.10	21944300
		2	5.33	5.23	5.43	22442000
		3	5.69	5.59	5.79	24171900
		4	6.60	6.50	6.70	75449400
		5	7.04	6.94	7.14	73811300