

CALIBRATION VERIFICATION SUMMARY

Contract: NOBI03

Lab Code: CHEM Case No.: P5306 SAS No.: P5306 SDG NO.: P5306

Continuing Calib Date: 12/18/2024 Initial Calibration Date(s): 11/25/2024 11/25/2024

Continuing Calib Time: 18:18 Initial Calibration Time(s): 11:32 12:25

GC Column: ZB-MR1 ID: 0.32 (mm)

COMPOUND	CCAL RT	AVG RT	RT WINDOW		DIFF RT
			FROM	TO	
Decachlorobiphenyl	9.06	9.05	8.95	9.15	-0.01
Tetrachloro-m-xylene	3.54	3.54	3.44	3.64	0.00
alpha-BHC	4.00	3.99	3.89	4.09	0.00
beta-BHC	4.53	4.52	4.42	4.62	-0.01
delta-BHC	4.77	4.77	4.67	4.87	0.00
gamma-BHC (Lindane)	4.33	4.33	4.23	4.43	0.00
Heptachlor	4.92	4.91	4.81	5.01	-0.01
Aldrin	5.26	5.26	5.16	5.36	0.00
Heptachlor epoxide	5.68	5.68	5.58	5.78	0.00
Endosulfan I	6.07	6.07	5.97	6.17	0.00
Dieldrin	6.35	6.34	6.24	6.44	-0.01
4,4'-DDE	6.19	6.19	6.09	6.29	0.00
Endrin	6.57	6.57	6.47	6.67	0.00
Endosulfan II	6.80	6.79	6.69	6.89	-0.01
4,4'-DDD	6.71	6.71	6.61	6.81	0.00
Endosulfan sulfate	7.16	7.16	7.06	7.26	0.00
4,4'-DDT	7.03	7.02	6.92	7.12	-0.01
Methoxychlor	7.50	7.50	7.40	7.60	0.00
Endrin ketone	7.65	7.64	7.54	7.74	0.00
Endrin aldehyde	6.93	6.92	6.82	7.02	-0.01
alpha-Chlordane	6.02	6.02	5.92	6.12	0.00
gamma-Chlordane	5.94	5.94	5.84	6.04	0.00

CALIBRATION VERIFICATION SUMMARY

Contract: NOBI03

Lab Code: CHEM Case No.: P5306 SAS No.: P5306 SDG NO.: P5306

Continuing Calib Date: 12/18/2024 Initial Calibration Date(s): 11/25/2024 11/25/2024

Continuing Calib Time: 18:18 Initial Calibration Time(s): 11:32 12:25

GC Column: ZB-MR2 ID: 0.32 (mm)

COMPOUND	CCAL RT	AVG RT	RT WINDOW		DIFF RT
			FROM	TO	
Decachlorobiphenyl	7.91	7.91	7.81	8.01	0.00
Tetrachloro-m-xylene	2.78	2.77	2.67	2.87	-0.01
alpha-BHC	3.28	3.28	3.18	3.38	0.00
beta-BHC	3.91	3.91	3.81	4.01	0.00
delta-BHC	4.14	4.14	4.04	4.24	0.00
gamma-BHC (Lindane)	3.61	3.61	3.51	3.71	0.00
Heptachlor	3.95	3.95	3.85	4.05	0.00
Aldrin	4.23	4.23	4.13	4.33	0.00
Heptachlor epoxide	4.73	4.73	4.63	4.83	0.00
Endosulfan I	5.10	5.10	5.00	5.20	0.00
Dieldrin	5.36	5.36	5.26	5.46	0.00
4,4'-DDE	5.23	5.23	5.13	5.33	0.00
Endrin	5.64	5.64	5.54	5.74	0.00
Endosulfan II	5.94	5.93	5.83	6.03	0.00
4,4'-DDD	5.79	5.79	5.69	5.89	0.00
Endosulfan sulfate	6.34	6.33	6.23	6.43	-0.01
4,4'-DDT	6.04	6.04	5.94	6.14	0.00
Methoxychlor	6.61	6.61	6.51	6.71	0.00
Endrin ketone	6.84	6.84	6.74	6.94	0.00
Endrin aldehyde	6.11	6.11	6.01	6.21	0.00
alpha-Chlordane	5.04	5.04	4.94	5.14	0.00
gamma-Chlordane	4.98	4.98	4.88	5.08	0.00