

Instrument ID: ECD_L

Daily Analysis Runlog For Sequence/QC Batch ID # PL112524

Review By	yogesh	Review On	11/26/2024 8:18:30 AM
Supervise By	Ankita	Supervise On	11/26/2024 4:22:56 PM
SubDirectory	PL112524	HP Acquire Method	HP Processing Method PL112524

STD. NAME	STD REF.#
Tune/Reschk	PP23793,PP23517
Initial Calibration Stds	PP23673,PP23674,PP23675,PP23676,PP23677,PP23678,PP23679,PP23680,PP23681,PP23682,PP23683
CCC	PP23686,PP23690,PP23695
Internal Standard/PEM	
ICV/I.BLK	PP23687,PP23693,PP23698
Surrogate Standard	
MS/MSD Standard	
LCS Standard	

Sr#	SampleId	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	HEXANE	HEXANE	PL093229.D	25 Nov 2024 10:39		AR\AJ	Ok
2	I.BLK	I.BLK	PL093230.D	25 Nov 2024 10:52		AR\AJ	Ok
3	PEM	PEM	PL093231.D	25 Nov 2024 11:05		AR\AJ	Ok,M
4	RESCHK	RESCHK	PL093232.D	25 Nov 2024 11:18		AR\AJ	Ok
5	PSTDICC100	PSTDICC100	PL093233.D	25 Nov 2024 11:32		AR\AJ	Ok
6	PSTDICC075	PSTDICC075	PL093234.D	25 Nov 2024 11:45		AR\AJ	Ok
7	PSTDICC050	PSTDICC050	PL093235.D	25 Nov 2024 11:58		AR\AJ	Ok
8	PSTDICC025	PSTDICC025	PL093236.D	25 Nov 2024 12:11		AR\AJ	Ok
9	PSTDICC005	PSTDICC005	PL093237.D	25 Nov 2024 12:25		AR\AJ	Ok,M
10	PCHLORICC1000	PCHLORICC1000	PL093238.D	25 Nov 2024 12:38		AR\AJ	Ok
11	PCHLORICC750	PCHLORICC750	PL093239.D	25 Nov 2024 12:51		AR\AJ	Ok
12	PCHLORICC500	PCHLORICC500	PL093240.D	25 Nov 2024 13:04		AR\AJ	Ok
13	PCHLORICC250	PCHLORICC250	PL093241.D	25 Nov 2024 13:18		AR\AJ	Ok
14	PCHLORICC050	PCHLORICC050	PL093242.D	25 Nov 2024 13:31		AR\AJ	Ok,M
15	PTOXICC1000	PTOXICC1000	PL093243.D	25 Nov 2024 13:44		AR\AJ	Ok
16	PTOXICC750	PTOXICC750	PL093244.D	25 Nov 2024 13:57		AR\AJ	Ok
17	PTOXICC500	PTOXICC500	PL093245.D	25 Nov 2024 14:11		AR\AJ	Ok
18	PTOXICC250	PTOXICC250	PL093246.D	25 Nov 2024 14:24		AR\AJ	Ok

Instrument ID: ECD_L

Daily Analysis Runlog For Sequence/QC Batch ID # PL112524

Review By	yogesh	Review On	11/26/2024 8:18:30 AM
Supervise By	Ankita	Supervise On	11/26/2024 4:22:56 PM
SubDirectory	PL112524	HP Acquire Method	HP Processing Method PL112524
STD. NAME	STD REF.#		
Tune/Reschk	PP23793,PP23517		
Initial Calibration Stds	PP23673,PP23674,PP23675,PP23676,PP23677,PP23678,PP23679,PP23680,PP23681,PP23682,PP23683		
CCC	PP23686,PP23690,PP23695		
Internal Standard/PEM			
ICV/I.BLK	PP23687,PP23693,PP23698		
Surrogate Standard			
MS/MSD Standard			
LCS Standard			

19	PTOXICC100	PTOXICC100	PL093247.D	25 Nov 2024 14:37		AR\AJ	Ok,M
20	PSTDICV050	ICVPL112524	PL093248.D	25 Nov 2024 14:50		AR\AJ	Ok
21	PCHLORICV500	ICVPL112524	PL093249.D	25 Nov 2024 15:03		AR\AJ	Ok,M
22	PTOXICV500	ICVPL112524	PL093250.D	25 Nov 2024 15:16		AR\AJ	Ok
23	I.BLK	I.BLK	PL093251.D	25 Nov 2024 15:30		AR\AJ	Ok
24	PEM	PEM	PL093252.D	25 Nov 2024 15:43		AR\AJ	Ok
25	PSTDCCC050	PSTDCCC050	PL093253.D	25 Nov 2024 15:56		AR\AJ	Ok
26	PB165184BS	PB165184BS	PL093254.D	25 Nov 2024 16:22		AR\AJ	Ok,M
27	PB165202BL	PB165202BL	PL093255.D	25 Nov 2024 16:37		AR\AJ	Ok
28	PB165202BS	PB165202BS	PL093256.D	25 Nov 2024 16:50	Recovery high for methoxychlor	AR\AJ	Ok,M
29	PB165202BSD	PB165202BSD	PL093257.D	25 Nov 2024 17:03	Recovery high for methoxychlor	AR\AJ	Ok,M
30	P4947-01	A3988	PL093258.D	25 Nov 2024 17:16		AR\AJ	Ok
31	I.BLK	I.BLK	PL093259.D	25 Nov 2024 17:29		AR\AJ	Ok
32	PSTDCCC050	PSTDCCC050	PL093260.D	25 Nov 2024 17:42		AR\AJ	Ok

M : Manual Integration

Instrument ID: ECD_L

Daily Analysis Runlog For Sequence/QC Batch ID # PL121824

Review By	Abdul	Review On	12/19/2024 9:44:00 AM
Supervise By	Ankita	Supervise On	12/19/2024 10:29:06 AM
SubDirectory	PL121824	HP Acquire Method	HP Processing Method pl112524 8081

STD. NAME	STD REF.#
Tune/Reschk	PP23793,PP23517
Initial Calibration Stds	PP23673,PP23674,PP23675,PP23676,PP23677,PP23678,PP23679,PP23680,PP23681,PP23682,PP23683
CCC	PP23686,PP23690,PP23695
Internal Standard/PEM	
ICV/I.BLK	PP23687,PP23693,PP23698
Surrogate Standard	
MS/MSD Standard	
LCS Standard	

Sr#	SampleId	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	HEXANE	HEXANE	PL093406.D	18 Dec 2024 10:22		AR\AJ	Ok
2	I.BLK	I.BLK	PL093407.D	18 Dec 2024 10:35		AR\AJ	Ok
3	PEM	PEM	PL093408.D	18 Dec 2024 10:48		AR\AJ	Ok,M
4	PSTDCCC050	PSTDCCC050	PL093409.D	18 Dec 2024 11:02		AR\AJ	Ok,M
5	PP24080	PP24080	PL093410.D	18 Dec 2024 11:37		AR\AJ	Ok,M
6	PP24081	PP24081	PL093411.D	18 Dec 2024 12:28		AR\AJ	Ok,M
7	PP24088	PP24088	PL093412.D	18 Dec 2024 13:03		AR\AJ	Ok,M
8	PP24091	PP24091	PL093413.D	18 Dec 2024 13:17		AR\AJ	Ok
9	I.BLK	I.BLK	PL093414.D	18 Dec 2024 14:10		AR\AJ	Ok
10	PSTDCCC050	PSTDCCC050	PL093415.D	18 Dec 2024 14:24		AR\AJ	Ok,M
11	PB165704BL	PB165704BL	PL093416.D	18 Dec 2024 15:49		AR\AJ	Ok
12	PB165704BS	PB165704BS	PL093417.D	18 Dec 2024 16:03		AR\AJ	Ok
13	P5312-01	SOIL-VNJ-222	PL093418.D	18 Dec 2024 16:18		AR\AJ	Ok,M
14	P5312-03	CONCRETE-VNJ-222	PL093419.D	18 Dec 2024 16:32		AR\AJ	Ok,M
15	P5316-01	TT-304-IDWSO-202412	PL093420.D	18 Dec 2024 16:45		AR\AJ	Ok,M
16	P5306-01	OU4-VSL-07-121224	PL093421.D	18 Dec 2024 16:58		AR\AJ	Ok,M
17	P5306-01MS	OU4-VSL-07-121224M	PL093422.D	18 Dec 2024 17:11		AR\AJ	Ok,M
18	P5306-01MSD	OU4-VSL-07-121224M	PL093423.D	18 Dec 2024 17:24		AR\AJ	Ok,M

Instrument ID: ECD_L

Daily Analysis Runlog For Sequence/QC Batch ID # PL121824

Review By	Abdul	Review On	12/19/2024 9:44:00 AM
Supervise By	Ankita	Supervise On	12/19/2024 10:29:06 AM
SubDirectory	PL121824	HP Acquire Method	HP Processing Method pl112524 8081
STD. NAME	STD REF.#		
Tune/Reschk	PP23793,PP23517		
Initial Calibration Stds	PP23673,PP23674,PP23675,PP23676,PP23677,PP23678,PP23679,PP23680,PP23681,PP23682,PP23683		
CCC	PP23686,PP23690,PP23695		
Internal Standard/PEM			
ICV/I.BLK	PP23687,PP23693,PP23698		
Surrogate Standard			
MS/MSD Standard			
LCS Standard			

19	P5306-03	OU4-VSL-08-121224	PL093424.D	18 Dec 2024 17:38		AR\AJ	Ok,M
20	P5306-05	OU4-VSL-09-121224	PL093425.D	18 Dec 2024 17:51		AR\AJ	Ok,M
21	I.BLK	I.BLK	PL093426.D	18 Dec 2024 18:04		AR\AJ	Ok
22	PSTDCCC050	PSTDCCC050	PL093427.D	18 Dec 2024 18:18		AR\AJ	Ok,M

M : Manual Integration

Instrument ID: ECD_L

Daily Analysis Runlog For Sequence/QC Batch ID # PL121924

Review By	Abdul	Review On	12/20/2024 9:20:52 AM
Supervise By	Ankita	Supervise On	12/20/2024 10:44:33 AM
SubDirectory	PL121924	HP Acquire Method	HP Processing Method pl112524 8081

STD. NAME	STD REF.#
Tune/Reschk	PP23793,PP23517
Initial Calibration Stds	PP23673,PP23674,PP23675,PP23676,PP23677,PP23678,PP23679,PP23680,PP23681,PP23682,PP23683
CCC	PP23686,PP23690,PP23695
Internal Standard/PEM	
ICV/I.BLK	PP23687,PP23693,PP23698
Surrogate Standard	
MS/MSD Standard	
LCS Standard	

Sr#	SampleId	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	HEXANE	HEXANE	PL093428.D	19 Dec 2024 12:12		AR\AJ	Ok
2	I.BLK	I.BLK	PL093429.D	19 Dec 2024 12:26		AR\AJ	Ok
3	PEM	PEM	PL093430.D	19 Dec 2024 12:39		AR\AJ	Ok,M
4	PSTDCCC050	PSTDCCC050	PL093431.D	19 Dec 2024 12:52		AR\AJ	Ok
5	PCHLORCCC500	PCHLORCCC500	PL093432.D	19 Dec 2024 13:55		AR\AJ	Ok,M
6	PTOXCCC500	PTOXCCC500	PL093433.D	19 Dec 2024 14:14		AR\AJ	Ok,M
7	PB165704BS	PB165704BS	PL093434.D	19 Dec 2024 14:55		AR\AJ	Ok,M
8	PB165704BS	PB165704BS	PL093435.D	19 Dec 2024 15:27		AR\AJ	Ok,M
9	P5306-07	OU4-VSL-10-121224	PL093436.D	19 Dec 2024 15:47		AR\AJ	Ok,M
10	P5306-09	OU4-VSL-11-121224	PL093437.D	19 Dec 2024 16:00		AR\AJ	Ok,M
11	P5306-11	OU4-VSL-12-121224	PL093438.D	19 Dec 2024 16:14		AR\AJ	Ok
12	P5306-13	OU4-VSL-13-121224	PL093439.D	19 Dec 2024 16:27		AR\AJ	Ok,M
13	P5306-15	OU4-VSL-14-121224	PL093440.D	19 Dec 2024 16:40		AR\AJ	Ok,M
14	I.BLK	I.BLK	PL093441.D	19 Dec 2024 16:54		AR\AJ	Ok
15	PSTDCCC050	PSTDCCC050	PL093442.D	19 Dec 2024 17:07		AR\AJ	Ok
16	PCHLORCCC500	PCHLORCCC500	PL093443.D	19 Dec 2024 17:21		AR\AJ	Ok,M
17	PTOXCCC500	PTOXCCC500	PL093444.D	19 Dec 2024 17:34		AR\AJ	Ok
18	PB165747BL	PB165747BL	PL093445.D	19 Dec 2024 17:48		AR\AJ	Ok

Instrument ID: ECD_L

Daily Analysis Runlog For Sequence/QC Batch ID # PL121924

Review By	Abdul	Review On	12/20/2024 9:20:52 AM
Supervise By	Ankita	Supervise On	12/20/2024 10:44:33 AM
SubDirectory	PL121924	HP Acquire Method	HP Processing Method pl112524 8081
STD. NAME	STD REF.#		
Tune/Reschk	PP23793,PP23517		
Initial Calibration Stds	PP23673,PP23674,PP23675,PP23676,PP23677,PP23678,PP23679,PP23680,PP23681,PP23682,PP23683		
CCC	PP23686,PP23690,PP23695		
Internal Standard/PEM			
ICV/I.BLK	PP23687,PP23693,PP23698		
Surrogate Standard			
MS/MSD Standard			
LCS Standard			

19	PB165747BS	PB165747BS	PL093446.D	19 Dec 2024 18:01	Comp#6 recovery fail	AR\AJ	Not Ok
20	P5330-01	TP-5	PL093447.D	19 Dec 2024 18:15		AR\AJ	Ok,M
21	P5339-01	TR-06-12182024	PL093448.D	19 Dec 2024 18:28		AR\AJ	Ok,M
22	P5342-01	CHRT26634	PL093449.D	19 Dec 2024 18:42		AR\AJ	Ok,M
23	P5342-05	RB21198	PL093450.D	19 Dec 2024 18:55		AR\AJ	Ok,M
24	P5343-01	VNJ210	PL093451.D	19 Dec 2024 19:08		AR\AJ	Ok,M
25	P5343-03	VNJ281	PL093452.D	19 Dec 2024 19:22		AR\AJ	Ok,M
26	P5343-03MS	VNJ281MS	PL093453.D	19 Dec 2024 19:36	Comp#12,17 recovery fail	AR\AJ	Ok,M
27	P5343-03MSD	VNJ281MSD	PL093454.D	19 Dec 2024 19:49	Comp#12,17 recovery fail	AR\AJ	Ok,M
28	I.BLK	I.BLK	PL093455.D	19 Dec 2024 20:03		AR\AJ	Ok
29	PEM	PEM	PL093456.D	19 Dec 2024 20:16		AR\AJ	Ok,M
30	PSTDCCC050	PSTDCCC050	PL093457.D	19 Dec 2024 20:30		AR\AJ	Ok,M
31	P5342-03	HT2651	PL093458.D	19 Dec 2024 21:10	end ccc fail , f flag in TCMX , DCB high in 1st column and TCMX high in 2nd column	AR\AJ	Not Ok
32	I.BLK	I.BLK	PL093459.D	19 Dec 2024 21:51		AR\AJ	Ok
33	PSTDCCC050	PSTDCCC050	PL093460.D	19 Dec 2024 22:04	CCC fail	AR\AJ	Not Ok

M : Manual Integration