

Laboratory Control Sample/Laboratory Control Sample Duplicate Summary

SW-846

SDG No.: P5306

Client: Nobis Group

Analytical Method: **8081B** Datafile : PL093417.D

Lab Sample ID	Parameter	Spike	Result	Units	Rec	RPD	Qual	RPD		Limits	
								Qual	Low	High	RPD
PB165704BS	alpha-BHC	16.66	16.8	ug/kg	101				45	137	
	beta-BHC	16.66	16.8	ug/kg	101				50	136	
	delta-BHC	16.66	15.7	ug/kg	94				47	139	
	gamma-BHC (Lindane)	16.66	16.6	ug/kg	100				49	135	
	Heptachlor	16.66	17.6	ug/kg	106				47	136	
	Aldrin	16.66	16.6	ug/kg	100				45	136	
	Heptachlor epoxide	16.66	17.4	ug/kg	104				52	136	
	Endosulfan I	16.66	17.8	ug/kg	107				53	132	
	Dieldrin	16.66	17.9	ug/kg	107				56	136	
	4,4'-DDE	16.66	17.6	ug/kg	106				56	134	
	Endrin	16.66	18.8	ug/kg	113				57	140	
	Endosulfan II	16.66	18.3	ug/kg	110				53	134	
	4,4'-DDD	16.66	17.9	ug/kg	107				56	139	
	Endosulfan sulfate	16.66	17.9	ug/kg	107				55	136	
	4,4'-DDT	16.66	18.7	ug/kg	112				50	141	
	Methoxychlor	16.66	18.4	ug/kg	110				52	143	
	Endrin ketone	16.66	18.5	ug/kg	111				55	136	
	Endrin aldehyde	16.66	17.3	ug/kg	104				35	137	
	alpha-Chlordane	16.66	17.8	ug/kg	107				54	133	
	gamma-Chlordane	16.66	17.9	ug/kg	107				53	135	