

6C

SEMIVOLATILE ORGANICS INITIAL CALIBRATION DATA

Lab Name: CHEMTECH Contract: NOBI03
Lab Code: CHEM Case No.: P5306 SAS No.: P5306 SDG No.: P5306
Instrument ID: BNA_F Calibration Date(s): 12/16/2024 12/16/2024
Calibration Time(s): 12:13 16:15

LAB FILE ID:		RRF2.5 = BF140855.D		RRF005 = BF140856.D		RRF010 = BF140857.D		
		RRF020 = BF140858.D		RRF040 = BF140859.D		RRF050 = BF140860.D		
COMPOUND	RRF2.5	RRF005	RRF010	RRF020	RRF040	RRF050	RRF	% RSD
2-Fluorophenol		1.211	1.236	1.279	1.199	1.231	1.215	4.4
Phenol-d6		1.692	1.661	1.661	1.567	1.576	1.609	4.5
Nitrobenzene-d5		0.406	0.401	0.420	0.393	0.397	0.400	4.4
2-Fluorobiphenyl		1.560	1.410	1.538	1.369	1.345	1.455	5.8
2,4,6-Tribromophenol		0.235	0.236	0.242	0.240	0.235	0.237	2.6
Terphenyl-d14		1.354	1.247	1.218	1.206	1.212	1.268	5.6