

### VOLATILE ORGANICS INITIAL CALIBRATION DATA

|                     |                  |                      |                   |
|---------------------|------------------|----------------------|-------------------|
| Lab Name:           | <u>CHEMTECH</u>  | Contract:            | <u>NOBI03</u>     |
| Lab Code:           | <u>CHEM</u>      | Case No.:            | <u>P5306</u>      |
| Instrument ID:      | <u>MSVOA_Y</u>   | SAS No.:             | <u>P5306</u>      |
| Heated Purge: (Y/N) | <u>Y</u>         | SDG No.:             | <u>P5306</u>      |
| GC Column:          | <u>RXI-624</u>   | Calibration Date(s): | <u>12/23/2024</u> |
| ID:                 | <u>0.25 (mm)</u> | Calibration Time(s): | <u>11:30</u>      |
|                     |                  |                      | <u>13:24</u>      |

| LAB FILE ID:                   | RRF005 = VY020684.D | RRF010 = VY020685.D | RRF020 = VY020686.D | RRF050 = VY020687.D | RRF100 = VY020688.D | RRF150 = VY020689.D |       |       |
|--------------------------------|---------------------|---------------------|---------------------|---------------------|---------------------|---------------------|-------|-------|
| COMPOUND                       | RRF005              | RRF010              | RRF020              | RRF050              | RRF100              | RRF150              | RRF   | % RSD |
| Dichlorodifluoromethane        | 0.338               | 0.391               | 0.389               | 0.424               | 0.391               | 0.378               | 0.385 | 7.2   |
| Chloromethane                  | 0.125               | 0.165               | 0.151               | 0.176               | 0.161               | 0.163               | 0.157 | 11.1  |
| Vinyl Chloride                 | 0.163               | 0.196               | 0.184               | 0.212               | 0.193               | 0.200               | 0.191 | 8.6   |
| Bromomethane                   | 0.111               | 0.133               | 0.123               | 0.150               | 0.144               | 0.151               | 0.135 | 11.9  |
| Chloroethane                   | 0.097               | 0.115               | 0.113               | 0.128               | 0.120               | 0.125               | 0.116 | 9.3   |
| Tetrahydrofuran                | 0.035               | 0.039               | 0.040               | 0.053               | 0.039               | 0.044               | 0.042 | 15.4  |
| Trichlorofluoromethane         | 0.597               | 0.687               | 0.671               | 0.765               | 0.698               | 0.711               | 0.688 | 8     |
| 1,1,2-Trichlorotrifluoroethane | 0.316               | 0.386               | 0.379               | 0.413               | 0.379               | 0.381               | 0.376 | 8.5   |
| 1,1-Dichloroethene             | 0.279               | 0.341               | 0.315               | 0.371               | 0.349               | 0.346               | 0.333 | 9.6   |
| Acrylonitrile                  | 0.051               | 0.056               | 0.058               | 0.076               | 0.059               | 0.064               | 0.061 | 13.8  |
| Acetone                        | 0.059               | 0.049               | 0.047               | 0.058               | 0.043               | 0.048               | 0.051 | 12.4  |
| Carbon Disulfide               | 0.654               | 0.768               | 0.769               | 0.978               | 0.907               | 0.913               | 0.831 | 14.6  |
| Methyl tert-butyl Ether        | 0.769               | 0.869               | 0.904               | 1.042               | 0.908               | 0.949               | 0.907 | 9.9   |
| Methylene Chloride             | 0.324               | 0.346               | 0.317               | 0.370               | 0.333               | 0.332               | 0.337 | 5.5   |
| trans-1,2-Dichloroethene       | 0.310               | 0.350               | 0.350               | 0.404               | 0.370               | 0.374               | 0.360 | 8.7   |
| 1,1-Dichloroethane             | 0.500               | 0.583               | 0.597               | 0.668               | 0.616               | 0.611               | 0.596 | 9.3   |
| 2-Butanone                     | 0.069               | 0.070               | 0.067               | 0.085               | 0.064               | 0.071               | 0.071 | 10.2  |
| Carbon Tetrachloride           | 0.472               | 0.585               | 0.562               | 0.654               | 0.595               | 0.597               | 0.577 | 10.4  |
| 2,2-Dichloropropane            | 0.720               | 0.768               | 0.725               | 0.798               | 0.735               | 0.726               | 0.746 | 4.2   |
| cis-1,2-Dichloroethene         | 0.389               | 0.415               | 0.418               | 0.472               | 0.433               | 0.442               | 0.428 | 6.6   |
| Chloroform                     | 0.787               | 0.799               | 0.753               | 0.841               | 0.760               | 0.760               | 0.783 | 4.2   |
| 1,1,1-Trichloroethane          | 0.706               | 0.835               | 0.809               | 0.903               | 0.829               | 0.828               | 0.818 | 7.8   |
| 1,1-Dichloropropene            | 0.302               | 0.358               | 0.352               | 0.402               | 0.364               | 0.364               | 0.357 | 8.9   |
| Benzene                        | 0.870               | 1.019               | 1.009               | 1.163               | 1.057               | 1.058               | 1.030 | 9.3   |
| 1,2-Dichloroethane             | 0.271               | 0.324               | 0.324               | 0.384               | 0.327               | 0.334               | 0.327 | 11    |
| Trichloroethene                | 0.261               | 0.306               | 0.307               | 0.339               | 0.310               | 0.310               | 0.306 | 8.3   |
| 1,2-Dichloropropane            | 0.166               | 0.197               | 0.205               | 0.231               | 0.206               | 0.209               | 0.202 | 10.5  |
| Dibromomethane                 | 0.118               | 0.131               | 0.138               | 0.163               | 0.144               | 0.144               | 0.140 | 10.6  |
| Bromodichloromethane           | 0.326               | 0.384               | 0.392               | 0.461               | 0.404               | 0.409               | 0.396 | 11    |
| 4-Methyl-2-Pentanone           | 0.110               | 0.111               | 0.116               | 0.148               | 0.114               | 0.127               | 0.121 | 12    |

\* Compounds with required minimum RRF and maximum %RSD values.  
All other compounds must meet a minimum RRF of 0.010.  
RRF of 1,4-Dioxane = Value should be divide by 1000.

### VOLATILE ORGANICS INITIAL CALIBRATION DATA

|                     |                 |                      |                   |
|---------------------|-----------------|----------------------|-------------------|
| Lab Name:           | <u>CHEMTECH</u> | Contract:            | <u>NOBI03</u>     |
| Lab Code:           | <u>CHEM</u>     | Case No.:            | <u>P5306</u>      |
| Instrument ID:      | <u>MSVOA_Y</u>  | SAS No.:             | <u>P5306</u>      |
| Heated Purge: (Y/N) | <u>Y</u>        | SDG No.:             | <u>P5306</u>      |
| GC Column:          | <u>RXI-624</u>  | ID:                  | <u>0.25 (mm)</u>  |
|                     |                 | Calibration Date(s): | <u>12/23/2024</u> |
|                     |                 | Calibration Time(s): | <u>11:30</u>      |

|                           |                     |                     |                     |                     |                     |                     |       |       |
|---------------------------|---------------------|---------------------|---------------------|---------------------|---------------------|---------------------|-------|-------|
| LAB FILE ID:              | RRF005 = VY020684.D | RRF010 = VY020685.D | RRF020 = VY020686.D | RRF050 = VY020687.D | RRF100 = VY020688.D | RRF150 = VY020689.D |       |       |
| COMPOUND                  | RRF005              | RRF010              | RRF020              | RRF050              | RRF100              | RRF150              | RRF   | % RSD |
| Toluene                   | 0.582               | 0.687               | 0.693               | 0.792               | 0.723               | 0.729               | 0.701 | 9.8   |
| t-1,3-Dichloropropene     | 0.277               | 0.323               | 0.349               | 0.411               | 0.366               | 0.376               | 0.351 | 13.3  |
| cis-1,3-Dichloropropene   | 0.304               | 0.369               | 0.386               | 0.452               | 0.402               | 0.405               | 0.386 | 12.7  |
| 1,1,2-Trichloroethane     | 0.154               | 0.175               | 0.176               | 0.214               | 0.180               | 0.186               | 0.181 | 10.8  |
| 1,3-Dichloropropane       | 0.242               | 0.286               | 0.291               | 0.353               | 0.296               | 0.305               | 0.296 | 12.1  |
| 2-Hexanone                | 0.057               | 0.067               | 0.073               | 0.100               | 0.078               | 0.087               | 0.077 | 19.4  |
| Dibromochloromethane      | 0.244               | 0.271               | 0.287               | 0.335               | 0.292               | 0.303               | 0.289 | 10.6  |
| 1,2-Dibromoethane         | 0.132               | 0.172               | 0.168               | 0.203               | 0.166               | 0.176               | 0.170 | 13.5  |
| Tetrachloroethene         | 0.274               | 0.339               | 0.331               | 0.362               | 0.341               | 0.346               | 0.332 | 9.1   |
| Chlorobenzene             | 0.811               | 0.932               | 0.924               | 1.022               | 0.956               | 0.961               | 0.934 | 7.5   |
| 1,1,1,2-Tetrachloroethane | 0.321               | 0.368               | 0.360               | 0.398               | 0.369               | 0.373               | 0.365 | 6.9   |
| Ethyl Benzene             | 1.397               | 1.698               | 1.668               | 1.841               | 1.722               | 1.731               | 1.676 | 8.9   |
| m/p-Xylenes               | 0.531               | 0.645               | 0.644               | 0.703               | 0.656               | 0.656               | 0.639 | 8.9   |
| o-Xylene                  | 0.493               | 0.610               | 0.599               | 0.670               | 0.621               | 0.611               | 0.600 | 9.7   |
| Styrene                   | 0.823               | 1.026               | 1.020               | 1.119               | 1.039               | 1.033               | 1.010 | 9.8   |
| Bromoform                 | 0.180               | 0.195               | 0.208               | 0.247               | 0.206               | 0.219               | 0.209 | 11    |
| Isopropylbenzene          | 2.983               | 3.415               | 3.372               | 3.667               | 3.515               | 3.453               | 3.401 | 6.7   |
| 1,1,2,2-Tetrachloroethane | 0.386               | 0.444               | 0.430               | 0.498               | 0.424               | 0.445               | 0.438 | 8.3   |
| 1,2,3-Trichloropropane    | 0.305               | 0.372               | 0.341               | 0.349               | 0.346               | 0.357               | 0.345 | 6.5   |
| Bromobenzene              | 0.667               | 0.779               | 0.752               | 0.853               | 0.794               | 0.802               | 0.774 | 8     |
| n-propylbenzene           | 3.382               | 3.920               | 3.786               | 4.131               | 3.928               | 3.859               | 3.834 | 6.5   |
| 2-Chlorotoluene           | 1.923               | 2.228               | 2.169               | 2.368               | 2.221               | 2.188               | 2.183 | 6.7   |
| 1,3,5-Trimethylbenzene    | 2.458               | 2.840               | 2.814               | 3.058               | 2.911               | 2.845               | 2.821 | 7     |
| 4-Chlorotoluene           | 1.938               | 2.214               | 2.232               | 2.421               | 2.262               | 2.290               | 2.226 | 7.1   |
| tert-Butylbenzene         | 2.340               | 2.609               | 2.608               | 2.842               | 2.703               | 2.647               | 2.625 | 6.3   |
| 1,2,4-Trimethylbenzene    | 2.330               | 2.749               | 2.772               | 3.039               | 2.858               | 2.786               | 2.756 | 8.5   |
| sec-Butylbenzene          | 3.092               | 3.493               | 3.553               | 3.813               | 3.671               | 3.584               | 3.534 | 6.9   |
| p-Isopropyltoluene        | 2.598               | 3.016               | 3.111               | 3.376               | 3.260               | 3.178               | 3.090 | 8.8   |
| 1,3-Dichlorobenzene       | 1.340               | 1.499               | 1.475               | 1.675               | 1.528               | 1.510               | 1.504 | 7.1   |
| 1,4-Dichlorobenzene       | 1.290               | 1.485               | 1.473               | 1.605               | 1.502               | 1.489               | 1.474 | 6.9   |

\* Compounds with required minimum RRF and maximum %RSD values.  
All other compounds must meet a minimum RRF of 0.010.  
RRF of 1,4-Dioxane = Value should be divide by 1000.

### VOLATILE ORGANICS INITIAL CALIBRATION DATA

Lab Name: CHEMTECH Contract: NOBI03  
 Lab Code: CHEM Case No.: P5306 SAS No.: P5306 SDG No.: P5306  
 Instrument ID: MSVOA\_Y Calibration Date(s): 12/23/2024 12/23/2024  
 Heated Purge: (Y/N) Y Calibration Time(s): 11:30 13:24  
 GC Column: RXI-624 ID: 0.25 (mm)

|                             |        |                     |        |                     |        |                     |       |       |
|-----------------------------|--------|---------------------|--------|---------------------|--------|---------------------|-------|-------|
| LAB FILE ID:                |        | RRF005 = VY020684.D |        | RRF010 = VY020685.D |        | RRF020 = VY020686.D |       |       |
|                             |        | RRF050 = VY020687.D |        | RRF100 = VY020688.D |        | RRF150 = VY020689.D |       |       |
| COMPOUND                    | RRF005 | RRF010              | RRF020 | RRF050              | RRF100 | RRF150              | RRF   | % RSD |
| n-Butylbenzene              | 2.232  | 2.504               | 2.602  | 2.825               | 2.750  | 2.705               | 2.603 | 8.2   |
| 1,2-Dichlorobenzene         | 1.121  | 1.291               | 1.286  | 1.431               | 1.311  | 1.308               | 1.291 | 7.7   |
| 1,2-Dibromo-3-Chloropropane | 0.075  | 0.080               | 0.079  | 0.096               | 0.080  | 0.090               | 0.083 | 9.3   |
| 1,2,4-Trichlorobenzene      | 0.563  | 0.667               | 0.713  | 0.882               | 0.866  | 0.908               | 0.767 | 18.2  |
| Hexachlorobutadiene         | 0.466  | 0.530               | 0.561  | 0.628               | 0.622  | 0.618               | 0.571 | 11.3  |
| 1,2,3-Trichlorobenzene      | 0.465  | 0.545               | 0.607  | 0.747               | 0.721  | 0.760               | 0.641 | 18.8  |
| 1,2-Dichloroethane-d4       |        | 0.407               | 0.359  | 0.511               | 0.469  | 0.468               | 0.443 | 13.5  |
| Dibromofluoromethane        |        | 0.286               | 0.250  | 0.330               | 0.303  | 0.299               | 0.294 | 10    |
| Toluene-d8                  |        | 0.891               | 0.815  | 1.244               | 1.209  | 1.174               | 1.066 | 18.6  |
| 4-Bromofluorobenzene        |        | 0.373               | 0.339  | 0.424               | 0.408  | 0.396               | 0.388 | 8.6   |
| trans-1,4-Dichloro-2-butene | 0.128  | 0.153               | 0.155  | 0.195               | 0.154  | 0.167               | 0.159 | 13.8  |

\* Compounds with required minimum RRF and maximum %RSD values.  
 All other compounds must meet a minimum RRF of 0.010.  
 RRF of 1,4-Dioxane = Value should be divide by 1000.