

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH Contract: NOBI03
Lab Code: CHEM Case No.: P5306 SAS No.: P5306 SDG NO.: P5306
Lab File ID: VY020687.D Date Analyzed: 12/23/2024
Instrument ID: MSVOA_Y Time Analyzed: 12:38
GC Column: RXI-624 ID: 0.25 (mm) Heated Purge: (Y/N) Y

	IS1 AREA #	RT #	IS2 AREA #	RT #	IS3 AREA #	RT #
12 HOUR STD	100940	7.73	139898	8.63	121636	11.43
UPPER LIMIT	201880	8.225	279796	9.128	243272	11.926
LOWER LIMIT	50470	7.225	69949	8.128	60818	10.926
EPA SAMPLE NO.						
OU4-VSL-12-121224	94664	7.72	149835	8.62	128778	11.43
OU4-VSL-14-121224	94774	7.72	152816	8.62	128401	11.42
VY1223SBL01	103684	7.72	162906	8.62	133896	11.43
VY1223SBS01	116113	7.72	161575	8.62	134443	11.43

IS1 = Pentafluorobenzene
IS2 = 1,4-Difluorobenzene
IS3 = Chlorobenzene-d5

AREA UPPER LIMIT = +100% of internal standard area
AREA LOWER LIMIT = -50% of internal standard area
RT UPPER LIMIT = +0.50 minutes of internal standard RT
RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.
* Values outside of QC limits.