

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH Contract: NOBI03
 Lab Code: CHEM Case No.: P5306 SAS No.: P5306 SDG NO.: P5306
 Lab File ID: VY020687.D Date Analyzed: 12/23/2024
 Instrument ID: MSVOA_Y Time Analyzed: 12:38
 GC Column: RXI-624 ID: 0.25 (mm) Heated Purge: (Y/N) Y

	IS4 AREA #	RT #				
12 HOUR STD	61579	13.352				
UPPER LIMIT	123158	13.852				
LOWER LIMIT	30789.5	12.852				
EPA SAMPLE NO.						
OU4-VSL-12-121224	55306	13.35				
OU4-VSL-14-121224	54196	13.35				
VY1223SBL01	51849	13.36				
VY1223SBS01	68792	13.35				

IS4 = 1,4-Dichlorobenzene-d4

AREA UPPER LIMIT = +100% of internal standard area
 AREA LOWER LIMIT = -50% of internal standard area
 RT UPPER LIMIT = +0.50 minutes of internal standard RT
 RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.
 * Values outside of QC limits.