

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY122324\
 Data File : VY020703.D
 Acq On : 23 Dec 2024 19:47
 Operator : SY/MD
 Sample : VY1223SBSD01
 Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 24 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 VY1223SBSD01

Manual Integrations
 APPROVED

Reviewed By : Semsettin Yesilyurt 12/26/2024
 Supervised By : Mahesh Dadoda 12/26/2024

Quant Time: Dec 25 01:03:17 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y122324S.M
 Quant Title : SW846 8260
 QLast Update : Wed Dec 25 00:58:45 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.720	168	104834	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.622	114	154339	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.420	117	131930	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.353	152	69344	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.067	65	52754	56.819	ug/l	0.00
Spiked Amount 50.000	Range 50 - 163		Recovery = 113.640%			
35) Dibromofluoromethane	7.646	113	48159	53.141	ug/l	0.00
Spiked Amount 50.000	Range 54 - 147		Recovery = 106.280%			
50) Toluene-d8	10.109	98	184098	55.931	ug/l	0.00
Spiked Amount 50.000	Range 58 - 134		Recovery = 111.860%			
62) 4-Bromofluorobenzene	12.414	95	63417	52.934	ug/l	0.00
Spiked Amount 50.000	Range 29 - 146		Recovery = 105.860%			
Target Compounds					Qvalue	
2) Dichlorodifluoromethane	1.867	85	16152	19.999	ug/l	99
3) Chloromethane	2.074	50	6233	18.943	ug/l	100
4) Vinyl Chloride	2.214	62	7104	17.709	ug/l	93
5) Bromomethane	2.605	94	5714	20.171	ug/l	95
6) Chloroethane	2.745	64	4299	17.632	ug/l	96
7) Trichlorofluoromethane	3.068	101	28952	20.069	ug/l	100
8) Diethyl Ether	3.464	74	7345	21.271	ug/l	97
9) 1,1,2-Trichlorotrifluo...	3.830	101	15648	19.863	ug/l	98
10) Methyl Iodide	4.013	142	17395	19.168	ug/l	100
11) Tert butyl alcohol	4.879	59	6977	93.870	ug/l #	93
12) 1,1-Dichloroethene	3.799	96	14011	20.040	ug/l	98
13) Acrolein	3.671	56	2497	141.348	ug/l	97
14) Allyl chloride	4.397	41	16858	20.562	ug/l	100
15) Acrylonitrile	5.074	53	13947	109.579	ug/l	98
16) Acetone	3.873	43	10717	101.056	ug/l	98
17) Carbon Disulfide	4.116	76	33010	18.937	ug/l #	93
18) Methyl Acetate	4.409	43	6183	22.535	ug/l	99
19) Methyl tert-butyl Ether	5.135	73	42423	22.307	ug/l	98
20) Methylene Chloride	4.635	84	15288	21.633	ug/l	86
21) trans-1,2-Dichloroethene	5.128	96	15798	20.956	ug/l	92
22) Diisopropyl ether	6.025	45	37677	22.479	ug/l #	96
23) Vinyl Acetate	5.970	43	107461	108.272	ug/l	99
24) 1,1-Dichloroethane	5.933	63	27342	21.882	ug/l #	95
25) 2-Butanone	6.903	43	15672	105.199	ug/l	91
26) 2,2-Dichloropropane	6.896	77	31303	20.024	ug/l	98
27) cis-1,2-Dichloroethene	6.903	96	19772	22.020	ug/l	98
28) Bromochloromethane	7.256	49	6466	19.677	ug/l	99
29) Tetrahydrofuran	7.274	42	9534	108.884	ug/l	98
30) Chloroform	7.427	83	34856	21.222	ug/l	97
31) Cyclohexane	7.713	56	21545	19.666	ug/l	96
32) 1,1,1-Trichloroethane	7.628	97	36239	21.120	ug/l	97
36) 1,1-Dichloropropene	7.841	75	22098	20.061	ug/l	97
37) Ethyl Acetate	6.994	43	7719	20.609	ug/l	97
38) Carbon Tetrachloride	7.829	117	35214	19.755	ug/l	95
39) Methylcyclohexane	9.116	83	27500	19.641	ug/l	99
40) Benzene	8.085	78	65063	20.474	ug/l	96

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY122324\
 Data File : VY020703.D
 Acq On : 23 Dec 2024 19:47
 Operator : SY/MD
 Sample : VY1223SBSD01
 Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 24 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 VY1223SBSD01

Quant Time: Dec 25 01:03:17 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y122324S.M
 Quant Title : SW846 8260
 QLast Update : Wed Dec 25 00:58:45 2024
 Response via : Initial Calibration

Manual Integrations
 APPROVED

Reviewed By :Semsettin Yesilyurt 12/26/2024
 Supervised By :Mahesh Dadoda 12/26/2024

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
41) Methacrylonitrile	7.244	41	4454	21.838	ug/l	94
42) 1,2-Dichloroethane	8.165	62	21724	21.495	ug/l	99
43) Isopropyl Acetate	8.207	43	15842	21.172	ug/l #	83
44) Trichloroethene	8.872	130	18586	19.707	ug/l	99
45) 1,2-Dichloropropane	9.146	63	13056	20.911	ug/l	98
46) Dibromomethane	9.238	93	9334	21.655	ug/l	97
47) Bromodichloromethane	9.427	83	26059	21.312	ug/l	94
48) Methyl methacrylate	9.225	41	7217	20.347	ug/l	96
49) 1,4-Dioxane	9.238	88	1762	435.830	ug/l #	90
51) 4-Methyl-2-Pentanone	10.006	43	38706	103.556	ug/l	98
52) Toluene	10.176	92	44815	20.709	ug/l	98
53) t-1,3-Dichloropropene	10.402	75	23137	21.385	ug/l	100
54) cis-1,3-Dichloropropene	9.859	75	25177	21.119	ug/l	92
55) 1,1,2-Trichloroethane	10.579	97	11682	20.931	ug/l #	90
56) Ethyl methacrylate	10.445	69	15165	21.301	ug/l	97
57) 1,3-Dichloropropane	10.725	76	19061	20.893	ug/l	95
58) 2-Chloroethyl Vinyl ether	9.719	63	27253	102.437	ug/l	98
59) 2-Hexanone	10.768	43	24776	104.066	ug/l	99
60) Dibromochloromethane	10.914	129	19154	21.495	ug/l	96
61) 1,2-Dibromoethane	11.018	107	11054	21.127	ug/l	100
64) Tetrachloroethene	10.652	164	17595	20.081	ug/l	94
65) Chlorobenzene	11.451	112	52606	21.338	ug/l	97
66) 1,1,1,2-Tetrachloroethane	11.524	131	20004	20.778	ug/l	97
67) Ethyl Benzene	11.524	91	91241	20.628	ug/l	95
68) m/p-Xylenes	11.633	106	69958	41.477	ug/l	100
69) o-Xylene	11.957	106	34079	21.509	ug/l	98
70) Styrene	11.975	104	56763	21.300	ug/l	99
71) Bromoform	12.139	173	11461	20.784	ug/l #	98
73) Isopropylbenzene	12.261	105	93189	19.757	ug/l	100
74) N-amyl acetate	12.078	43	13493	20.358	ug/l	98
75) 1,1,2,2-Tetrachloroethane	12.511	83	12470	20.545	ug/l	97
76) 1,2,3-Trichloropropane	12.560	75	9949m	20.804	ug/l	
77) Bromobenzene	12.536	156	21199	19.737	ug/l	99
78) n-propylbenzene	12.603	91	103950	19.548	ug/l	99
79) 2-Chlorotoluene	12.688	91	60340	19.932	ug/l	100
80) 1,3,5-Trimethylbenzene	12.743	105	79726	20.379	ug/l	100
81) trans-1,4-Dichloro-2-b...	12.310	75	4098	18.628	ug/l	97
82) 4-Chlorotoluene	12.786	91	62684	20.303	ug/l	99
83) tert-Butylbenzene	13.005	119	74013	20.332	ug/l	98
84) 1,2,4-Trimethylbenzene	13.048	105	79650	20.842	ug/l	99
85) sec-Butylbenzene	13.182	105	98630	20.123	ug/l	99
86) p-Isopropyltoluene	13.298	119	88167	20.575	ug/l	99
87) 1,3-Dichlorobenzene	13.292	146	43865	21.023	ug/l	98
88) 1,4-Dichlorobenzene	13.371	146	42330	20.707	ug/l	98
89) n-Butylbenzene	13.621	91	72816	20.171	ug/l	99
90) Hexachloroethane	13.883	117	15902	19.561	ug/l	95
91) 1,2-Dichlorobenzene	13.664	146	37259	20.807	ug/l	99
92) 1,2-Dibromo-3-Chloropr...	14.279	75	2361	20.414	ug/l	94
93) 1,2,4-Trichlorobenzene	14.932	180	21965	20.661	ug/l	97
94) Hexachlorobutadiene	15.029	225	16383	20.697	ug/l	98
95) Naphthalene	15.151	128	34561	19.827	ug/l	100
96) 1,2,3-Trichlorobenzene	15.340	180	17939	20.184	ug/l	99

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY122324\
Data File : VY020703.D
Acq On : 23 Dec 2024 19:47
Operator : SY/MD
Sample : VY1223SBSD01
Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 24 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
VY1223SBSD01

Manual Integrations
APPROVED

Reviewed By :Semsettin Yesilyurt 12/26/2024
Supervised By :Mahesh Dadoda 12/26/2024

Quant Time: Dec 25 01:03:17 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y122324S.M
Quant Title : SW846 8260
QLast Update : Wed Dec 25 00:58:45 2024
Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
----------	------	------	----------	------	-------	----------

(#) = qualifier out of range (m) = manual integration (+) = signals summed


Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY122324\
Data File : VY020703.D
Acq On : 23 Dec 2024 19:47
Operator : SY/MD
Sample : VY1223SBSD01
Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 24 Sample Multiplier: 1


Instrument :
MSVOA_Y
ClientSampleId :
VY1223SBSD01

Manual Integrations
APPROVED

Reviewed By : Semsettin Yesilyurt 12/26/2024
Supervised By : Mahesh Dadoda 12/26/2024


Quant Time: Dec 25 01:03:17 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y122324S.M
Quant Title : SW846 8260
QLast Update : Wed Dec 25 00:58:45 2024
Response via : Initial Calibration





#1
Pentafluorobenzene
Concen: 50.000 ug/l
RT: 7.720 min Scan# 98
Delta R.T. -0.005 min
Lab File: VY020703.D
Acq: 23 Dec 2024 19:47




Instrument :
MSVOA_Y
ClientSampleId :
VY1223SBSD01



Tgt Ion:168 Resp: 104834
Ion Ratio Lower Upper
168 100
99 58.0 43.8 65.6




Manual Integrations
APPROVED


Reviewed By :Semsettin Yesilyurt 12/26/2024
Supervised By :Mahesh Dadoda 12/26/2024



#2
Dichlorodifluoromethane
Concen: 19.999 ug/l
RT: 1.867 min Scan# 24
Delta R.T. -0.006 min
Lab File: VY020703.D
Acq: 23 Dec 2024 19:47


Tgt Ion: 85 Resp: 16152
Ion Ratio Lower Upper
85 100
87 31.7 15.4 46.4





#3
Chloromethane
Concen: 18.943 ug/l
RT: 2.074 min Scan# 58
Delta R.T. -0.006 min
Lab File: VY020703.D
Acq: 23 Dec 2024 19:47




Instrument :
MSVOA_Y
ClientSampleId :
VY1223SBSD01



Tgt Ion: 50 Resp: 6233
Ion Ratio Lower Upper
50 100
52 34.9 28.2 42.2




Manual Integrations
APPROVED


Reviewed By :Semsettin Yesilyurt 12/26/2024
Supervised By :Mahesh Dadoda 12/26/2024



#4
Vinyl Chloride
Concen: 17.709 ug/l
RT: 2.214 min Scan# 81
Delta R.T. 0.000 min
Lab File: VY020703.D
Acq: 23 Dec 2024 19:47

Tgt Ion: 62 Resp: 7104
Ion Ratio Lower Upper
62 100
64 29.1 26.6 40.0





#5

Bromomethane

Concen: 20.171 ug/l

RT: 2.605 min Scan# 14

Delta R.T. -0.006 min

Lab File: VY020703.D

Acq: 23 Dec 2024 19:47

Instrument :

MSVOA_Y

ClientSampleId :

VY1223SBSD01

Tgt Ion: 94 Resp: 5714

Ion Ratio Lower Upper

94 100





96 84.5 71.0 106.6

Manual Integrations

APPROVED

Reviewed By :Semsettin Yesilyurt 12/26/2024

Supervised By :Mahesh Dadoda 12/26/2024



#6

Chloroethane

Concen: 17.632 ug/l

RT: 2.745 min Scan# 168

Delta R.T. -0.006 min

Lab File: VY020703.D




Acq: 23 Dec 2024 19:47


Tgt Ion: 64 Resp: 4299

Ion Ratio Lower Upper

64 100


66 32.7 27.8 41.8





Trichlorofluoromethane
Concen: 20.069 ug/l
RT: 3.068 min Scan# 223
Delta R.T. -0.012 min
Lab File: VY020703.D
Acq: 23 Dec 2024 19:47




Instrument :
MSVOA_Y
ClientSampleId :
VY1223SBS01



Tgt Ion: 101 Resp: 28952
Ion Ratio Lower Upper
101 100
103 64.6 51.4 77.0




Manual Integrations
APPROVED


Reviewed By : Semsettin Yesilyurt 12/26/2024
Supervised By : Mahesh Dadoda 12/26/2024



Diethyl Ether
Concen: 21.271 ug/l
RT: 3.464 min Scan# 286
Delta R.T. -0.006 min
Lab File: VY020703.D
Acq: 23 Dec 2024 19:47

Tgt Ion: 74 Resp: 7345
Ion Ratio Lower Upper
74 100
45 74.4 35.8 107.4









#9
1,1,2-Trichlorotrifluoroethane
Concen: 19.863 ug/l
RT: 3.830 min Scan# 347
Delta R.T. -0.006 min
Lab File: VY020703.D
Acq: 23 Dec 2024 19:47

Instrument :
MSVOA_Y
ClientSampleId :
VY1223SBS01

Tgt Ion:101 Resp: 15648
Ion Ratio Lower Upper
101 100
85 49.6 38.2 57.2
151 87.2 68.4 102.6




Manual Integrations
APPROVED


Reviewed By :Semsettin Yesilyurt 12/26/2024
Supervised By :Mahesh Dadoda 12/26/2024



#10
Methyl Iodide
Concen: 19.168 ug/l
RT: 4.013 min Scan# 376
Delta R.T. -0.012 min
Lab File: VY020703.D
Acq: 23 Dec 2024 19:47

Tgt Ion:142 Resp: 17395
Ion Ratio Lower Upper
142 100
127 61.1 48.8 73.2
141 14.5 11.9 17.9





#11

Tert butyl alcohol

Concen: 93.870 ug/l

RT: 4.879 min Scan# 51

Delta R.T. 0.000 min

Lab File: VY020703.D

Acq: 23 Dec 2024 19:47

Instrument :

MSVOA_Y

ClientSampleId :

VY1223SBS01

Tgt Ion: 59 Resp: 6977

Ion Ratio Lower Upper

59 100





57 5.9 2.8 4.2#

Manual Integrations

APPROVED

Reviewed By :Semsettin Yesilyurt 12/26/2024

Supervised By :Mahesh Dadoda 12/26/2024



#12

1,1-Dichloroethene

Concen: 20.040 ug/l

RT: 3.799 min Scan# 341

Delta R.T. -0.012 min

Lab File: VY020703.D

Acq: 23 Dec 2024 19:47




Tgt Ion: 96 Resp: 14011


Ion Ratio Lower Upper

96 100

61 155.9 127.1 190.7


98 65.0 53.4 80.0





#13
Acrolein
Concen: 141.348 ug/l
RT: 3.671 min Scan# 320
Delta R.T. 0.000 min
Lab File: VY020703.D
Acq: 23 Dec 2024 19:47




Instrument :
MSVOA_Y
ClientSampleId :
VY1223SBS01



Tgt Ion: 56 Resp: 2497
Ion Ratio Lower Upper
56 100
55 68.6 57.0 85.4




Manual Integrations
APPROVED


Reviewed By :Semsettin Yesilyurt 12/26/2024
Supervised By :Mahesh Dadoda 12/26/2024



#14
Allyl chloride
Concen: 20.562 ug/l
RT: 4.397 min Scan# 439
Delta R.T. -0.012 min
Lab File: VY020703.D
Acq: 23 Dec 2024 19:47

Tgt Ion: 41 Resp: 16858
Ion Ratio Lower Upper
41 100
39 91.2 73.5 110.3
76 52.3 41.8 62.8





#15

Acrylonitrile

Concen: 109.579 ug/l

RT: 5.074 min Scan# 551

Delta R.T. -0.006 min

Lab File: VY020703.D

Acq: 23 Dec 2024 19:47

Instrument :

MSVOA_Y

ClientSampleId :





VY1223SBS01

Tgt Ion: 53	Resp: 13947
Ion Ratio	Lower Upper
53 100	
52 82.4	67.1 100.7
51 37.2	31.4 47.0

Manual Integrations

APPROVED

Reviewed By :Semsettin Yesilyurt 12/26/2024
Supervised By :Mahesh Dadoda 12/26/2024



#16

Acetone

Concen: 101.056 ug/l




RT: 3.873 min Scan# 353


Delta R.T. -0.012 min

Lab File: VY020703.D

Acq: 23 Dec 2024 19:47


Tgt Ion: 43	Resp: 10717
Ion Ratio	Lower Upper
43 100	
58 32.2	25.0 37.6





#17
Carbon Disulfide
Concen: 18.937 ug/l
RT: 4.116 min Scan# 395
Delta R.T. -0.012 min
Lab File: VY020703.D
Acq: 23 Dec 2024 19:47




Instrument :
MSVOA_Y
ClientSampleId :
VY1223SBS01



Tgt Ion: 76 Resp: 33010
Ion Ratio Lower Upper
76 100
78 11.2 6.9 10.3#




Manual Integrations
APPROVED


Reviewed By :Semsettin Yesilyurt 12/26/2024
Supervised By :Mahesh Dadoda 12/26/2024



#18
Methyl Acetate
Concen: 22.535 ug/l
RT: 4.409 min Scan# 441
Delta R.T. 0.006 min
Lab File: VY020703.D
Acq: 23 Dec 2024 19:47

Tgt Ion: 43 Resp: 6183
Ion Ratio Lower Upper
43 100
74 32.2 25.3 37.9









#19
Methyl tert-butyl Ether
Concen: 22.307 ug/l
RT: 5.135 min Scan# 560
Delta R.T. 0.000 min
Lab File: VY020703.D
Acq: 23 Dec 2024 19:47

Instrument :
MSVOA_Y
ClientSampleId :
VY1223SBS01

Tgt Ion: 73 Resp: 42423
Ion Ratio Lower Upper
73 100
57 18.8 14.4 21.6




Manual Integrations
APPROVED


Reviewed By :Semsettin Yesilyurt 12/26/2024
Supervised By :Mahesh Dadoda 12/26/2024



#20
Methylene Chloride
Concen: 21.633 ug/l
RT: 4.635 min Scan# 478
Delta R.T. 0.000 min
Lab File: VY020703.D
Acq: 23 Dec 2024 19:47


Tgt Ion: 84 Resp: 15288
Ion Ratio Lower Upper
84 100
49 86.9 86.4 129.6
51 30.7 25.5 38.3
86 63.8 56.6 85.0





#21
trans-1,2-Dichloroethene
Concen: 20.956 ug/l
RT: 5.128 min Scan# 559
Delta R.T. -0.012 min
Lab File: VY020703.D
Acq: 23 Dec 2024 19:47




Instrument :
MSVOA_Y
ClientSampleId :
VY1223SBSD01




Tgt Ion: 96 Resp: 15798
Ion Ratio Lower Upper
96 100
61 108.1 96.1 144.1
98 62.7 52.7 79.1

Manual Integrations
APPROVED



Reviewed By :Semsettin Yesilyurt 12/26/2024
Supervised By :Mahesh Dadoda 12/26/2024




#22
Diisopropyl ether
Concen: 22.479 ug/l
RT: 6.025 min Scan# 706
Delta R.T. -0.012 min
Lab File: VY020703.D
Acq: 23 Dec 2024 19:47



Tgt Ion: 45 Resp: 37677
Ion Ratio Lower Upper
45 100
43 68.4 51.3 76.9
87 38.1 30.6 46.0
59 16.0 9.6 14.4#





#23

Vinyl Acetate

Concen: 108.272 ug/l

RT: 5.970 min Scan# 698

Delta R.T. -0.006 min

Lab File: VY020703.D

Acq: 23 Dec 2024 19:47

Instrument :

MSVOA_Y

ClientSampleId :

VY1223SBS01

Tgt Ion: 43 Resp: 107461

Ion Ratio Lower Upper

43 100





86 14.7 12.0 18.0

Manual Integrations

APPROVED

Reviewed By :Semsettin Yesilyurt 12/26/2024

Supervised By :Mahesh Dadoda 12/26/2024



#24

1,1-Dichloroethane

Concen: 21.882 ug/l

RT: 5.933 min Scan# 691

Delta R.T. 0.000 min

Lab File: VY020703.D

Acq: 23 Dec 2024 19:47




Tgt Ion: 63 Resp: 27342


Ion Ratio Lower Upper

63 100

98 6.7 4.0 12.0

100 6.8 2.1 6.2#





#25

2-Butanone

Concen: 105.199 ug/l

RT: 6.903 min Scan# 851

Delta R.T. -0.006 min

Lab File: VY020703.D

Acq: 23 Dec 2024 19:47

Instrument :

MSVOA_Y

ClientSampleId :

VY1223SBSD01

Tgt Ion: 43 Resp: 15672

Ion Ratio Lower Upper

43 100





72 31.8 29.5 44.3

Manual Integrations

APPROVED

Reviewed By :Semsettin Yesilyurt 12/26/2024

Supervised By :Mahesh Dadoda 12/26/2024



#26

2,2-Dichloropropane

Concen: 20.024 ug/l

RT: 6.896 min Scan# 849

Delta R.T. -0.006 min

Lab File: VY020703.D




Acq: 23 Dec 2024 19:47


Tgt Ion: 77 Resp: 31303

Ion Ratio Lower Upper

77 100


97 20.9 10.8 32.4





#27
cis-1,2-Dichloroethene
Concen: 22.020 ug/l
RT: 6.903 min Scan# 851
Delta R.T. -0.006 min
Lab File: VY020703.D
Acq: 23 Dec 2024 19:47




Instrument :
MSVOA_Y
ClientSampleId :
VY1223SBS01



Tgt Ion: 96 Resp: 19772
Ion Ratio Lower Upper
96 100
61 135.1 0.0 277.2
98 64.4 0.0 131.0




Manual Integrations
APPROVED


Reviewed By :Semsettin Yesilyurt 12/26/2024
Supervised By :Mahesh Dadoda 12/26/2024



#28
Bromochloromethane
Concen: 19.677 ug/l
RT: 7.256 min Scan# 908
Delta R.T. -0.006 min
Lab File: VY020703.D
Acq: 23 Dec 2024 19:47

Tgt Ion: 49 Resp: 6466
Ion Ratio Lower Upper
49 100
129 2.0 0.0 5.4
130 109.6 87.2 130.8









#29
Tetrahydrofuran
Concen: 108.884 ug/l
RT: 7.274 min Scan# 911
Delta R.T. 0.000 min
Lab File: VY020703.D
Acq: 23 Dec 2024 19:47

Instrument :
MSVOA_Y
ClientSampleId :
VY1223SBS01

Tgt Ion: 42 Resp: 9534
Ion Ratio Lower Upper
42 100
72 59.0 47.4 71.0
71 56.9 42.8 64.2




Manual Integrations
APPROVED


Reviewed By :Semsettin Yesilyurt 12/26/2024
Supervised By :Mahesh Dadoda 12/26/2024



#30
Chloroform
Concen: 21.222 ug/l
RT: 7.427 min Scan# 936
Delta R.T. -0.006 min
Lab File: VY020703.D
Acq: 23 Dec 2024 19:47

Tgt Ion: 83 Resp: 34856
Ion Ratio Lower Upper
83 100
85 64.7 50.2 75.2





#31

Cyclohexane

Concen: 19.666 ug/l

RT: 7.713 min Scan# 98

Delta R.T. -0.006 min

Lab File: VY020703.D


Acq: 23 Dec 2024 19:47

Instrument :

MSVOA_Y

ClientSampleId :

VY1223SBS01






Tgt Ion:	56	Resp:	21545
Ion Ratio	Lower	Upper	
56	100		
69	39.7	31.3	46.9
84	99.7	83.9	125.9

Manual Integrations

APPROVED

Reviewed By :Semsettin Yesilyurt 12/26/2024
Supervised By :Mahesh Dadoda 12/26/2024



#32

1,1,1-Trichloroethane

Concen: 21.120 ug/l




RT: 7.628 min Scan# 969


Delta R.T. -0.006 min

Lab File: VY020703.D

Acq: 23 Dec 2024 19:47


Tgt Ion:	97	Resp:	36239
Ion Ratio	Lower	Upper	
97	100		
99	66.0	51.0	76.6
61	39.2	30.0	45.0





#33
1,2-Dichloroethane-d4
Concen: 56.819 ug/l
RT: 8.067 min Scan# 1042
Delta R.T. -0.006 min
Lab File: VY020703.D
Acq: 23 Dec 2024 19:47




Instrument :
MSVOA_Y
ClientSampleId :
VY1223SBSD01




Tgt Ion: 65 Resp: 52754
Ion Ratio Lower Upper
65 100
67 48.5 0.0 98.0

Manual Integrations
APPROVED



Reviewed By :Semsettin Yesilyurt 12/26/2024
Supervised By :Mahesh Dadoda 12/26/2024




#34
1,4-Difluorobenzene
Concen: 50.000 ug/l
RT: 8.622 min Scan# 1132
Delta R.T. -0.006 min
Lab File: VY020703.D
Acq: 23 Dec 2024 19:47



Tgt Ion:114 Resp: 154339
Ion Ratio Lower Upper
114 100
63 18.3 0.0 38.4
88 15.4 0.0 30.4





#35

Dibromofluoromethane

Concen: 53.141 ug/l

RT: 7.646 min Scan# 97

Delta R.T. 0.000 min

Lab File: VY020703.D

Acq: 23 Dec 2024 19:47

Instrument :

MSVOA_Y

ClientSampleId :





VY1223SBS01

Tgt Ion:	113	Resp:	48159
Ion Ratio	Lower	Upper	
113	100		
111	99.6	79.8	119.8
192	21.6	17.0	25.6

Manual Integrations

APPROVED

Reviewed By :Semsettin Yesilyurt 12/26/2024
Supervised By :Mahesh Dadoda 12/26/2024



#36

1,1-Dichloropropene

Concen: 20.061 ug/l




RT: 7.841 min Scan# 1004


Delta R.T. -0.012 min

Lab File: VY020703.D

Acq: 23 Dec 2024 19:47


Tgt Ion:	75	Resp:	22098
Ion Ratio	Lower	Upper	
75	100		
110	37.1	19.6	58.7
77	29.9	25.1	37.7





#37
Ethyl Acetate
Concen: 20.609 ug/l
RT: 6.994 min Scan# 865
Delta R.T. 0.000 min
Lab File: VY020703.D
Acq: 23 Dec 2024 19:47




Instrument :
MSVOA_Y
ClientSampleId :
VY1223SBS01



Tgt Ion: 43 Resp: 7719
Ion Ratio Lower Upper
43 100
61 13.2 11.7 17.5
70 12.5 11.0 16.4




Manual Integrations
APPROVED


Reviewed By :Semsettin Yesilyurt 12/26/2024
Supervised By :Mahesh Dadoda 12/26/2024



#38
Carbon Tetrachloride
Concen: 19.755 ug/l
RT: 7.829 min Scan# 1002
Delta R.T. -0.006 min
Lab File: VY020703.D
Acq: 23 Dec 2024 19:47

Tgt Ion:117 Resp: 35214
Ion Ratio Lower Upper
117 100
119 97.4 74.4 111.6
121 33.0 24.2 36.2









#39
Methylcyclohexane
Concen: 19.641 ug/l
RT: 9.116 min Scan# 1214
Delta R.T. -0.006 min
Lab File: VY020703.D
Acq: 23 Dec 2024 19:47

Instrument :
MSVOA_Y
ClientSampleId :
VY1223SBSD01

Tgt Ion	Ratio	Lower	Upper
83	100		
55	67.8	55.0	82.4
98	49.1	39.0	58.6




Manual Integrations
APPROVED


Reviewed By :Semsettin Yesilyurt 12/26/2024
Supervised By :Mahesh Dadoda 12/26/2024



#40
Benzene
Concen: 20.474 ug/l
RT: 8.085 min Scan# 1044
Delta R.T. -0.006 min
Lab File: VY020703.D
Acq: 23 Dec 2024 19:47


Tgt Ion	Ratio	Lower	Upper
78	100		
77	26.0	19.2	28.8





#41
Methacrylonitrile
Concen: 21.838 ug/l
RT: 7.244 min Scan# 904
Delta R.T. 0.012 min
Lab File: VY020703.D
Acq: 23 Dec 2024 19:47




Instrument :
MSVOA_Y
ClientSampleId :
VY1223SBSD01



Tgt Ion: 41	Resp: 4454
Ion Ratio	Lower Upper
41 100	
39 61.2	46.6 70.0
67 88.7	74.7 112.1
52 34.2	32.2 48.4




Manual Integrations
APPROVED


Reviewed By :Semsettin Yesilyurt 12/26/2024
Supervised By :Mahesh Dadoda 12/26/2024



#42
1,2-Dichloroethane
Concen: 21.495 ug/l
RT: 8.165 min Scan# 1057
Delta R.T. -0.006 min
Lab File: VY020703.D
Acq: 23 Dec 2024 19:47

Tgt	Ion: 62	Resp:	21724
Ion	Ratio	Lower	Upper
62	100		
98	8.6	0.0	16.6









#43
Isopropyl Acetate
Concen: 21.172 ug/l
RT: 8.207 min Scan# 1064
Delta R.T. 0.000 min
Lab File: VY020703.D
Acq: 23 Dec 2024 19:47

Instrument : MSVOA_Y
ClientSampleId : VY1223SBSD01

Tgt Ion: 43	Resp: 15842
Ion Ratio	Lower Upper
43 100	
61 18.8	27.0 40.4
87 17.7	14.2 21.4




Manual Integrations
APPROVED


Reviewed By :Semsettin Yesilyurt 12/26/2024
Supervised By :Mahesh Dadoda 12/26/2024



#44
Trichloroethene
Concen: 19.707 ug/l
RT: 8.872 min Scan# 1173
Delta R.T. -0.006 min
Lab File: VY020703.D
Acq: 23 Dec 2024 19:47


Tgt Ion:130	Resp: 18586
Ion Ratio	Lower Upper
130 100	
95 98.6	0.0 194.8





#45
 1,2-Dichloropropane
 Concen: 20.911 ug/l
 RT: 9.146 min Scan# 1218
 Delta R.T. 0.000 min
 Lab File: VY020703.D
 Acq: 23 Dec 2024 19:47




Instrument :
 MSVOA_Y
 ClientSampleId :
 VY1223SBS01



Tgt Ion:	63	Resp:	13056
Ion Ratio	Lower	Upper	
63	100		
65	31.8	24.6	36.8




Manual Integrations
 APPROVED


Reviewed By :Semsettin Yesilyurt 12/26/2024
 Supervised By :Mahesh Dadoda 12/26/2024



#46
 Dibromomethane
 Concen: 21.655 ug/l
 RT: 9.238 min Scan# 1233
 Delta R.T. -0.006 min
 Lab File: VY020703.D
 Acq: 23 Dec 2024 19:47

Tgt Ion:	93	Resp:	9334
Ion Ratio	Lower	Upper	
93	100		
95	81.9	67.8	101.8
174	107.9	89.4	134.0









#47
Bromodichloromethane
Concen: 21.312 ug/l
RT: 9.427 min Scan# 1265
Delta R.T. -0.006 min
Lab File: VY020703.D
Acq: 23 Dec 2024 19:47

Instrument :
MSVOA_Y
ClientSampleId :
VY1223SBSD01

Tgt Ion	83	Resp	26059
Ion Ratio	83	100	
Lower	85	64.5	47.6
Upper	127	9.7	7.9




Manual Integrations
APPROVED


Reviewed By :Semsettin Yesilyurt 12/26/2024
Supervised By :Mahesh Dadoda 12/26/2024



#48
Methyl methacrylate
Concen: 20.347 ug/l
RT: 9.225 min Scan# 1231
Delta R.T. -0.006 min
Lab File: VY020703.D
Acq: 23 Dec 2024 19:47


Tgt Ion	41	Resp	7217
Ion Ratio	41	100	
Lower	69	105.4	83.1
Upper	39	77.6	57.3





#49
1,4-Dioxane
Concen: 435.830 ug/l
RT: 9.238 min Scan# 1233
Delta R.T. -0.006 min
Lab File: VY020703.D
Acq: 23 Dec 2024 19:47




Instrument :
MSVOA_Y
ClientSampleId :
VY1223SBSD01



Tgt Ion: 88 Resp: 1762
Ion Ratio Lower Upper
88 100
43 38.4 22.1 33.1#
58 54.5 46.7 70.1




Manual Integrations
APPROVED


Reviewed By :Semsettin Yesilyurt 12/26/2024
Supervised By :Mahesh Dadoda 12/26/2024



#50
Toluene-d8
Concen: 55.931 ug/l
RT: 10.109 min Scan# 1376
Delta R.T. -0.006 min
Lab File: VY020703.D
Acq: 23 Dec 2024 19:47

Tgt Ion: 98 Resp: 184098
Ion Ratio Lower Upper
98 100
100 65.9 52.2 78.4









#51
4-Methyl-2-Pentanone
Concen: 103.556 ug/l
RT: 10.006 min Scan# 1359
Delta R.T. 0.000 min
Lab File: VY020703.D
Acq: 23 Dec 2024 19:47

Instrument :
MSVOA_Y
ClientSampleId :
VY1223SBS01

Tgt Ion: 43 Resp: 38706
Ion Ratio Lower Upper
43 100
58 45.0 35.1 52.7




Manual Integrations
APPROVED


Reviewed By :Semsettin Yesilyurt 12/26/2024
Supervised By :Mahesh Dadoda 12/26/2024



#52
Toluene
Concen: 20.709 ug/l
RT: 10.176 min Scan# 1387
Delta R.T. -0.006 min
Lab File: VY020703.D
Acq: 23 Dec 2024 19:47

Tgt Ion: 92 Resp: 44815
Ion Ratio Lower Upper
92 100
91 174.2 137.7 206.5









#53
t-1,3-Dichloropropene
Concen: 21.385 ug/l
RT: 10.402 min Scan# 1424
Delta R.T. 0.000 min
Lab File: VY020703.D
Acq: 23 Dec 2024 19:47

Instrument :
MSVOA_Y
ClientSampleId :
VY1223SBS01

Tgt Ion: 75 Resp: 23137
Ion Ratio Lower Upper
75 100
77 31.4 25.1 37.7




Manual Integrations
APPROVED


Reviewed By :Semsettin Yesilyurt 12/26/2024
Supervised By :Mahesh Dadoda 12/26/2024



#54
cis-1,3-Dichloropropene
Concen: 21.119 ug/l
RT: 9.859 min Scan# 1335
Delta R.T. -0.006 min
Lab File: VY020703.D
Acq: 23 Dec 2024 19:47

Tgt Ion: 75 Resp: 25177
Ion Ratio Lower Upper
75 100
77 27.3 25.6 38.4
39 50.4 36.2 54.2





#55

1,1,2-Trichloroethane

Concen: 20.931 ug/l

RT: 10.579 min Scan# 1453

Delta R.T. 0.000 min

Lab File: VY020703.D

Acq: 23 Dec 2024 19:47

Instrument :

MSVOA_Y

ClientSampleId :

VY1223SBS01

Tgt Ion: 97 Resp: 11682

Ion Ratio Lower Upper

97 100

83 84.5 63.1 94.7

85 50.8 39.5 59.3





99 73.6 46.4 69.6#

Manual Integrations

APPROVED

Reviewed By :Semsettin Yesilyurt 12/26/2024

Supervised By :Mahesh Dadoda 12/26/2024



#56

Ethyl methacrylate

Concen: 21.301 ug/l

RT: 10.445 min Scan# 1431

Delta R.T. 0.000 min

Lab File: VY020703.D

Acq: 23 Dec 2024 19:47




Tgt Ion: 69 Resp: 15165


Ion Ratio Lower Upper

69 100

41 56.1 43.8 65.6

39 42.5 32.0 48.0





#57

1,3-Dichloropropane

Concen: 20.893 ug/l

RT: 10.725 min Scan# 1477

Delta R.T. 0.000 min

Lab File: VY020703.D

Acq: 23 Dec 2024 19:47

Instrument :

MSVOA_Y

ClientSampleId :

VY1223SBSD01

Tgt Ion: 76 Resp: 19061

Ion Ratio Lower Upper

76 100





78 34.2 25.0 37.6

Manual Integrations

APPROVED

Reviewed By :Semsettin Yesilyurt 12/26/2024

Supervised By :Mahesh Dadoda 12/26/2024



#58

2-Chloroethyl Vinyl ether

Concen: 102.437 ug/l

RT: 9.719 min Scan# 1312

Delta R.T. 0.000 min

Lab File: VY020703.D




Acq: 23 Dec 2024 19:47


Tgt Ion: 63 Resp: 27253

Ion Ratio Lower Upper

63 100

106 35.3 29.1 43.7





#59

2-Hexanone

Concen: 104.066 ug/l

RT: 10.768 min Scan# 14

Delta R.T. 0.000 min

Lab File: VY020703.D

Acq: 23 Dec 2024 19:47

Instrument :

MSVOA_Y

ClientSampleId :

VY1223SBS01

Tgt Ion: 43 Resp: 24776

Ion Ratio Lower Upper

43 100





58 59.6 30.1 90.5

Manual Integrations

APPROVED

Reviewed By :Semsettin Yesilyurt 12/26/2024

Supervised By :Mahesh Dadoda 12/26/2024



#60

Dibromochloromethane

Concen: 21.495 ug/l

RT: 10.914 min Scan# 1508

Delta R.T. -0.006 min

Lab File: VY020703.D




Acq: 23 Dec 2024 19:47


Tgt Ion:129 Resp: 19154

Ion Ratio Lower Upper

129 100

127 80.0 38.5 115.5









#61
1,2-Dibromoethane
Concen: 21.127 ug/l
RT: 11.018 min Scan# 1526
Delta R.T. -0.006 min
Lab File: VY020703.D
Acq: 23 Dec 2024 19:47

Instrument :
MSVOA_Y
ClientSampleId :
VY1223SBS01

Tgt Ion: 107 Resp: 11054
Ion Ratio Lower Upper
107 100
109 94.7 75.5 113.3




Manual Integrations
APPROVED


Reviewed By : Semsettin Yesilyurt 12/26/2024
Supervised By : Mahesh Dadoda 12/26/2024



#62
4-Bromofluorobenzene
Concen: 52.934 ug/l
RT: 12.414 min Scan# 1754
Delta R.T. 0.000 min
Lab File: VY020703.D
Acq: 23 Dec 2024 19:47

Tgt Ion: 95 Resp: 63417
Ion Ratio Lower Upper
95 100
174 90.6 0.0 179.2
176 90.1 0.0 179.8









#63
Chlorobenzene-d5
Concen: 50.000 ug/l
RT: 11.420 min Scan# 1592
Delta R.T. -0.006 min
Lab File: VY020703.D
Acq: 23 Dec 2024 19:47

Instrument :
MSVOA_Y
ClientSampleId :
VY1223SBS01

Tgt Ion:117 Resp: 131936
Ion Ratio Lower Upper
117 100
82 53.1 43.4 65.2
119 32.0 26.6 39.8




Manual Integrations
APPROVED


Reviewed By :Semsettin Yesilyurt 12/26/2024
Supervised By :Mahesh Dadoda 12/26/2024



#64
Tetrachloroethene
Concen: 20.081 ug/l
RT: 10.652 min Scan# 1465
Delta R.T. 0.000 min
Lab File: VY020703.D
Acq: 23 Dec 2024 19:47

Tgt Ion:164 Resp: 17595
Ion Ratio Lower Upper
164 100
166 130.5 99.6 149.4
129 97.7 73.0 109.4
131 95.4 72.3 108.5





#65

Chlorobenzene

Concen: 21.338 ug/l

RT: 11.451 min Scan# 1596

Delta R.T. 0.000 min

Lab File: VY020703.D

Acq: 23 Dec 2024 19:47

Instrument :

MSVOA_Y

ClientSampleId :

VY1223SBS01

Tgt Ion:112 Resp: 52606

Ion Ratio Lower Upper

112 100





114 33.5 25.4 38.0

Manual Integrations

APPROVED

Reviewed By :Semsettin Yesilyurt 12/26/2024

Supervised By :Mahesh Dadoda 12/26/2024



#66

1,1,1,2-Tetrachloroethane

Concen: 20.778 ug/l

RT: 11.524 min Scan# 1608

Delta R.T. 0.000 min

Lab File: VY020703.D

Acq: 23 Dec 2024 19:47




Tgt Ion:131 Resp: 20004


Ion Ratio Lower Upper

131 100

133 99.8 48.6 145.9

119 60.1 28.4 85.3









#67
Ethyl Benzene
Concen: 20.628 ug/l
RT: 11.524 min Scan# 1609
Delta R.T. -0.006 min
Lab File: VY020703.D
Acq: 23 Dec 2024 19:47

Instrument :
MSVOA_Y
ClientSampleId :
VY1223SBS01

Tgt Ion: 91 Resp: 91241
Ion Ratio Lower Upper
91 100
106 33.2 24.5 36.7




Manual Integrations
APPROVED


Reviewed By :Semsettin Yesilyurt 12/26/2024
Supervised By :Mahesh Dadoda 12/26/2024



#68
m/p-Xylenes
Concen: 41.477 ug/l
RT: 11.633 min Scan# 1626
Delta R.T. 0.000 min
Lab File: VY020703.D
Acq: 23 Dec 2024 19:47

Tgt Ion:106 Resp: 69958
Ion Ratio Lower Upper
106 100
91 207.7 166.6 250.0









#69
o-Xylene
Concen: 21.509 ug/l
RT: 11.957 min Scan# 1680
Delta R.T. -0.006 min
Lab File: VY020703.D
Acq: 23 Dec 2024 19:47

Instrument :
MSVOA_Y
ClientSampleId :
VY1223SBS01

Tgt Ion:106 Resp: 34079
Ion Ratio Lower Upper
106 100
91 213.4 108.6 325.8




Manual Integrations
APPROVED


Reviewed By :Semsettin Yesilyurt 12/26/2024
Supervised By :Mahesh Dadoda 12/26/2024



#70
Styrene
Concen: 21.300 ug/l
RT: 11.975 min Scan# 1682
Delta R.T. 0.000 min
Lab File: VY020703.D
Acq: 23 Dec 2024 19:47

Tgt Ion:104 Resp: 56763
Ion Ratio Lower Upper
104 100
78 53.1 43.0 64.4
103 54.1 44.4 66.6





#71

Bromoform

Concen: 20.784 ug/l

RT: 12.139 min Scan# 1709

Delta R.T. 0.000 min

Lab File: VY020703.D

Acq: 23 Dec 2024 19:47

Instrument :

MSVOA_Y





ClientSampleId :

VY1223SBS01

Tgt Ion:	173	Resp:	11461
Ion Ratio	Lower	Upper	
173	100		
175	49.7	24.2	72.6
254	0.0	0.0	0.0

Manual Integrations
APPROVED

Reviewed By :Semsettin Yesilyurt 12/26/2024
Supervised By :Mahesh Dadoda 12/26/2024



#72

1,4-Dichlorobenzene-d4

Concen: 50.000 ug/l




RT: 13.353 min Scan# 1908


Delta R.T. 0.000 min

Lab File: VY020703.D

Acq: 23 Dec 2024 19:47

Tgt Ion:	152	Resp:	69344
Ion Ratio	Lower	Upper	
152	100		
115	58.2	30.3	90.9
150	163.4	0.0	349.0





#73

Isopropylbenzene

Concen: 19.757 ug/l

RT: 12.261 min Scan# 1729

Delta R.T. 0.000 min

Lab File: VY020703.D

Acq: 23 Dec 2024 19:47

Instrument :

MSVOA_Y

ClientSampleId :





VY1223SBS01

Tgt Ion	Ratio	Lower	Upper
105	100		
120	26.7	13.3	39.9

Manual Integrations

APPROVED

Reviewed By :Semsettin Yesilyurt 12/26/2024
Supervised By :Mahesh Dadoda 12/26/2024



#74

N-ethyl acetate

Concen: 20.358 ug/l




RT: 12.078 min Scan# 1699


Delta R.T. 0.000 min

Lab File: VY020703.D

Acq: 23 Dec 2024 19:47

Tgt Ion	Ratio	Lower	Upper
43	100		
70	52.3	39.7	59.5
55	26.8	21.5	32.3
61	27.5	21.4	32.2









#75
1,1,2,2-Tetrachloroethane
Concen: 20.545 ug/l
RT: 12.511 min Scan# 1770
Delta R.T. 0.000 min
Lab File: VY020703.D
Acq: 23 Dec 2024 19:47

Instrument :
MSVOA_Y
ClientSampleId :
VY1223SBSD01

Tgt Ion	83	Resp	12470
Ion	Ratio	Lower	Upper
83	100		
131	12.3	6.9	20.7
85	63.0	32.5	97.4




Manual Integrations
APPROVED


Reviewed By :Semsettin Yesilyurt 12/26/2024
Supervised By :Mahesh Dadoda 12/26/2024



#76
1,2,3-Trichloropropane
Concen: 20.804 ug/l m
RT: 12.560 min Scan# 1778
Delta R.T. 0.000 min
Lab File: VY020703.D
Acq: 23 Dec 2024 19:47


Tgt Ion	75	Resp	9949
Ion	Ratio	Lower	Upper
75	100		
77	0.0	0.0	0.0





#77
Bromobenzene
Concen: 19.737 ug/l
RT: 12.536 min Scan# 1774
Delta R.T. 0.000 min
Lab File: VY020703.D
Acq: 23 Dec 2024 19:47




Instrument :
MSVOA_Y
ClientSampleId :
VY1223SBSD01



Tgt Ion: 156 Resp: 21199
Ion Ratio Lower Upper
156 100
77 157.3 78.3 235.0
158 99.1 48.3 144.9




Manual Integrations
APPROVED


Reviewed By :Semsettin Yesilyurt 12/26/2024
Supervised By :Mahesh Dadoda 12/26/2024



#78
n-propylbenzene
Concen: 19.548 ug/l
RT: 12.603 min Scan# 1785
Delta R.T. 0.000 min
Lab File: VY020703.D
Acq: 23 Dec 2024 19:47

Tgt Ion: 91 Resp: 103950
Ion Ratio Lower Upper
91 100
120 25.2 12.3 36.8









#79
2-Chlorotoluene
Concen: 19.932 ug/l
RT: 12.688 min Scan# 1799
Delta R.T. 0.000 min
Lab File: VY020703.D
Acq: 23 Dec 2024 19:47

Instrument :
MSVOA_Y
ClientSampleId :
VY1223SBS01

Tgt Ion: 91 Resp: 60340
Ion Ratio Lower Upper
91 100
126 35.6 17.8 53.3




Manual Integrations
APPROVED


Reviewed By :Semsettin Yesilyurt 12/26/2024
Supervised By :Mahesh Dadoda 12/26/2024



#80
1,3,5-Trimethylbenzene
Concen: 20.379 ug/l
RT: 12.743 min Scan# 1808
Delta R.T. 0.000 min
Lab File: VY020703.D
Acq: 23 Dec 2024 19:47

Tgt Ion:105 Resp: 79726
Ion Ratio Lower Upper
105 100
120 47.5 23.9 71.7









#81
trans-1,4-Dichloro-2-butene
Concen: 18.628 ug/l
RT: 12.310 min Scan# 1737
Delta R.T. 0.000 min
Lab File: VY020703.D
Acq: 23 Dec 2024 19:47

Instrument :
MSVOA_Y
ClientSampleId :
VY1223SBSD01

Tgt Ion	Ratio	Lower	Upper
75	100		
53	105.6	83.1	124.7
89	54.9	40.6	61.0




Manual Integrations
APPROVED


Reviewed By :Semsettin Yesilyurt 12/26/2024
Supervised By :Mahesh Dadoda 12/26/2024



#82
4-Chlorotoluene
Concen: 20.303 ug/l
RT: 12.786 min Scan# 1815
Delta R.T. 0.000 min
Lab File: VY020703.D
Acq: 23 Dec 2024 19:47

Tgt Ion	Ratio	Lower	Upper
91	100		
126	35.2	17.8	53.5





#83

tert-Butylbenzene

Concen: 20.332 ug/l

RT: 13.005 min Scan# 1851

Delta R.T. 0.000 min

Lab File: VY020703.D

Acq: 23 Dec 2024 19:47

Instrument :

MSVOA_Y

ClientSampleId :





VY1223SBS01

Tgt Ion:	119	Resp:	74013
Ion Ratio	Lower	Upper	
119	100		
91	65.1	33.6	100.8
134	26.6	13.4	40.2

Manual Integrations

APPROVED

Reviewed By :Semsettin Yesilyurt 12/26/2024
Supervised By :Mahesh Dadoda 12/26/2024



#84

1,2,4-Trimethylbenzene

Concen: 20.842 ug/l




RT: 13.048 min Scan# 1858


Delta R.T. 0.000 min

Lab File: VY020703.D

Acq: 23 Dec 2024 19:47

Tgt Ion:	105	Resp:	79650
Ion Ratio	Lower	Upper	
105	100		
120	42.9	21.8	65.3









#85
sec-Butylbenzene
Concen: 20.123 ug/l
RT: 13.182 min Scan# 1880
Delta R.T. 0.000 min
Lab File: VY020703.D
Acq: 23 Dec 2024 19:47

Instrument :
MSVOA_Y
ClientSampleId :
VY1223SBSD01

Tgt Ion:105 Resp: 98630
Ion Ratio Lower Upper
105 100
134 21.8 10.7 32.1




Manual Integrations
APPROVED


Reviewed By :Semsettin Yesilyurt 12/26/2024
Supervised By :Mahesh Dadoda 12/26/2024



#86
p-Isopropyltoluene
Concen: 20.575 ug/l
RT: 13.298 min Scan# 1899
Delta R.T. 0.000 min
Lab File: VY020703.D
Acq: 23 Dec 2024 19:47

Tgt Ion:119 Resp: 88167
Ion Ratio Lower Upper
119 100
134 26.3 13.4 40.2
91 25.3 12.9 38.6





#87

1,3-Dichlorobenzene

Concen: 21.023 ug/l

RT: 13.292 min Scan# 1899

Delta R.T. -0.006 min

Lab File: VY020703.D

Acq: 23 Dec 2024 19:47

Instrument :

MSVOA_Y

ClientSampleId :





VY1223SBS01

Tgt	Ion:146	Resp:	43865
Ion	Ratio	Lower	Upper
146	100		
111	39.8	19.3	57.8
148	61.4	31.1	93.5

Manual Integrations

APPROVED

Reviewed By :Semsettin Yesilyurt 12/26/2024
Supervised By :Mahesh Dadoda 12/26/2024



#88

1,4-Dichlorobenzene

Concen: 20.707 ug/l




RT: 13.371 min Scan# 1911


Delta R.T. -0.006 min

Lab File: VY020703.D

Acq: 23 Dec 2024 19:47

Tgt	Ion:146	Resp:	42330
Ion	Ratio	Lower	Upper
146	100		
111	40.0	19.7	59.0
148	63.7	32.8	98.3









#89
n-Butylbenzene
Concen: 20.171 ug/l
RT: 13.621 min Scan# 1953
Delta R.T. -0.006 min
Lab File: VY020703.D
Acq: 23 Dec 2024 19:47

Instrument :
MSVOA_Y
ClientSampleId :
VY1223SBS01

Tgt Ion	Ratio	Lower	Upper
91	100		
92	52.1	26.1	78.2
134	28.7	14.8	44.4




Manual Integrations
APPROVED


Reviewed By :Semsettin Yesilyurt 12/26/2024
Supervised By :Mahesh Dadoda 12/26/2024



#90
Hexachloroethane
Concen: 19.561 ug/l
RT: 13.883 min Scan# 1995
Delta R.T. -0.006 min
Lab File: VY020703.D
Acq: 23 Dec 2024 19:47

Tgt Ion	Ratio	Lower	Upper
117	100		
201	94.8	44.9	134.7









#91
1,2-Dichlorobenzene
Concen: 20.807 ug/l
RT: 13.664 min Scan# 1960
Delta R.T. -0.006 min
Lab File: VY020703.D
Acq: 23 Dec 2024 19:47

Instrument :
MSVOA_Y
ClientSampleId :
VY1223SBS01

Tgt Ion:	146	Resp:	37259
Ion Ratio	Lower	Upper	
146	100		
111	40.5	20.3	60.8
148	64.6	31.9	95.5




Manual Integrations
APPROVED


Reviewed By :Semsettin Yesilyurt 12/26/2024
Supervised By :Mahesh Dadoda 12/26/2024



#92
1,2-Dibromo-3-Chloropropane
Concen: 20.414 ug/l
RT: 14.279 min Scan# 2060
Delta R.T. 0.000 min
Lab File: VY020703.D
Acq: 23 Dec 2024 19:47

Tgt Ion:	75	Resp:	2361
Ion Ratio	Lower	Upper	
75	100		
155	80.4	43.4	130.2
157	112.2	58.6	175.7





#93

1,2,4-Trichlorobenzene

Concen: 20.661 ug/l

RT: 14.932 min Scan# 2165

Delta R.T. 0.000 min

Lab File: VY020703.D

Acq: 23 Dec 2024 19:47

Instrument :

MSVOA_Y

ClientSampleId :





VY1223SBS01

Tgt Ion:	180	Resp:	21965
Ion Ratio	Lower	Upper	
180	100		
182	93.1	47.9	143.7
145	29.4	15.6	46.8

Manual Integrations

APPROVED

Reviewed By :Semsettin Yesilyurt 12/26/2024
Supervised By :Mahesh Dadoda 12/26/2024



#94

Hexachlorobutadiene

Concen: 20.697 ug/l




RT: 15.029 min Scan# 2183


Delta R.T. -0.006 min

Lab File: VY020703.D

Acq: 23 Dec 2024 19:47

Tgt Ion:	225	Resp:	16383
Ion Ratio	Lower	Upper	
225	100		
223	63.0	31.3	93.8
227	66.3	32.1	96.5





#95

Naphthalene

Concen: 19.827 ug/l

RT: 15.151 min Scan# 2203

Delta R.T. 0.000 min

Lab File: VY020703.D

Acq: 23 Dec 2024 19:47

Instrument :

MSVOA_Y

ClientSampleId :





VY1223SBS01

Tgt Ion:	128	Resp:	34561
Ion Ratio	Lower	Upper	
128	100		
127	13.2	10.7	16.1
129	10.6	8.6	12.8

Manual Integrations

APPROVED

Reviewed By :Semsettin Yesilyurt 12/26/2024
Supervised By :Mahesh Dadoda 12/26/2024



#96

1,2,3-Trichlorobenzene

Concen: 20.184 ug/l

RT: 15.340 min Scan# 2234

Delta R.T. 0.000 min

Lab File: VY020703.D

Acq: 23 Dec 2024 19:47

Tgt Ion:	180	Resp:	17939
Ion Ratio	Lower	Upper	
180	100		
182	95.1	47.4	142.2
145	32.3	16.7	50.1

