

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH Contract: NOBI03
 Lab Code: CHEM Case No.: P5306 SAS No.: P5306 SDG NO.: P5306
 Lab File ID: VY020664.D Date Analyzed: 12/20/2024
 Instrument ID: MSVOA_Y Time Analyzed: 09:09
 GC Column: RXI-624 ID: 0.25 (mm) Heated Purge: (Y/N) Y

	IS4 AREA #	RT #				
12 HOUR STD	119359	13.353				
UPPER LIMIT	238718	13.853				
LOWER LIMIT	59679.5	12.853				
EPA SAMPLE NO.						
OU4-VSL-08-121224	105748	13.35				
OU4-VSL-09-121224	109147	13.35				
OU4-VSL-10-121224	108630	13.35				
OU4-VSL-11-121224RE	10185 *	13.35				
OU4-VSL-13-121224RE	1297 *	13.35				
VY1220SBL01	109647	13.35				
VY1220SBS01	116933	13.35				

IS4 = 1,4-Dichlorobenzene-d4

AREA UPPER LIMIT = +100% of internal standard area
 AREA LOWER LIMIT = -50% of internal standard area
 RT UPPER LIMIT = +0.50 minutes of internal standard RT
 RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.
 * Values outside of QC limits.