

Instrument ID: MSVOA_Y

Daily Analysis Runlog For Sequence/QC Batch ID # VY121724

Review By	Maresh Dadoda	Review On	12/18/2024 8:38:34 AM
Supervise By	Semsettin Yesilyurt	Supervise On	12/18/2024 8:51:04 AM
SubDirectory	VY121724	HP Acquire Method	HP Processing Method 82y121724s.m

STD. NAME	STD REF.#
Tune/Reschk	VP132182
Initial Calibration Stds	VP132183,VP132184,VP132185,VP132186,VP132187,VP132188
CCC	
Internal Standard/PEM	VP131783
ICV/I.BLK	VP132189
Surrogate Standard	
MS/MSD Standard	
LCS Standard	

Sr#	Sampleld	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	BFB	BFB	VY020611.D	17 Dec 2024 09:00		SY/MD	Ok
2	VSTDIC005	VSTDIC005	VY020612.D	17 Dec 2024 09:38		SY/MD	Ok
3	VSTDIC010	VSTDIC010	VY020613.D	17 Dec 2024 10:01	Com #16,30 on LR	SY/MD	Ok,M
4	VSTDIC020	VSTDIC020	VY020614.D	17 Dec 2024 10:23	Com #11 on QR	SY/MD	Ok,M
5	VSTDIC050	VSTDIC050	VY020615.D	17 Dec 2024 10:51	Com #13 fail for method	SY/MD	Ok,M
6	VSTDIC100	VSTDIC100	VY020616.D	17 Dec 2024 11:14		SY/MD	Ok,M
7	VSTDIC150	VSTDIC150	VY020617.D	17 Dec 2024 11:37		SY/MD	Ok,M
8	VIBLK	VIBLK	VY020618.D	17 Dec 2024 12:24		SY/MD	Ok
9	VSTDIC005	VSTDIC005	VY020619.D	17 Dec 2024 14:51	%D fail for com #11	SY/MD	Ok,M
10	VSTDICV050	VSTDICV050	VY020620.D	17 Dec 2024 17:41	ICV fail for com #11,16	SY/MD	Not Ok

M : Manual Integration