

Instrument ID: MSVOA_Y

Daily Analysis Runlog For Sequence/QC Batch ID # VY121924

Review By	Maresh Dadoda	Review On	12/20/2024 10:52:52 AM
Supervise By	Semsettin Yesilyurt	Supervise On	12/20/2024 10:57:22 AM
SubDirectory	VY121924	HP Acquire Method	HP Processing Method 82y121724s.m

STD. NAME	STD REF.#
Tune/Reschk Initial Calibration Stds	VP132238
CCC Internal Standard/PEM ICV/I.BLK Surrogate Standard MS/MSD Standard LCS Standard	VP132239,VP132240 VP131783

Sr#	SampleID	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	BFB	BFB	VY020644.D	19 Dec 2024 09:13		SY/MD	Ok
2	VSTDCCC050	VSTDCCC050	VY020645.D	19 Dec 2024 09:43		SY/MD	Ok,M
3	VY1219SBL01	VY1219SBL01	VY020646.D	19 Dec 2024 11:07		SY/MD	Ok
4	VY1219SBS01	VY1219SBS01	VY020647.D	19 Dec 2024 11:53	BS low for com #12	SY/MD	Ok,M
5	VY1219SBSD01	VY1219SBSD01	VY020648.D	19 Dec 2024 12:16		SY/MD	Ok,M
6	P5312-03RE	CONCRETE-VNJ-222RE	VY020649.D	19 Dec 2024 12:52	vial- B BS low; Internal standard fail	SY/MD	Confirms
7	P5312-01RE	SOIL-VNJ-222RE	VY020650.D	19 Dec 2024 13:15	vial- B BS low; Internal standard fail	SY/MD	Confirms
8	P5301-03RE	HR-04-121624RE	VY020651.D	19 Dec 2024 13:39	vial- B BS low; ISTD Fail	SY/MD	Confirms
9	P5306-01	OU4-VSL-07-121224	VY020652.D	19 Dec 2024 14:02	vial- B End CCC fail	SY/MD	Ok
10	P5306-03	OU4-VSL-08-121224	VY020653.D	19 Dec 2024 14:26	vial- A End CCC fail; Not purge	SY/MD	Not Ok
11	P5306-05	OU4-VSL-09-121224	VY020654.D	19 Dec 2024 14:49	vial- A End CCC fail; Surrogate fail	SY/MD	Not Ok
12	P5306-07	OU4-VSL-10-121224	VY020655.D	19 Dec 2024 15:12	vial- A End CCC fail; Internal standard fail; Surrogate fail	SY/MD	Not Ok
13	P5306-09	OU4-VSL-11-121224	VY020656.D	19 Dec 2024 15:36	vial- A End CCC fail; Surrogate fail	SY/MD	ReRun
14	P5306-11	OU4-VSL-12-121224	VY020657.D	19 Dec 2024 15:59	vial- A End CCC fail; Surrogate fail	SY/MD	Not Ok
15	P5306-13	OU4-VSL-13-121224	VY020658.D	19 Dec 2024 16:23	vial- A End CCC fail; Surrogate fail	SY/MD	ReRun
16	P5343-01	VNJ210	VY020659.D	19 Dec 2024 16:46	vial- A Internal standard fail	SY/MD	Ok

Instrument ID: MSVOA_Y

Daily Analysis Runlog For Sequence/QC Batch ID # VY121924

Review By	Maresh Dadoda	Review On	12/20/2024 10:52:52 AM
Supervise By	Semsettin Yesilyurt	Supervise On	12/20/2024 10:57:22 AM
SubDirectory	VY121924	HP Acquire Method	HP Processing Method 82y121724s.m
STD. NAME	STD REF.#		
Tune/Reschk Initial Calibration Stds	VP132238		
CCC	VP132239,VP132240		
Internal Standard/PEM	VP131783		
ICV/I.BLK			
Surrogate Standard			
MS/MSD Standard			
LCS Standard			

17	P5277-01	COMP-1A	VY020660.D	19 Dec 2024 17:09	vial- B	SY/MD	Ok
18	P5343-03	VNJ281	VY020661.D	19 Dec 2024 17:33		SY/MD	Ok
19	VSTDCCC050	VSTDCCC050EC	VY020662.D	19 Dec 2024 17:56	CCC fail for many comp.	SY/MD	Ok,M

M : Manual Integration