

Instrument ID: MSVOA\_Y

**Daily Analysis Runlog For Sequence/QC Batch ID # VY122024**

|              |                     |                   |                                             |
|--------------|---------------------|-------------------|---------------------------------------------|
| Review By    | Semsettin Yesilyurt | Review On         | 12/23/2024 7:25:27 AM                       |
| Supervise By | Maresh Dadoda       | Supervise On      | 12/24/2024 12:28:45 AM                      |
| SubDirectory | VY122024            | HP Acquire Method | MSVOA_Y HP Processing Method 82y12022024s.m |

| STD. NAME                                                                                          | STD REF.#                     |
|----------------------------------------------------------------------------------------------------|-------------------------------|
| Tune/Reschk<br>Initial Calibration Stds                                                            | VP132264                      |
| CCC<br>Internal Standard/PEM<br>ICV/I.BLK<br>Surrogate Standard<br>MS/MSD Standard<br>LCS Standard | VP132265,VP132266<br>VP131783 |

| Sr# | SampleID     | ClientID            | Data File Name | Date-Time         | Comment                                      | Operator | Status   |
|-----|--------------|---------------------|----------------|-------------------|----------------------------------------------|----------|----------|
| 1   | BFB          | BFB                 | VY020663.D     | 20 Dec 2024 08:37 |                                              | SY/MD    | Ok       |
| 2   | VSTDCCC050   | VSTDCCC050          | VY020664.D     | 20 Dec 2024 09:09 |                                              | SY/MD    | Ok,M     |
| 3   | VY1220SBL01  | VY1220SBL01         | VY020665.D     | 20 Dec 2024 11:24 |                                              | SY/MD    | Ok       |
| 4   | VY1220SBS01  | VY1220SBS01         | VY020666.D     | 20 Dec 2024 13:12 |                                              | SY/MD    | Ok,M     |
| 5   | VY1220SBSD01 | VY1220SBSD01        | VY020667.D     | 20 Dec 2024 13:34 |                                              | SY/MD    | Ok,M     |
| 6   | P5279-03     | ROCKAWAY-PARK       | VY020668.D     | 20 Dec 2024 14:10 | vial-B                                       | SY/MD    | Ok       |
| 7   | P5316-01     | TT-304-IDWSO-202412 | VY020669.D     | 20 Dec 2024 14:34 | vial-A Surrogate Fail                        | SY/MD    | ReRun    |
| 8   | P5316-02     | TT-304-IDWSO-202412 | VY020670.D     | 20 Dec 2024 14:57 | vial-A Surrogate Fail                        | SY/MD    | ReRun    |
| 9   | P5316-03     | TT-304-IDWSO-202412 | VY020671.D     | 20 Dec 2024 15:21 | vial-A Surrogate Fail                        | SY/MD    | ReRun    |
| 10  | P5316-04     | TT-304-IDWSO-202412 | VY020672.D     | 20 Dec 2024 15:44 | vial-A Internal Standard and Surrogate Fail  | SY/MD    | ReRun    |
| 11  | P5343-03     | VNJ281              | VY020673.D     | 20 Dec 2024 16:07 | vial-A Not purged                            | SY/MD    | Not Ok   |
| 12  | P5343-01     | VNJ210              | VY020674.D     | 20 Dec 2024 16:31 | vial-B Not purged                            | SY/MD    | Not Ok   |
| 13  | P5316-04RE   | TT-304-IDWSO-202412 | VY020675.D     | 20 Dec 2024 16:52 | vial-B Surrogate Fail                        | SY/MD    | Confirms |
| 14  | P5306-01     | OU4-VSL-07-121224   | VY020676.D     | 20 Dec 2024 17:16 | vial-B NOT PURGE                             | SY/MD    | Not Ok   |
| 15  | P5306-03     | OU4-VSL-08-121224   | VY020677.D     | 20 Dec 2024 17:39 | vial-B Surrogate Fail                        | SY/MD    | Ok       |
| 16  | P5306-05     | OU4-VSL-09-121224   | VY020678.D     | 20 Dec 2024 18:02 | vial-B                                       | SY/MD    | Ok       |
| 17  | P5306-13RE   | OU4-VSL-13-121224RE | VY020679.D     | 20 Dec 2024 18:26 | vial-B Surrogate Fail;Internal Standard Fail | SY/MD    | Confirms |

Instrument ID: MSVOA\_Y

**Daily Analysis Runlog For Sequence/QC Batch ID # VY122024**

| Review By                               | Semsettin Yesilyurt | Review On         | 12/23/2024 7:25:27 AM  |                      |                |
|-----------------------------------------|---------------------|-------------------|------------------------|----------------------|----------------|
| Supervise By                            | Mahesh Dadoda       | Supervise On      | 12/24/2024 12:28:45 AM |                      |                |
| SubDirectory                            | VY122024            | HP Acquire Method | MSVOA_Y                | HP Processing Method | 82y12022024s.m |
| STD. NAME                               |                     | STD REF.#         |                        |                      |                |
| Tune/Reschk<br>Initial Calibration Stds |                     | VP132264          |                        |                      |                |
| CCC                                     |                     | VP132265,VP132266 |                        |                      |                |
| Internal Standard/PEM                   |                     | VP131783          |                        |                      |                |
| ICV/I.BLK                               |                     |                   |                        |                      |                |
| Surrogate Standard                      |                     |                   |                        |                      |                |
| MS/MSD Standard                         |                     |                   |                        |                      |                |
| LCS Standard                            |                     |                   |                        |                      |                |

|    |            |                     |            |                   |                                              |       |          |
|----|------------|---------------------|------------|-------------------|----------------------------------------------|-------|----------|
| 18 | P5306-07   | OU4-VSL-10-121224   | VY020680.D | 20 Dec 2024 18:49 | vial-B                                       | SY/MD | Ok       |
| 19 | P5306-09RE | OU4-VSL-11-121224RE | VY020681.D | 20 Dec 2024 19:13 | vial-B Surrogate Fail;Internal Standard Fail | SY/MD | Confirms |
| 20 | VSTDCCC050 | VSTDCCC050EC        | VY020682.D | 20 Dec 2024 19:35 |                                              | SY/MD | Ok,M     |

M : Manual Integration