

CALIBRATION VERIFICATION SUMMARY

Contract: ATCE02

Lab Code: CHEM Case No.: P5307 SAS No.: P5307 SDG NO.: P5307

Continuing Calib Date: 12/19/2024 Initial Calibration Date(s): 12/06/2024 12/06/2024

Continuing Calib Time: 04:30 Initial Calibration Time(s): 14:19 22:34

GC Column: ZB-MR1 ID: 0.32 (mm)

COMPOUND		CCAL RT	AVG RT	RT WINDOW		DIFF RT
				FROM	TO	
Aroclor-1242-1	(1)	4.80	4.81	4.71	4.91	0.01
Aroclor-1242-2	(2)	4.82	4.83	4.73	4.93	0.01
Aroclor-1242-3	(3)	4.88	4.88	4.78	4.98	0.00
Aroclor-1242-4	(4)	5.00	5.01	4.91	5.11	0.01
Aroclor-1242-5	(5)	5.65	5.66	5.56	5.76	0.01
Tetrachloro-m-xylene		3.71	3.71	3.61	3.81	0.00
Decachlorobiphenyl		8.77	8.79	8.69	8.89	0.02

CALIBRATION VERIFICATION SUMMARY

Contract: ATCE02

Lab Code: CHEM Case No.: P5307 SAS No.: P5307 SDG NO.: P5307

Continuing Calib Date: 12/19/2024 Initial Calibration Date(s): 12/06/2024 12/06/2024

Continuing Calib Time: 04:30 Initial Calibration Time(s): 14:19 22:34

GC Column: ZB-MR2 ID: 0.32 (mm)

COMPOUND		CCAL RT	AVG RT	RT WINDOW		DIFF RT
				FROM	TO	
Aroclor-1242-1	(1)	4.79	4.80	4.70	4.90	0.01
Aroclor-1242-2	(2)	4.81	4.82	4.72	4.92	0.01
Aroclor-1242-3	(3)	4.99	4.99	4.89	5.09	0.00
Aroclor-1242-4	(4)	5.07	5.08	4.98	5.18	0.01
Aroclor-1242-5	(5)	5.59	5.60	5.50	5.70	0.01
Tetrachloro-m-xylene		3.70	3.71	3.61	3.81	0.01
Decachlorobiphenyl		8.72	8.74	8.64	8.84	0.02