

CALIBRATION VERIFICATION SUMMARY

Contract: ATCE02

Lab Code: CHEM Case No.: P5307 SAS No.: P5307 SDG NO.: P5307

GC Column: ZB-MR1 ID: 0.32 (mm) Initi. Calib. Date(s): 12/06/2024 12/06/2024

Client Sample No.: CCAL01 Date Analyzed: 12/17/2024

Lab Sample No.: AR1660CCC500 Data File : PO108586.D Time Analyzed: 15:44

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	4.803	4.709	4.909	495.840	500.000	-0.8
Aroclor-1016-2	4.824	4.729	4.929	503.310	500.000	0.7
Aroclor-1016-3	4.879	4.785	4.985	500.100	500.000	0.0
Aroclor-1016-4	5.001	4.907	5.107	499.840	500.000	0.0
Aroclor-1016-5	5.258	5.165	5.365	490.250	500.000	-2.0
Aroclor-1260-1	6.301	6.210	6.410	472.980	500.000	-5.4
Aroclor-1260-2	6.489	6.398	6.598	459.400	500.000	-8.1
Aroclor-1260-3	6.859	6.769	6.969	427.540	500.000	-14.5
Aroclor-1260-4	7.120	7.029	7.229	412.970	500.000	-17.4
Aroclor-1260-5	7.361	7.270	7.470	419.680	500.000	-16.1
Decachlorobiphenyl	8.775	8.691	8.891	40.270	50.000	-19.5
Tetrachloro-m-xylene	3.707	3.610	3.810	50.400	50.000	0.8

CALIBRATION VERIFICATION SUMMARY

Contract: ATCE02

Lab Code: CHEM Case No.: P5307 SAS No.: P5307 SDG NO.: P5307

GC Column: ZB-MR2 ID: 0.32 (mm) Initi. Calib. Date(s): 12/06/2024 12/06/2024

Client Sample No.: CCAL01 Date Analyzed: 12/17/2024

Lab Sample No.: AR1660CCC500 Data File : PO108586.D Time Analyzed: 15:44

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	4.791	4.698	4.898	559.110	500.000	11.8
Aroclor-1016-2	4.810	4.718	4.918	563.680	500.000	12.7
Aroclor-1016-3	4.986	4.894	5.094	555.610	500.000	11.1
Aroclor-1016-4	5.027	4.935	5.135	526.750	500.000	5.4
Aroclor-1016-5	5.242	5.150	5.350	573.850	500.000	14.8
Aroclor-1260-1	6.276	6.186	6.386	526.370	500.000	5.3
Aroclor-1260-2	6.462	6.373	6.573	519.600	500.000	3.9
Aroclor-1260-3	6.616	6.527	6.727	510.490	500.000	2.1
Aroclor-1260-4	7.088	7.000	7.200	516.890	500.000	3.4
Aroclor-1260-5	7.328	7.239	7.439	520.890	500.000	4.2
Decachlorobiphenyl	8.724	8.641	8.841	48.600	50.000	-2.8
Tetrachloro-m-xylene	3.704	3.608	3.808	55.660	50.000	11.3