

CALIBRATION VERIFICATION SUMMARY

Contract: ATCE02

Lab Code: CHEM Case No.: P5307 SAS No.: P5307 SDG NO.: P5307

GC Column: ZB-MR1 ID: 0.32 (mm) Initi. Calib. Date(s): 12/06/2024 12/06/2024

Client Sample No.: CCAL02 Date Analyzed: 12/17/2024

Lab Sample No.: AR1660CCC500 Data File : PO108605.D Time Analyzed: 22:15

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	4.803	4.709	4.909	487.400	500.000	-2.5
Aroclor-1016-2	4.823	4.729	4.929	497.760	500.000	-0.4
Aroclor-1016-3	4.878	4.785	4.985	491.990	500.000	-1.6
Aroclor-1016-4	5.000	4.907	5.107	497.790	500.000	-0.4
Aroclor-1016-5	5.258	5.165	5.365	509.060	500.000	1.8
Aroclor-1260-1	6.300	6.210	6.410	462.050	500.000	-7.6
Aroclor-1260-2	6.489	6.398	6.598	459.330	500.000	-8.1
Aroclor-1260-3	6.858	6.769	6.969	458.520	500.000	-8.3
Aroclor-1260-4	7.119	7.029	7.229	434.850	500.000	-13.0
Aroclor-1260-5	7.359	7.270	7.470	432.630	500.000	-13.5
Decachlorobiphenyl	8.773	8.691	8.891	41.830	50.000	-16.3
Tetrachloro-m-xylene	3.706	3.610	3.810	49.770	50.000	-0.5

CALIBRATION VERIFICATION SUMMARY

Contract: ATCE02

Lab Code: CHEM Case No.: P5307 SAS No.: P5307 SDG NO.: P5307

GC Column: ZB-MR2 ID: 0.32 (mm) Initi. Calib. Date(s): 12/06/2024 12/06/2024

Client Sample No.: CCAL02 Date Analyzed: 12/17/2024

Lab Sample No.: AR1660CCC500 Data File : PO108605.D Time Analyzed: 22:15

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	4.790	4.698	4.898	563.720	500.000	12.7
Aroclor-1016-2	4.810	4.718	4.918	564.290	500.000	12.9
Aroclor-1016-3	4.986	4.894	5.094	554.250	500.000	10.9
Aroclor-1016-4	5.027	4.935	5.135	517.490	500.000	3.5
Aroclor-1016-5	5.242	5.150	5.350	587.380	500.000	17.5
Aroclor-1260-1	6.276	6.186	6.386	556.620	500.000	11.3
Aroclor-1260-2	6.462	6.373	6.573	556.630	500.000	11.3
Aroclor-1260-3	6.616	6.527	6.727	552.090	500.000	10.4
Aroclor-1260-4	7.088	7.000	7.200	556.140	500.000	11.2
Aroclor-1260-5	7.327	7.239	7.439	556.990	500.000	11.4
Decachlorobiphenyl	8.725	8.641	8.841	50.700	50.000	1.4
Tetrachloro-m-xylene	3.704	3.608	3.808	55.950	50.000	11.9