

CALIBRATION VERIFICATION SUMMARY

Contract: ATCE02

Lab Code: CHEM Case No.: P5307 SAS No.: P5307 SDG NO.: P5307

GC Column: ZB-MR1 ID: 0.32 (mm) Initi. Calib. Date(s): 12/06/2024 12/06/2024

Client Sample No.: CCAL03 Date Analyzed: 12/18/2024

Lab Sample No.: AR1660CCC500 Data File : PO108608.D Time Analyzed: 09:44

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	4.804	4.709	4.909	515.210	500.000	3.0
Aroclor-1016-2	4.823	4.729	4.929	515.890	500.000	3.2
Aroclor-1016-3	4.879	4.785	4.985	516.890	500.000	3.4
Aroclor-1016-4	5.001	4.907	5.107	517.750	500.000	3.6
Aroclor-1016-5	5.258	5.165	5.365	508.190	500.000	1.6
Aroclor-1260-1	6.302	6.210	6.410	483.180	500.000	-3.4
Aroclor-1260-2	6.490	6.398	6.598	481.360	500.000	-3.7
Aroclor-1260-3	6.860	6.769	6.969	472.890	500.000	-5.4
Aroclor-1260-4	7.120	7.029	7.229	466.740	500.000	-6.7
Aroclor-1260-5	7.361	7.270	7.470	458.880	500.000	-8.2
Decachlorobiphenyl	8.776	8.691	8.891	44.440	50.000	-11.1
Tetrachloro-m-xylene	3.706	3.610	3.810	53.080	50.000	6.2

CALIBRATION VERIFICATION SUMMARY

Contract: ATCE02

Lab Code: CHEM Case No.: P5307 SAS No.: P5307 SDG NO.: P5307

GC Column: ZB-MR2 ID: 0.32 (mm) Initi. Calib. Date(s): 12/06/2024 12/06/2024

Client Sample No.: CCAL03 Date Analyzed: 12/18/2024

Lab Sample No.: AR1660CCC500 Data File : PO108608.D Time Analyzed: 09:44

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
Aroclor-1016-1	4.790	4.698	4.898	536.850	500.000	7.4
Aroclor-1016-2	4.809	4.718	4.918	537.110	500.000	7.4
Aroclor-1016-3	4.985	4.894	5.094	529.270	500.000	5.9
Aroclor-1016-4	5.026	4.935	5.135	527.460	500.000	5.5
Aroclor-1016-5	5.240	5.150	5.350	526.070	500.000	5.2
Aroclor-1260-1	6.275	6.186	6.386	529.960	500.000	6.0
Aroclor-1260-2	6.462	6.373	6.573	531.840	500.000	6.4
Aroclor-1260-3	6.615	6.527	6.727	522.150	500.000	4.4
Aroclor-1260-4	7.088	7.000	7.200	534.260	500.000	6.9
Aroclor-1260-5	7.327	7.239	7.439	539.730	500.000	7.9
Decachlorobiphenyl	8.723	8.641	8.841	49.980	50.000	0.0
Tetrachloro-m-xylene	3.703	3.608	3.808	53.730	50.000	7.5