

Surrogate Summary

SDG No.: **P5307**

Client: **ATC Group Services LLC**

Analytical Method: **8082A**

Lab Sample ID	Client ID	Parameter	Column	Spike	Result	Rec	Qual	Limits	
								Low	High
I.BLK-PO108361.D	PIBLK-PO108361.D	Tetrachloro-m-xylene	1	20	23.0	115		60	140
		Decachlorobiphenyl	1	20	23.6	118		60	140
		Tetrachloro-m-xylene	2	20	21.6	108		60	140
		Decachlorobiphenyl	2	20	23.6	118		60	140
I.BLK-PO108587.D	PIBLK-PO108587.D	Tetrachloro-m-xylene	1	20	19.2	96		60	140
		Decachlorobiphenyl	1	20	16.9	84		60	140
		Tetrachloro-m-xylene	2	20	20.4	102		60	140
		Decachlorobiphenyl	2	20	20.3	101		60	140
PB165696BL	PB165696BL	Tetrachloro-m-xylene	1	20	18.7	93		32	144
		Decachlorobiphenyl	1	20	17.0	85		32	175
		Tetrachloro-m-xylene	2	20	19.5	98		32	144
		Decachlorobiphenyl	2	20	20.5	102		32	175
PB165696BS	PB165696BS	Tetrachloro-m-xylene	1	20	17.1	86		32	144
		Decachlorobiphenyl	1	20	17.2	86		32	175
		Tetrachloro-m-xylene	2	20	18.0	90		32	144
		Decachlorobiphenyl	2	20	20.8	104		32	175
P5307-01	1A-1B-1C-ROOF-2	Tetrachloro-m-xylene	1	20	20.1	100		32	144
		Decachlorobiphenyl	1	20	12.5	62		32	175
		Tetrachloro-m-xylene	2	20	21.6	108		32	144
		Decachlorobiphenyl	2	20	16.9	85		32	175
P5307-03	3A-3B-3C-1907	Tetrachloro-m-xylene	1	20	18.2	91		32	144
		Decachlorobiphenyl	1	20	14.8	74		32	175
		Tetrachloro-m-xylene	2	20	20.8	104		32	144
		Decachlorobiphenyl	2	20	18.3	91		32	175
P5307-04	4A-4B-4C-1907	Tetrachloro-m-xylene	1	20	19.2	96		32	144
		Decachlorobiphenyl	1	20	15.3	76		32	175
		Tetrachloro-m-xylene	2	20	20.7	103		32	144
		Decachlorobiphenyl	2	20	19.0	95		32	175
P5307-05	5A-5B-5C-1907	Tetrachloro-m-xylene	1	20	18.1	91		32	144
		Decachlorobiphenyl	1	20	15.5	78		32	175
		Tetrachloro-m-xylene	2	20	19.7	98		32	144
		Decachlorobiphenyl	2	20	19.5	97		32	175
P5307-06	6A-6B-6C-1952	Tetrachloro-m-xylene	1	20	18.9	94		32	144
		Decachlorobiphenyl	1	20	15.7	78		32	175
		Tetrachloro-m-xylene	2	20	19.8	99		32	144
		Decachlorobiphenyl	2	20	20.6	103		32	175
P5307-07	1907-BLDG-GRAY	Tetrachloro-m-xylene	1	20	16.2	81		32	144
		Decachlorobiphenyl	1	20	15.5	78		32	175
		Tetrachloro-m-xylene	2	20	17.5	87		32	144
		Decachlorobiphenyl	2	20	18.6	93		32	175
P5307-09	9A-9B-9C-1907	Tetrachloro-m-xylene	1	20	17.4	87		32	144

Surrogate Summary

SDG No.: **P5307**

Client: **ATC Group Services LLC**

Analytical Method: **8082A**

Lab Sample ID	Client ID	Parameter	Column	Spike	Result	Rec	Qual	Limits	
								Low	High
P5307-09	9A-9B-9C-1907	Decachlorobiphenyl	1	20	13.0	65		32	175
		Tetrachloro-m-xylene	2	20	19.1	95		32	144
		Decachlorobiphenyl	2	20	16.1	81		32	175
P5307-12	12A-12B-12C-1952	Tetrachloro-m-xylene	1	20	15.8	79		32	144
		Decachlorobiphenyl	1	20	16.0	80		32	175
		Tetrachloro-m-xylene	2	20	18.0	90		32	144
		Decachlorobiphenyl	2	20	20.2	101		32	175
P5307-13	13A-13B-13C-1952	Tetrachloro-m-xylene	1	20	19.3	97		32	144
		Decachlorobiphenyl	1	20	17.0	85		32	175
		Tetrachloro-m-xylene	2	20	21.0	105		32	144
		Decachlorobiphenyl	2	20	21.3	106		32	175
P5307-14	14A-14B-14C-1907	Tetrachloro-m-xylene	1	20	17.1	86		32	144
		Decachlorobiphenyl	1	20	14.7	73		32	175
		Tetrachloro-m-xylene	2	20	18.7	94		32	144
		Decachlorobiphenyl	2	20	18.7	94		32	175
P5307-15	15A-15B-15C-ROOF-7	Tetrachloro-m-xylene	1	20	19.1	96		32	144
		Decachlorobiphenyl	1	20	15.9	80		32	175
		Tetrachloro-m-xylene	2	20	20.8	104		32	144
		Decachlorobiphenyl	2	20	20.6	103		32	175
I.BLK-PO108606.D	PIBLK-PO108606.D	Tetrachloro-m-xylene	1	20	18.9	94		60	140
		Decachlorobiphenyl	1	20	17.4	87		60	140
		Tetrachloro-m-xylene	2	20	20.3	102		60	140
		Decachlorobiphenyl	2	20	21.1	106		60	140
I.BLK-PO108611.D	PIBLK-PO108611.D	Tetrachloro-m-xylene	1	20	20.5	103		60	140
		Decachlorobiphenyl	1	20	19.2	96		60	140
		Tetrachloro-m-xylene	2	20	20.6	103		60	140
		Decachlorobiphenyl	2	20	22.0	110		60	140
P5307-02	2A-2B-2C-ROOF-2	Tetrachloro-m-xylene	1	20	19.8	99		32	144
		Decachlorobiphenyl	1	20	19.2	96		32	175
		Tetrachloro-m-xylene	2	20	20.5	102		32	144
		Decachlorobiphenyl	2	20	21.4	107		32	175
P5307-08	1952-BLDG	Tetrachloro-m-xylene	1	20	17.7	88		32	144
		Decachlorobiphenyl	1	20	16.1	80		32	175
		Tetrachloro-m-xylene	2	20	18.3	91		32	144
		Decachlorobiphenyl	2	20	18.4	92		32	175
P5307-10	1907-BLDG-OFF-WHITE	Tetrachloro-m-xylene	1	20	17.2	86		32	144
		Decachlorobiphenyl	1	20	11.3	57		32	175
		Tetrachloro-m-xylene	2	20	18.0	90		32	144
		Decachlorobiphenyl	2	20	13.8	69		32	175
P5307-11	11A-11B-11C-1952-BLDG	Tetrachloro-m-xylene	1	20	15.1	76		32	144
		Decachlorobiphenyl	1	20	16.5	82		32	175

Surrogate Summary

SDG No.: P5307

Client: ATC Group Services LLC

Analytical Method: 8082A

Lab Sample ID	Client ID	Parameter	Column	Spike	Result	Rec	Qual	Limits	
								Low	High
P5307-11	11A-11B-11C-1952-BLDG	Tetrachloro-m-xylene	2	20	17.7	89		32	144
		Decachlorobiphenyl	2	20	19.7	98		32	175
I.BLK-PO108624.D	PIBLK-PO108624.D	Tetrachloro-m-xylene	1	20	20.4	102		60	140
		Decachlorobiphenyl	1	20	20.2	101		60	140
		Tetrachloro-m-xylene	2	20	21.3	107		60	140
		Decachlorobiphenyl	2	20	23.4	117		60	140