

Report of Analysis

Client:	Tetra Tech NUS, Inc.	Date Collected:	12/13/24
Project:	CTO WE13	Date Received:	12/17/24
Client Sample ID:	BP-VPB-190A-GW-678-680	SDG No.:	P5314
Lab Sample ID:	P5314-02	Matrix:	Water
Analytical Method:	SW8270ESIM	% Solid:	0
Sample Wt/Vol:	800 Units: mL	Final Vol:	1000 uL
Soil Aliquot Vol:	uL	Test:	SVOC-SIMGroup1
Extraction Type :	Decanted : N	Level :	LOW
Injection Volume :	GPC Factor : 1.0	GPC Cleanup :	N PH :
Prep Method :			

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
BN035785.D	1	12/18/24 12:50	12/23/24 15:41	PB165724

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units
TARGETS							
123-91-1	1,4-Dioxane	1.40		0.090	0.25	0.25	ug/L
SURROGATES							
7297-45-2	2-Methylnaphthalene-d10	0.39		30 - 150		96%	SPK: 0.4
93951-69-0	Fluoranthene-d10	0.41		30 - 150		101%	SPK: 0.4
4165-60-0	Nitrobenzene-d5	0.41		55 - 111		102%	SPK: 0.4
321-60-8	2-Fluorobiphenyl	0.39		53 - 106		97%	SPK: 0.4
1718-51-0	Terphenyl-d14	0.56	*	58 - 132		140%	SPK: 0.4
INTERNAL STANDARDS							
3855-82-1	1,4-Dichlorobenzene-d4	3450	7.264				
1146-65-2	Naphthalene-d8	8950	10.009				
15067-26-2	Acenaphthene-d10	5550	13.924				
1517-22-2	Phenanthrene-d10	10300	16.698				
1719-03-5	Chrysene-d12	8360	20.947				
1520-96-3	Perylene-d12	8390	23.032				

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

() = Laboratory InHouse Limit

A = Aldol-Condensation Reaction Products