

8B

SEMIVOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH

Lab Code: CHEM Case No.: P5314 SAS No.: P5314 SDG NO.: P5314

EPA Sample No.: SSTDCCC0.4 Date Analyzed: 12/23/2024

Lab File ID: BN035776.D Time Analyzed: 10:22

Instrument ID: BNA_N GC Column: ZB-GR ID: 0.25 (mm)

	IS1 (DCB) AREA #	RT #	IS2 (NPT) AREA #	RT #	IS3 (ANT) AREA #	RT #
12 HOUR STD	3884	7.264	10099	10.01	5893	13.92
UPPER LIMIT	7768	7.764	20198	10.509	11786	14.424
LOWER LIMIT	1942	6.764	5049.5	9.509	2946.5	13.424
EPA SAMPLE NO.						
01 BP-VPB-190A-GW-678-680	3452	7.26	8948	10.01	5551	13.92
02 VPB190A-HYD-20241217	4132	7.26	10947	10.01	6411	13.93

IS1 (DCB) = 1,4-Dichlorobenzene-d4

IS2 (NPT) = Naphthalene-d8

IS3 (ANT) = Acenaphthene-d10

AREA UPPER LIMIT = +100% of internal standard area

AREA LOWER LIMIT = -50% of internal standard area

RT UPPER LIMIT = +0.50 minutes of internal standard RT

RT UPPER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.

* Values outside of QC limits.