

5B

SEMIVOLATILE ORGANIC INSTRUMENT PERFORMANCE CHECK
DECAFLUOROTRIPHENYLPHOSPHINE (DFTPP)

Lab Name: CHEMTECH

Contract: TETR06

Lab Code: CHEM

SAS No.: P5314

SDG NO.: P5314

Lab File ID: BN035349.D

DFTPP Injection Date: 11/27/2024

Instrument ID: BNA_N

DFTPP Injection Time: 14:55

m/e	ION ABUNDANCE CRITERIA	% RELATIVE ABUNDANCE
51	10.0 - 80.0% of mass 198	21.6
68	Less than 2.0% of mass 69	0.0 (0.0) 1
69	Mass 69 relative abundance	28.9
70	Less than 2.0% of mass 69	0.1 (0.4) 1
127	10.0 - 80.0% of mass 198	39.4
197	Less than 2.0% of mass 198	0.0
198	Base Peak, 100% relative abundance	100
199	5.0 to 9.0% of mass 198	6.7
275	10.0 - 60.0% of mass 198	27.9
365	Greater than 1% of mass 198	4.6
441	Present, but less than mass 443	11.5
442	Greater than 50% of mass 198	100
443	15.0 - 24.0% of mass 442	14 (19.7) 2

1-Value is % mass 69

2-Value is % mass 442

THIS CHECK APPLIES TO THE FOLLOWING SAMPLES, MS, MSD, BLANKS, AND STANDARDS:

EPA SAMPLE NO.	LAB SAMPLE ID	LAB FILE ID	DATE ANALYZED	TIME ANALYZED
SSTDICC0.1	SSTDICC0.1	BN035350.D	11/27/2024	15:34
SSTDICC0.2	SSTDICC0.2	BN035351.D	11/27/2024	16:10
SSTDICCC0.4	SSTDICCC0.4	BN035352.D	11/27/2024	16:46
SSTDICC0.8	SSTDICC0.8	BN035353.D	11/27/2024	17:21
SSTDICC1.6	SSTDICC1.6	BN035354.D	11/27/2024	17:57
SSTDICC3.2	SSTDICC3.2	BN035355.D	11/27/2024	18:33
SSTDICC5.0	SSTDICC5.0	BN035356.D	11/27/2024	19:09