

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH Contract: TETR06
Lab Code: CHEM Case No.: P5314 SAS No.: P5314 SDG NO.: P5314
Lab File ID: VX044438.D Date Analyzed: 12/20/2024
Instrument ID: MSVOA_X Time Analyzed: 09:16
GC Column: DB-624UI ID: 0.18 (mm) Heated Purge: (Y/N) N

	IS1 AREA #	RT #	IS2 AREA #	RT #	IS3 AREA #	RT #
12 HOUR STD	136445	5.54	226018	6.76	197543	10.06
UPPER LIMIT	272890	6.044	452036	7.257	395086	10.555
LOWER LIMIT	68222.5	5.044	113009	6.257	98771.5	9.555
EPA SAMPLE NO.						
BP-VPB-190A-TB-20241213	117644	5.55	208582	6.76	178072	10.06
BP-VPB-190A-GW-678-680	116922	5.55	205157	6.76	179312	10.06
BP-VPB-190A-GW-678-680DL	113918	5.54	202874	6.76	177737	10.06
BP-VPB-190A-GW-698-700	121735	5.55	211457	6.76	182367	10.06
VPB190A-HYD-20241217	120419	5.55	211996	6.76	183684	10.05
VX1220WBL01	120864	5.55	218417	6.76	193836	10.05
VX1220WBS01	146515	5.55	249173	6.76	212384	10.05
VX1220WBSD01	138719	5.54	231232	6.76	202087	10.05

IS1 = Pentafluorobenzene
IS2 = 1,4-Difluorobenzene
IS3 = Chlorobenzene-d5

AREA UPPER LIMIT = +100% of internal standard area
AREA LOWER LIMIT = -50% of internal standard area
RT UPPER LIMIT = +0.50 minutes of internal standard RT
RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.
* Values outside of QC limits.