

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH Contract: TETR06
Lab Code: CHEM Case No.: P5314 SAS No.: P5314 SDG NO.: P5314
Lab File ID: VX044438.D Date Analyzed: 12/20/2024
Instrument ID: MSVOA_X Time Analyzed: 09:16
GC Column: DB-624UI ID: 0.18 (mm) Heated Purge: (Y/N) N

	IS4 AREA #	RT #				
12 HOUR STD	95339	12.018				
UPPER LIMIT	190678	12.518				
LOWER LIMIT	47669.5	11.518				
EPA SAMPLE NO.						
BP-VPB-190A-TB-20241213	71905	12.02				
BP-VPB-190A-GW-678-680	74123	12.02				
BP-VPB-190A-GW-678-680DL	76153	12.02				
BP-VPB-190A-GW-698-700	69187	12.02				
VPB190A-HYD-20241217	74241	12.02				
VX1220WBL01	81005	12.02				
VX1220WBS01	96885	12.02				
VX1220WBSD01	96082	12.02				

IS4 = 1,4-Dichlorobenzene-d4

AREA UPPER LIMIT = +100% of internal standard area

AREA LOWER LIMIT = -50% of internal standard area

RT UPPER LIMIT = +0.50 minutes of internal standard RT

RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.

* Values outside of QC limits.