

Instrument ID: CV1

Daily Analysis Runlog For Sequence/QC Batch ID # LB134050

Review By	jaswal	Review On	12/20/2024 9:49:17 PM
Supervise By	mohan	Supervise On	12/20/2024 9:49:50 PM

STD. NAME	STD REF.#
ICAL Standard	MP83730,MP83731,MP83732,MP83733,MP83734,MP83735
ICV Standard	MP83736
CCV Standard	MP83738
ICSA Standard	
CRI Standard	MP83740
LCS Standard	
Chk Standard	MP83737,MP83739,MP83741,MP83743

Sr#	SampleId	ClientID	QcType	Date	Comment	Operator	Status
1	S0	S0	CAL1	12/20/24 13:58		Mohan	OK
2	S0.2	S0.2	CAL2	12/20/24 14:00		Mohan	OK
3	S2.5	S2.5	CAL3	12/20/24 14:03		Mohan	OK
4	S5	S5	CAL4	12/20/24 14:05		Mohan	OK
5	S7.5	S7.5	CAL5	12/20/24 14:07		Mohan	OK
6	S10	S10	CAL6	12/20/24 14:10		Mohan	OK
7	ICV28	ICV28	ICV	12/20/24 14:13		Mohan	OK
8	ICB28	ICB28	ICB	12/20/24 14:15		Mohan	OK
9	CCV57	CCV57	CCV	12/20/24 14:17		Mohan	OK
10	CCB57	CCB57	CCB	12/20/24 14:20		Mohan	OK
11	CRA	CRA	CRDL	12/20/24 14:22		Mohan	OK
12	HighStd	HighStd	HIGH STD	12/20/24 14:24		Mohan	OK
13	ChkStd	ChkStd	SAM	12/20/24 14:26		Mohan	OK
14	PB165798BL	PB165798BL	MB	12/20/24 14:29		Mohan	OK
15	PB165798BS	PB165798BS	LCS	12/20/24 14:31		Mohan	OK
16	P5316-01	TT-304-IDWSO-2024	SAM	12/20/24 14:34		Mohan	OK
17	P5339-01	TR-06-12182024	SAM	12/20/24 14:36		Mohan	OK
18	P5355-01	RBR251688	SAM	12/20/24 14:38		Mohan	OK

Instrument ID: CV1

Daily Analysis Runlog For Sequence/QC Batch ID # LB134050

Review By	jaswal	Review On	12/20/2024 9:49:17 PM
Supervise By	mohan	Supervise On	12/20/2024 9:49:50 PM
STD. NAME	STD REF.#		
ICAL Standard	MP83730,MP83731,MP83732,MP83733,MP83734,MP83735		
ICV Standard	MP83736		
CCV Standard	MP83738		
ICSA Standard			
CRI Standard	MP83740		
LCS Standard			
Chk Standard	MP83737,MP83739,MP83741,MP83743		

19	P5362-01	WC-SOIL-20241219	SAM	12/20/24 14:40		Mohan	OK
20	CCV58	CCV58	CCV	12/20/24 14:43		Mohan	OK
21	CCB58	CCB58	CCB	12/20/24 14:45		Mohan	OK
22	P5365-01	TAPFTA-SB01I-4.5-12	SAM	12/20/24 14:47		Mohan	OK
23	P5365-01DUP	TAPFTA-SB01I-4.5-12	DUP	12/20/24 14:50		Mohan	OK
24	P5365-01MS	TAPFTA-SB01I-4.5-12	MS	12/20/24 14:52		Mohan	OK
25	P5365-01MSD	TAPFTA-SB01I-4.5-12	MSD	12/20/24 14:54		Mohan	OK
26	P5365-01L	TAPFTA-SB01I-4.5-12	SD	12/20/24 14:56		Mohan	OK
27	P5365-01A	TAPFTA-SB01I-4.5-12	PS	12/20/24 14:59		Mohan	OK
28	CCV59	CCV59	CCV	12/20/24 15:01		Mohan	OK
29	CCB59	CCB59	CCB	12/20/24 15:03		Mohan	OK

Instrument ID: P4

Daily Analysis Runlog For Sequence/QC Batch ID # LB134054

Review By	jaswal	Review On	12/25/2024 1:47:05 AM
Supervise By	kareem	Supervise On	12/27/2024 10:17:53 AM

STD. NAME	STD REF.#
ICAL Standard	MP83552,MP83553,MP83554,MP83555,MP83556,MP83558
ICV Standard	MP83559
CCV Standard	MP83562
ICSA Standard	MP83560,MP83561
CRI Standard	MP83558
LCS Standard	
Chk Standard	MP83513,MP83514

Sr#	SampleId	ClientID	QcType	Date	Comment	Operator	Status
1	S0	S0	CAL1	12/20/24 15:41		Kareem	OK
2	S1	S1	CAL2	12/20/24 15:46		Kareem	OK
3	S2	S2	CAL3	12/20/24 15:50		Kareem	OK
4	S3	S3	CAL4	12/20/24 15:54		Kareem	OK
5	S4	S4	CAL5	12/20/24 15:58		Kareem	OK
6	S5	S5	CAL6	12/20/24 16:03		Kareem	OK
7	ICV01	ICV01	ICV	12/20/24 16:07	ICV fail for Ba,Mg,V (200.7)	Kareem	OK
8	LLICV01	LLICV01	LLICV	12/20/24 16:11		Kareem	OK
9	ICB01	ICB01	ICB	12/20/24 16:21		Kareem	OK
10	CRI01	CRI01	CRDL	12/20/24 16:26		Kareem	OK
11	ICSA01	ICSA01	ICSA	12/20/24 16:30		Kareem	OK
12	ICSAB01	ICSAB01	ICSAB	12/20/24 16:34		Kareem	OK
13	ICSADL	ICSADL	ICSA	12/20/24 16:38		Kareem	OK
14	ICSABDL	ICSABDL	ICSAB	12/20/24 16:43		Kareem	OK
15	CCV01	CCV01	CCV	12/20/24 16:47		Kareem	OK
16	CCB01	CCB01	CCB	12/20/24 16:51		Kareem	OK
17	P5312-01	SOIL-VNJ-222	SAM	12/20/24 16:55		Kareem	OK
18	P5343-01	VNJ210	SAM	12/20/24 17:00		Kareem	OK

Instrument ID: P4

Daily Analysis Runlog For Sequence/QC Batch ID # LB134054

Review By	jaswal	Review On	12/25/2024 1:47:05 AM
Supervise By	kareem	Supervise On	12/27/2024 10:17:53 AM

STD. NAME	STD REF.#
ICAL Standard	MP83552,MP83553,MP83554,MP83555,MP83556,MP83558
ICV Standard	MP83559
CCV Standard	MP83562
ICSA Standard	MP83560,MP83561
CRI Standard	MP83558
LCS Standard	
Chk Standard	MP83513,MP83514

19	P5316-01	TT-304-IDWSO-2024	SAM	12/20/24 17:04		Kareem	OK
20	P5343-03	VNJ281	SAM	12/20/24 17:08		Kareem	OK
21	P5312-03	CONCRETE-VNJ-222	SAM	12/20/24 17:12		Kareem	OK
22	P5279-02	ROCKAWAY-PARK	SAM	12/20/24 17:16		Kareem	OK
23	P5192-02	EFF-WASTE WATER	SAM	12/20/24 17:21		Kareem	OK
24	P5327-01	12175-12176-12177-1	SAM	12/20/24 17:25		Kareem	OK
25	CCV02	CCV02	CCV	12/20/24 17:29		Kareem	OK
26	CCB02	CCB02	CCB	12/20/24 17:34		Kareem	OK
27	PB165738BL	PB165738BL	MB	12/20/24 17:38		Kareem	OK
28	PB165738BS	PB165738BS	LCS	12/20/24 17:42	0.1 ML OF M6010 AND M6001 WERE ADDED TO 10ML OF THE SAMPLE	Kareem	OK
29	PB165754BL	PB165754BL	MB	12/20/24 17:46		Kareem	OK
30	PB165754BS	PB165754BS	LCS	12/20/24 17:51	0.1 ML OF M6010 AND M6001 WERE ADDED TO 10ML OF THE SAMPLE	Kareem	OK
31	PB165720BL	PB165720BL	MB	12/20/24 17:55		Kareem	OK
32	PB165720BS	PB165720BS	LCS	12/20/24 17:59	0.1 ML OF M6010 AND M6001 WERE ADDED TO 10ML OF THE SAMPLE	Kareem	OK
33	PB165679TB	PB165679TB	MB	12/20/24 18:03		Kareem	OK
34	PB165680TB	PB165680TB	MB	12/20/24 18:08		Kareem	OK
35	CCV03	CCV03	CCV	12/20/24 18:12		Kareem	OK

Instrument ID: P4

Daily Analysis Runlog For Sequence/QC Batch ID # LB134054

Review By	jaswal	Review On	12/25/2024 1:47:05 AM
Supervise By	kareem	Supervise On	12/27/2024 10:17:53 AM

STD. NAME	STD REF.#
ICAL Standard	MP83552,MP83553,MP83554,MP83555,MP83556,MP83558
ICV Standard	MP83559
CCV Standard	MP83562
ICSA Standard	MP83560,MP83561
CRI Standard	MP83558
LCS Standard	
Chk Standard	MP83513,MP83514

36	CCB03	CCB03	CCB	12/20/24 18:16		Kareem	OK
37	P5341-01	STORMWATER-AQ-C	SAM	12/20/24 18:21		Kareem	OK
38	P5341-03RE	TOTE-425RE	SAM	12/20/24 18:25	NOT USE	Kareem	Not Ok
39	P5343-06	PUMP-WATER-271	SAM	12/20/24 18:29		Kareem	OK
40	P5291-13RE	WC-20241213RE	SAM	12/20/24 18:33	NOT USE	Kareem	Not Ok
41	P5299-01	SB-01	SAM	12/20/24 18:38		Kareem	OK
42	P5327-01DUP	12175-12176-12177-1	DUP	12/20/24 18:42		Kareem	OK
43	P5327-01L	12175-12176-12177-1	SD	12/20/24 18:46		Kareem	OK
44	P5327-01MS	12175-12176-12177-1	MS	12/20/24 18:50	0.1 ML OF M6010 AND M6001 WERE ADDED TO 10ML OF THE SAMPLE	Kareem	OK
45	P5327-01MSD	12175-12176-12177-1	MSD	12/20/24 18:54	0.1 ML OF M6010 AND M6001 WERE ADDED TO 10ML OF THE SAMPLE	Kareem	OK
46	P5327-01A	12175-12176-12177-1	PS	12/20/24 18:58	0.1 ML OF M6010 AND M6001 WERE ADDED TO 10ML OF THE SAMPLE	Kareem	OK
47	CCV04	CCV04	CCV	12/20/24 19:02		Kareem	OK
48	CCB04	CCB04	CCB	12/20/24 19:11		Kareem	OK
49	P5343-06L	PUMP-WATER-271L	SD	12/20/24 19:16		Kareem	OK
50	P5343-06DUP	PUMP-WATER-271DU	DUP	12/20/24 19:20		Kareem	OK
51	P5343-06MS	PUMP-WATER-271MS	MS	12/20/24 19:24	0.1 ML OF M6010 AND M6001 WERE ADDED TO 10ML OF THE SAMPLE	Kareem	OK

Instrument ID: P4

Daily Analysis Runlog For Sequence/QC Batch ID # LB134054

Review By	jaswal	Review On	12/25/2024 1:47:05 AM
Supervise By	kareem	Supervise On	12/27/2024 10:17:53 AM

STD. NAME	STD REF.#
ICAL Standard	MP83552,MP83553,MP83554,MP83555,MP83556,MP83558
ICV Standard	MP83559
CCV Standard	MP83562
ICSA Standard	MP83560,MP83561
CRI Standard	MP83558
LCS Standard	
Chk Standard	MP83513,MP83514

52	P5343-06MSD	PUMP-WATER-271MSD	MSD	12/20/24 19:28	0.1 ML OF M6010 AND M6001 WERE ADDED TO 10ML OF THE SAMPLE	Kareem	OK
53	P5343-06A	PUMP-WATER-271A	PS	12/20/24 19:32	0.1 ML OF M6010 AND M6001 WERE ADDED TO 10ML OF THE SAMPLE	Kareem	OK
54	P5291-13DUP	WC-20241213DUP	DUP	12/20/24 19:37		Kareem	OK
55	P5291-13L	WC-20241213L	SD	12/20/24 19:41		Kareem	OK
56	P5291-13MS	WC-20241213MS	MS	12/20/24 19:45	0.1 ML OF M6010 AND M6001 WERE ADDED TO 10ML OF THE SAMPLE	Kareem	OK
57	P5291-13MSD	WC-20241213MSD	MSD	12/20/24 19:49	0.1 ML OF M6010 AND M6001 WERE ADDED TO 10ML OF THE SAMPLE	Kareem	OK
58	P5291-13A	WC-20241213A	PS	12/20/24 19:53	0.1 ML OF M6010 AND M6001 WERE ADDED TO 10ML OF THE SAMPLE	Kareem	OK
59	CCV05	CCV05	CCV	12/20/24 19:57		Kareem	OK
60	CCB05	CCB05	CCB	12/20/24 20:01		Kareem	OK
61	P5299-02	SB-02	SAM	12/20/24 20:06		Kareem	OK
62	P5299-03	SB-01	SAM	12/20/24 20:10		Kareem	OK
63	P5299-04	SB-02	SAM	12/20/24 20:14		Kareem	OK
64	P5299-04DUP	SB-02DUP	DUP	12/20/24 20:18		Kareem	OK
65	P5299-04L	SB-02L	SD	12/20/24 20:22		Kareem	OK
66	P5299-04MS	SB-02MS	MS	12/20/24 20:26	0.1 ML OF M6010 AND M6001 WERE ADDED TO 10ML OF THE SAMPLE	Kareem	OK

Instrument ID: P4

Daily Analysis Runlog For Sequence/QC Batch ID # LB134054

Review By	jaswal	Review On	12/25/2024 1:47:05 AM
Supervise By	kareem	Supervise On	12/27/2024 10:17:53 AM

STD. NAME	STD REF.#
ICAL Standard	MP83552,MP83553,MP83554,MP83555,MP83556,MP83558
ICV Standard	MP83559
CCV Standard	MP83562
ICSA Standard	MP83560,MP83561
CRI Standard	MP83558
LCS Standard	
Chk Standard	MP83513,MP83514

67	P5299-04MSD	SB-02MSD	MSD	12/20/24 20:30	0.1 ML OF M6010 AND M6001 WERE ADDED TO 10ML OF THE SAMPLE	Kareem	OK
68	P5299-04A	SB-02A	PS	12/20/24 20:34	0.1 ML OF M6010 AND M6001 WERE ADDED TO 10ML OF THE SAMPLE	Kareem	OK
69	PB165722BL	PB165722BL	MB	12/20/24 20:38		Kareem	OK
70	PB165722BS	PB165722BS	LCS	12/20/24 20:43	0.1 ML OF M6010 AND M6001 WERE ADDED TO 10ML OF THE SAMPLE	Kareem	OK
71	CCV06	CCV06	CCV	12/20/24 20:47		Kareem	OK
72	CCB06	CCB06	CCB	12/20/24 20:51		Kareem	OK
73	P5330-01	TP-5	SAM	12/20/24 20:55		Kareem	OK
74	P5342-01	CHRT26634	SAM	12/20/24 20:59		Kareem	OK
75	P5342-03	HT2651	SAM	12/20/24 21:04		Kareem	OK
76	P5342-05	RB21198	SAM	12/20/24 21:08		Kareem	OK
77	P5343-03DUP	VNJ281DUP	DUP	12/20/24 21:12		Kareem	OK
78	P5343-03L	VNJ281L	SD	12/20/24 21:16		Kareem	OK
79	P5343-03MSD	VNJ281MSD	MSD	12/20/24 21:24	0.1 ML OF M6010 AND M6001 WERE ADDED TO 10ML OF THE SAMPLE	Kareem	OK
80	P5343-03A	VNJ281A	PS	12/20/24 21:28	0.1 ML OF M6010 AND M6001 WERE ADDED TO 10ML OF THE SAMPLE	Kareem	OK
81	PB165753BL	PB165753BL	MB	12/20/24 21:32		Kareem	OK
82	CCV07	CCV07	CCV	12/20/24 21:37		Kareem	OK

Instrument ID: P4

Daily Analysis Runlog For Sequence/QC Batch ID # LB134054

Review By	jaswal	Review On	12/25/2024 1:47:05 AM
Supervise By	kareem	Supervise On	12/27/2024 10:17:53 AM

STD. NAME	STD REF.#
ICAL Standard	MP83552,MP83553,MP83554,MP83555,MP83556,MP83558
ICV Standard	MP83559
CCV Standard	MP83562
ICSA Standard	MP83560,MP83561
CRI Standard	MP83558
LCS Standard	
Chk Standard	MP83513,MP83514

83	CCB07	CCB07	CCB	12/20/24 22:04		Kareem	OK
84	PB165753BS	PB165753BS	LCS	12/20/24 22:08	0.1 ML OF M6010 AND M6001 WERE ADDED TO 10ML OF THE SAMPLE	Kareem	OK
85	P5337-01	ETGI-357	SAM	12/20/24 22:12		Kareem	OK
86	P5339-01	TR-06-12182024	SAM	12/20/24 22:17		Kareem	OK
87	P5355-01	RBR251688	SAM	12/20/24 22:21	MS/MSD fail for more than parameters	Kareem	Not Ok
88	P5355-01DUP	RBR251688DUP	DUP	12/20/24 22:25	MS/MSD fail for more than parameters	Kareem	Not Ok
89	P5355-01L	RBR251688L	SD	12/20/24 22:29	MS/MSD fail for more than parameters	Kareem	Not Ok
90	P5355-01MS	RBR251688MS	MS	12/20/24 22:33	MS/MSD fail for more than parameters	Kareem	Not Ok
91	P5355-01MSD	RBR251688MSD	MSD	12/20/24 22:38	MS/MSD fail for more than parameters	Kareem	Not Ok
92	P5355-01A	RBR251688A	PS	12/20/24 22:42	MS/MSD fail for more than parameters	Kareem	Not Ok
93	PB165786BL	PB165786BL	MB	12/20/24 22:46		Kareem	OK
94	CCV08	CCV08	CCV	12/20/24 23:01		Kareem	OK
95	CCB08	CCB08	CCB	12/20/24 23:12		Kareem	OK
96	PB165786BS	PB165786BS	LCS	12/20/24 23:17	0.1 ML OF M6010 AND M6001 WERE ADDED TO 10ML OF THE SAMPLE	Kareem	OK
97	P5341-03	TOTE-425	SAM	12/20/24 23:21		Kareem	OK
98	P5291-13	WC-20241213	SAM	12/20/24 23:25		Kareem	OK

Instrument ID: P4

Daily Analysis Runlog For Sequence/QC Batch ID # LB134054

Review By	jaswal	Review On	12/25/2024 1:47:05 AM
Supervise By	kareem	Supervise On	12/27/2024 10:17:53 AM
STD. NAME	STD REF.#		
ICAL Standard	MP83552,MP83553,MP83554,MP83555,MP83556,MP83558		
ICV Standard	MP83559		
CCV Standard	MP83562		
ICSA Standard	MP83560,MP83561		
CRI Standard	MP83558		
LCS Standard			
Chk Standard	MP83513,MP83514		

99	LR1	LR1	HIGH STD	12/20/24 23:29		Kareem	OK
100	LR2	LR2	HIGH STD	12/20/24 23:41		Kareem	OK
101	P5343-03MS	VNJ281MS	MS	12/20/24 23:46	0.1 ML OF M6010 AND M6001 WERE ADDED TO 10ML OF THE SAMPLE	Kareem	OK
102	CCV09	CCV09	CCV	12/20/24 23:55		Kareem	OK
103	CCB09	CCB09	CCB	12/21/24 00:05		Kareem	OK

Instrument ID: P4

Daily Analysis Runlog For Sequence/QC Batch ID # LB134066

Review By	mohan	Review On	12/26/2024 10:44:26 PM
Supervise By	kareem	Supervise On	12/27/2024 11:00:25 AM

STD. NAME	STD REF.#
ICAL Standard	MP83552,MP83553,MP83554,MP83555,MP83556,MP83558
ICV Standard	MP83559
CCV Standard	MP83562
ICSA Standard	MP83560,MP83561
CRI Standard	MP83558
LCS Standard	
Chk Standard	MP83565,MP83566

Sr#	SampleId	ClientID	QcType	Date	Comment	Operator	Status
1	S0	S0	CAL1	12/23/24 10:51		Kareem	OK
2	S1	S1	CAL2	12/23/24 10:56		Kareem	OK
3	S2	S2	CAL3	12/23/24 11:00		Kareem	OK
4	S3	S3	CAL4	12/23/24 11:04		Kareem	OK
5	S4	S4	CAL5	12/23/24 11:09		Kareem	OK
6	S5	S5	CAL6	12/23/24 11:13		Kareem	OK
7	ICV01	ICV01	ICV	12/23/24 11:17		Kareem	OK
8	LLICV01	LLICV01	LLICV	12/23/24 11:37		Kareem	OK
9	ICB01	ICB01	ICB	12/23/24 11:44		Kareem	OK
10	CRI01	CRI01	CRDL	12/23/24 11:51		Kareem	OK
11	ICSA01	ICSA01	ICSA	12/23/24 11:55		Kareem	OK
12	ICSAB01	ICSAB01	ICSAB	12/23/24 12:06		Kareem	OK
13	ICSADL	ICSADL	ICSA	12/23/24 12:10		Kareem	OK
14	ICSABDL	ICSABDL	ICSAB	12/23/24 12:14		Kareem	OK
15	CCV01	CCV01	CCV	12/23/24 12:31		Kareem	OK
16	CCB01	CCB01	CCB	12/23/24 12:35		Kareem	OK
17	PB165789BL	PB165789BL	MB	12/23/24 12:51		Kareem	OK
18	PB165789BS	PB165789BS	LCS	12/23/24 12:55	0.1ML OF M6001 AND M6010 WERE ADDED TO 10ML OF SAMPLE	Kareem	OK

Instrument ID: P4

Daily Analysis Runlog For Sequence/QC Batch ID # LB134066

Review By	mohan	Review On	12/26/2024 10:44:26 PM
Supervise By	kareem	Supervise On	12/27/2024 11:00:25 AM

STD. NAME	STD REF.#
ICAL Standard	MP83552,MP83553,MP83554,MP83555,MP83556,MP83558
ICV Standard	MP83559
CCV Standard	MP83562
ICSA Standard	MP83560,MP83561
CRI Standard	MP83558
LCS Standard	
Chk Standard	MP83565,MP83566

19	PB165659BL	PB165659BL	MB	12/23/24 12:59		Kareem	OK
20	PB165659BS	PB165659BS	LCS	12/23/24 13:03	0.1ML OF M6001 AND M6010 WERE ADDED TO 10ML OF SAMPLE	Kareem	OK
21	P5330-04	TP-5	SAM	12/23/24 13:08		Kareem	OK
22	P5337-03	ETGI-357	SAM	12/23/24 13:12		Kareem	OK
23	P5338-01	USED-OIL-DEBRIS	SAM	12/23/24 13:16		Kareem	OK
24	P5338-02	USED-MOBILE-TRAN	SAM	12/23/24 13:21		Kareem	OK
25	P5341-02	STORMWATER-SOLI	SAM	12/23/24 13:25		Kareem	OK
26	P5355-02	RBR251688	SAM	12/23/24 13:39		Kareem	OK
27	CCV02	CCV02	CCV	12/23/24 13:43		Kareem	OK
28	CCB02	CCB02	CCB	12/23/24 13:47		Kareem	OK
29	P5355-01	RBR251688	SAM	12/23/24 13:52		Kareem	OK
30	P5355-01DL	RBR251688DL	SAM	12/23/24 13:58	NOT USE	Kareem	Not Ok
31	CCV03	CCV03	CCV	12/23/24 14:07		Kareem	OK
32	CCB03	CCB03	CCB	12/23/24 14:11		Kareem	OK
33	P5343-05	OILY-RAGS-274	SAM	12/23/24 14:15		Kareem	OK
34	P5355-02DUP	RBR251688DUP	DUP	12/23/24 14:20		Kareem	OK
35	P5355-02L	RBR251688L	SD	12/23/24 14:24		Kareem	OK
36	P5355-02MS	RBR251688MS	MS	12/23/24 14:28	0.1ML OF M6001 AND M6010 WERE ADDED TO 10ML OF SAMPLE	Kareem	OK

Instrument ID: P4

Daily Analysis Runlog For Sequence/QC Batch ID # LB134066

Review By	mohan	Review On	12/26/2024 10:44:26 PM
Supervise By	kareem	Supervise On	12/27/2024 11:00:25 AM

STD. NAME	STD REF.#
ICAL Standard	MP83552,MP83553,MP83554,MP83555,MP83556,MP83558
ICV Standard	MP83559
CCV Standard	MP83562
ICSA Standard	MP83560,MP83561
CRI Standard	MP83558
LCS Standard	
Chk Standard	MP83565,MP83566

37	P5355-02MSD	RBR251688MSD	MSD	12/23/24 14:33	0.1ML OF M6001 AND M6010 WERE ADDED TO 10ML OF SAMPLE	Kareem	OK
38	P5355-02A	RBR251688A	PS	12/23/24 14:37	0.1ML OF M6001 AND M6010 WERE ADDED TO 10ML OF SAMPLE	Kareem	OK
39	P5306-17	OU4-VSL-06R-12122	SAM	12/23/24 14:41		Kareem	OK
40	P5306-17DUP	OU4-VSL-06R-12122	DUP	12/23/24 14:45		Kareem	OK
41	P5306-17L	OU4-VSL-06R-12122	SD	12/23/24 14:49		Kareem	OK
42	P5306-17MS	OU4-VSL-06R-12122	MS	12/23/24 14:53	0.1ML OF M6001 AND M6010 WERE ADDED TO 10ML OF SAMPLE	Kareem	OK
43	CCV04	CCV04	CCV	12/23/24 15:06		Kareem	OK
44	CCB04	CCB04	CCB	12/23/24 15:10		Kareem	OK
45	P5306-17MSD	OU4-VSL-06R-12122	MSD	12/23/24 15:14	0.1ML OF M6001 AND M6010 WERE ADDED TO 10ML OF SAMPLE	Kareem	OK
46	P5306-17A	OU4-VSL-06R-12122	PS	12/23/24 15:18	0.1ML OF M6001 AND M6010 WERE ADDED TO 10ML OF SAMPLE	Kareem	OK
47	P5355-01DUP	RBR251688DUP	DUP	12/23/24 15:22		Kareem	OK
48	P5355-01L	RBR251688L	SD	12/23/24 15:27		Kareem	OK
49	P5355-01MS	RBR251688MS	MS	12/23/24 15:31	0.1ML OF M6001 AND M6010 WERE ADDED TO 10ML OF SAMPLE	Kareem	OK
50	P5355-01MSD	RBR251688MSD	MSD	12/23/24 15:36	0.1ML OF M6001 AND M6010 WERE ADDED TO 10ML OF SAMPLE	Kareem	OK

Instrument ID: P4

Daily Analysis Runlog For Sequence/QC Batch ID # LB134066

Review By	mohan	Review On	12/26/2024 10:44:26 PM
Supervise By	kareem	Supervise On	12/27/2024 11:00:25 AM

STD. NAME	STD REF.#
ICAL Standard	MP83552,MP83553,MP83554,MP83555,MP83556,MP83558
ICV Standard	MP83559
CCV Standard	MP83562
ICSA Standard	MP83560,MP83561
CRI Standard	MP83558
LCS Standard	
Chk Standard	MP83565,MP83566

51	P5355-01A	RBR251688A	PS	12/23/24 15:40	0.1ML OF M6001 AND M6010 WERE ADDED TO 10ML OF SAMPLE	Kareem	OK
52	P5355-01DUPDL	RBR251688DUPDL	DUP	12/23/24 15:49	NOT USE	Kareem	Not Ok
53	P5355-01LDL	RBR251688LDL	SD	12/23/24 15:53	NOT USE	Kareem	Not Ok
54	CCV05	CCV05	CCV	12/23/24 16:06		Kareem	OK
55	CCB05	CCB05	CCB	12/23/24 16:10		Kareem	OK
56	P5355-01MSDL	RBR251688MSDL	MS	12/23/24 16:15	NOT USE	Kareem	Not Ok
57	P5355-01MSDDL	RBR251688MSDDL	MSD	12/23/24 16:19	NOT USE	Kareem	Not Ok
58	P5355-01ADL	RBR251688ADL	PS	12/23/24 16:23	NOT USE	Kareem	Not Ok
59	LR1	LR1	HIGH STD	12/23/24 16:27		Kareem	OK
60	LR2	LR2	HIGH STD	12/23/24 16:32		Kareem	OK
61	PB165733TB	PB165733TB	MB	12/23/24 16:36		Kareem	OK
62	CCV06	CCV06	CCV	12/23/24 16:54		Kareem	OK
63	CCB06	CCB06	CCB	12/23/24 16:58		Kareem	OK