

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: Tetra Tech NUS, Inc. **SDG No.:** P5316
Contract: TETR06 **Lab Code:** CHEM **Case No.:** P5316 **SAS No.:** P5316
Initial Calibration Source: EPA
Continuing Calibration Source: PLASMA-PURE

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
ICV28	Mercury	4.01	4.0	100	90 - 110	CV	12/20/2024	14:13	LB134050

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: Tetra Tech NUS, Inc. **SDG No.:** P5316
Contract: TETR06 **Lab Code:** CHEM **Case No.:** P5316 **SAS No.:** P5316
Initial Calibration Source: EPA
Continuing Calibration Source: PLASMA-PURE

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
CCV57	Mercury	5.05	5.0	101	90 - 110	CV	12/20/2024	14:17	LB134050
CCV58	Mercury	4.92	5.0	98	90 - 110	CV	12/20/2024	14:43	LB134050
CCV59	Mercury	4.90	5.0	98	90 - 110	CV	12/20/2024	15:01	LB134050

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: Tetra Tech NUS, Inc. SDG No.: P5316
Contract: TETR06 Lab Code: CHEM Case No.: P5316 SAS No.: P5316
Initial Calibration Source: EPA
Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
ICV01	Arsenic	1020	1000	102	90 - 110	P	12/20/2024	16:07	LB134054
	Barium	471	520	91	90 - 110	P	12/20/2024	16:07	LB134054
	Beryllium	486	510	95	90 - 110	P	12/20/2024	16:07	LB134054
	Cadmium	502	510	98	90 - 110	P	12/20/2024	16:07	LB134054
	Chromium	534	520	103	90 - 110	P	12/20/2024	16:07	LB134054
	Copper	522	510	102	90 - 110	P	12/20/2024	16:07	LB134054
	Lead	1000	1000	100	90 - 110	P	12/20/2024	16:07	LB134054
	Manganese	494	520	95	90 - 110	P	12/20/2024	16:07	LB134054
	Nickel	510	530	96	90 - 110	P	12/20/2024	16:07	LB134054
	Selenium	1020	1000	102	90 - 110	P	12/20/2024	16:07	LB134054
	Silver	251	250	100	90 - 110	P	12/20/2024	16:07	LB134054
	Zinc	1010	1000	101	90 - 110	P	12/20/2024	16:07	LB134054

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: Tetra Tech NUS, Inc. SDG No.: P5316
Contract: TETR06 Lab Code: CHEM Case No.: P5316 SAS No.: P5316
Initial Calibration Source: EPA
Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
LLICV01	Arsenic	19.0	20.0	95	80 - 120	P	12/20/2024	16:11	LB134054
	Barium	87.9	100	88	80 - 120	P	12/20/2024	16:11	LB134054
	Beryllium	5.75	6.0	96	80 - 120	P	12/20/2024	16:11	LB134054
	Cadmium	5.80	6.0	97	80 - 120	P	12/20/2024	16:11	LB134054
	Chromium	9.98	10.0	100	80 - 120	P	12/20/2024	16:11	LB134054
	Copper	21.8	20.0	109	80 - 120	P	12/20/2024	16:11	LB134054
	Lead	12.1	12.0	101	80 - 120	P	12/20/2024	16:11	LB134054
	Manganese	19.3	20.0	97	80 - 120	P	12/20/2024	16:11	LB134054
	Nickel	38.9	40.0	97	80 - 120	P	12/20/2024	16:11	LB134054
	Selenium	19.7	20.0	98	80 - 120	P	12/20/2024	16:11	LB134054
	Silver	10.3	10.0	103	80 - 120	P	12/20/2024	16:11	LB134054
	Zinc	38.9	40.0	97	80 - 120	P	12/20/2024	16:11	LB134054

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: Tetra Tech NUS, Inc. **SDG No.:** P5316

Contract: TETR06 **Lab Code:** CHEM **Case No.:** P5316 **SAS No.:** P5316

Initial Calibration Source: EPA

Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
CCV01	Arsenic	4800	5000	96	90 - 110	P	12/20/2024	16:47	LB134054
	Barium	9130	10000	91	90 - 110	P	12/20/2024	16:47	LB134054
	Beryllium	237	250	95	90 - 110	P	12/20/2024	16:47	LB134054
	Cadmium	2420	2500	97	90 - 110	P	12/20/2024	16:47	LB134054
	Chromium	1000	1000	100	90 - 110	P	12/20/2024	16:47	LB134054
	Copper	1220	1250	97	90 - 110	P	12/20/2024	16:47	LB134054
	Lead	4830	5000	96	90 - 110	P	12/20/2024	16:47	LB134054
	Manganese	2410	2500	96	90 - 110	P	12/20/2024	16:47	LB134054
	Nickel	2410	2500	96	90 - 110	P	12/20/2024	16:47	LB134054
	Selenium	4790	5000	96	90 - 110	P	12/20/2024	16:47	LB134054
	Silver	1220	1250	98	90 - 110	P	12/20/2024	16:47	LB134054
	Zinc	2440	2500	98	90 - 110	P	12/20/2024	16:47	LB134054
CCV02	Arsenic	5010	5000	100	90 - 110	P	12/20/2024	17:29	LB134054
	Barium	9240	10000	92	90 - 110	P	12/20/2024	17:29	LB134054
	Beryllium	242	250	97	90 - 110	P	12/20/2024	17:29	LB134054
	Cadmium	2490	2500	100	90 - 110	P	12/20/2024	17:29	LB134054
	Chromium	1020	1000	102	90 - 110	P	12/20/2024	17:29	LB134054
	Copper	1260	1250	101	90 - 110	P	12/20/2024	17:29	LB134054
	Lead	4970	5000	99	90 - 110	P	12/20/2024	17:29	LB134054
	Manganese	2460	2500	99	90 - 110	P	12/20/2024	17:29	LB134054
	Nickel	2480	2500	99	90 - 110	P	12/20/2024	17:29	LB134054
	Selenium	5020	5000	100	90 - 110	P	12/20/2024	17:29	LB134054
	Silver	1250	1250	100	90 - 110	P	12/20/2024	17:29	LB134054
	Zinc	2500	2500	100	90 - 110	P	12/20/2024	17:29	LB134054
CCV03	Arsenic	4960	5000	99	90 - 110	P	12/20/2024	18:12	LB134054
	Barium	9430	10000	94	90 - 110	P	12/20/2024	18:12	LB134054
	Beryllium	244	250	98	90 - 110	P	12/20/2024	18:12	LB134054
	Cadmium	2480	2500	99	90 - 110	P	12/20/2024	18:12	LB134054
	Chromium	1020	1000	102	90 - 110	P	12/20/2024	18:12	LB134054
	Copper	1250	1250	100	90 - 110	P	12/20/2024	18:12	LB134054
	Lead	4950	5000	99	90 - 110	P	12/20/2024	18:12	LB134054
	Manganese	2480	2500	99	90 - 110	P	12/20/2024	18:12	LB134054
	Nickel	2480	2500	99	90 - 110	P	12/20/2024	18:12	LB134054
	Selenium	4960	5000	99	90 - 110	P	12/20/2024	18:12	LB134054

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: Tetra Tech NUS, Inc. **SDG No.:** P5316

Contract: TETR06 **Lab Code:** CHEM **Case No.:** P5316 **SAS No.:** P5316

Initial Calibration Source: EPA

Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
CCV03	Silver	1250	1250	100	90 - 110	P	12/20/2024	18:12	LB134054
	Zinc	2500	2500	100	90 - 110	P	12/20/2024	18:12	LB134054
CCV04	Arsenic	4910	5000	98	90 - 110	P	12/20/2024	19:02	LB134054
	Barium	9150	10000	92	90 - 110	P	12/20/2024	19:02	LB134054
	Beryllium	240	250	96	90 - 110	P	12/20/2024	19:02	LB134054
	Cadmium	2440	2500	98	90 - 110	P	12/20/2024	19:02	LB134054
	Chromium	1010	1000	101	90 - 110	P	12/20/2024	19:02	LB134054
	Copper	1240	1250	99	90 - 110	P	12/20/2024	19:02	LB134054
	Lead	4890	5000	98	90 - 110	P	12/20/2024	19:02	LB134054
	Manganese	2450	2500	98	90 - 110	P	12/20/2024	19:02	LB134054
	Nickel	2440	2500	98	90 - 110	P	12/20/2024	19:02	LB134054
	Selenium	4920	5000	98	90 - 110	P	12/20/2024	19:02	LB134054
	Silver	1240	1250	99	90 - 110	P	12/20/2024	19:02	LB134054
CCV05	Zinc	2470	2500	99	90 - 110	P	12/20/2024	19:02	LB134054
	Arsenic	5060	5000	101	90 - 110	P	12/20/2024	19:57	LB134054
	Barium	9370	10000	94	90 - 110	P	12/20/2024	19:57	LB134054
	Beryllium	242	250	97	90 - 110	P	12/20/2024	19:57	LB134054
	Cadmium	2510	2500	100	90 - 110	P	12/20/2024	19:57	LB134054
	Chromium	1030	1000	103	90 - 110	P	12/20/2024	19:57	LB134054
	Copper	1270	1250	102	90 - 110	P	12/20/2024	19:57	LB134054
	Lead	5020	5000	100	90 - 110	P	12/20/2024	19:57	LB134054
	Manganese	2490	2500	99	90 - 110	P	12/20/2024	19:57	LB134054
	Nickel	2510	2500	100	90 - 110	P	12/20/2024	19:57	LB134054
	Selenium	5070	5000	102	90 - 110	P	12/20/2024	19:57	LB134054
CCV06	Silver	1260	1250	101	90 - 110	P	12/20/2024	19:57	LB134054
	Zinc	2520	2500	101	90 - 110	P	12/20/2024	19:57	LB134054
	Arsenic	4830	5000	96	90 - 110	P	12/20/2024	20:47	LB134054
	Barium	9020	10000	90	90 - 110	P	12/20/2024	20:47	LB134054
	Beryllium	237	250	95	90 - 110	P	12/20/2024	20:47	LB134054
	Cadmium	2410	2500	96	90 - 110	P	12/20/2024	20:47	LB134054
	Chromium	995	1000	100	90 - 110	P	12/20/2024	20:47	LB134054
	Copper	1220	1250	98	90 - 110	P	12/20/2024	20:47	LB134054
	Lead	4800	5000	96	90 - 110	P	12/20/2024	20:47	LB134054
	Manganese	2420	2500	97	90 - 110	P	12/20/2024	20:47	LB134054

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: Tetra Tech NUS, Inc. **SDG No.:** P5316

Contract: TETR06 **Lab Code:** CHEM **Case No.:** P5316 **SAS No.:** P5316

Initial Calibration Source: EPA

Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
CCV06	Nickel	2400	2500	96	90 - 110	P	12/20/2024	20:47	LB134054
	Selenium	4830	5000	97	90 - 110	P	12/20/2024	20:47	LB134054
	Silver	1210	1250	97	90 - 110	P	12/20/2024	20:47	LB134054
	Zinc	2420	2500	97	90 - 110	P	12/20/2024	20:47	LB134054
CCV07	Arsenic	4880	5000	98	90 - 110	P	12/20/2024	21:37	LB134054
	Barium	9110	10000	91	90 - 110	P	12/20/2024	21:37	LB134054
	Beryllium	242	250	97	90 - 110	P	12/20/2024	21:37	LB134054
	Cadmium	2430	2500	97	90 - 110	P	12/20/2024	21:37	LB134054
	Chromium	1010	1000	101	90 - 110	P	12/20/2024	21:37	LB134054
	Copper	1230	1250	99	90 - 110	P	12/20/2024	21:37	LB134054
	Lead	4860	5000	97	90 - 110	P	12/20/2024	21:37	LB134054
	Manganese	2460	2500	98	90 - 110	P	12/20/2024	21:37	LB134054
	Nickel	2430	2500	97	90 - 110	P	12/20/2024	21:37	LB134054
	Selenium	4890	5000	98	90 - 110	P	12/20/2024	21:37	LB134054
	Silver	1230	1250	99	90 - 110	P	12/20/2024	21:37	LB134054
	Zinc	2460	2500	98	90 - 110	P	12/20/2024	21:37	LB134054
	Arsenic	5020	5000	100	90 - 110	P	12/20/2024	23:01	LB134054
	Barium	9290	10000	93	90 - 110	P	12/20/2024	23:01	LB134054
CCV08	Beryllium	238	250	95	90 - 110	P	12/20/2024	23:01	LB134054
	Cadmium	2480	2500	99	90 - 110	P	12/20/2024	23:01	LB134054
	Chromium	1020	1000	102	90 - 110	P	12/20/2024	23:01	LB134054
	Copper	1260	1250	101	90 - 110	P	12/20/2024	23:01	LB134054
	Lead	4960	5000	99	90 - 110	P	12/20/2024	23:01	LB134054
	Manganese	2450	2500	98	90 - 110	P	12/20/2024	23:01	LB134054
	Nickel	2480	2500	99	90 - 110	P	12/20/2024	23:01	LB134054
	Selenium	5050	5000	101	90 - 110	P	12/20/2024	23:01	LB134054
	Silver	1250	1250	100	90 - 110	P	12/20/2024	23:01	LB134054
	Zinc	2480	2500	99	90 - 110	P	12/20/2024	23:01	LB134054
	Arsenic	4940	5000	99	90 - 110	P	12/20/2024	23:55	LB134054
	Barium	9220	10000	92	90 - 110	P	12/20/2024	23:55	LB134054
	Beryllium	239	250	96	90 - 110	P	12/20/2024	23:55	LB134054
	Cadmium	2450	2500	98	90 - 110	P	12/20/2024	23:55	LB134054
CCV09	Chromium	1020	1000	102	90 - 110	P	12/20/2024	23:55	LB134054
	Copper	1250	1250	100	90 - 110	P	12/20/2024	23:55	LB134054

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: Tetra Tech NUS, Inc. **SDG No.:** P5316
Contract: TETR06 **Lab Code:** CHEM **Case No.:** P5316 **SAS No.:** P5316
Initial Calibration Source: EPA
Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
CCV09	Lead	4900	5000	98	90 - 110	P	12/20/2024	23:55	LB134054
	Manganese	2460	2500	98	90 - 110	P	12/20/2024	23:55	LB134054
	Nickel	2450	2500	98	90 - 110	P	12/20/2024	23:55	LB134054
	Selenium	4950	5000	99	90 - 110	P	12/20/2024	23:55	LB134054
	Silver	1240	1250	99	90 - 110	P	12/20/2024	23:55	LB134054
	Zinc	2460	2500	98	90 - 110	P	12/20/2024	23:55	LB134054

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: Tetra Tech NUS, Inc. SDG No.: P5316
Contract: TETR06 Lab Code: CHEM Case No.: P5316 SAS No.: P5316
Initial Calibration Source: EPA
Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
ICV01	Arsenic	1050	1000	105	90 - 110	P	12/23/2024	11:17	LB134066
	Barium	484	520	93	90 - 110	P	12/23/2024	11:17	LB134066
	Beryllium	472	510	93	90 - 110	P	12/23/2024	11:17	LB134066
	Cadmium	508	510	100	90 - 110	P	12/23/2024	11:17	LB134066
	Chromium	532	520	102	90 - 110	P	12/23/2024	11:17	LB134066
	Copper	530	510	104	90 - 110	P	12/23/2024	11:17	LB134066
	Lead	1020	1000	102	90 - 110	P	12/23/2024	11:17	LB134066
	Manganese	490	520	94	90 - 110	P	12/23/2024	11:17	LB134066
	Nickel	519	530	98	90 - 110	P	12/23/2024	11:17	LB134066
	Selenium	1050	1000	105	90 - 110	P	12/23/2024	11:17	LB134066
	Silver	260	250	104	90 - 110	P	12/23/2024	11:17	LB134066
	Zinc	1030	1000	103	90 - 110	P	12/23/2024	11:17	LB134066

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: Tetra Tech NUS, Inc. SDG No.: P5316
Contract: TETR06 Lab Code: CHEM Case No.: P5316 SAS No.: P5316
Initial Calibration Source: EPA
Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
LLICV01	Arsenic	20.6	20.0	103	80 - 120	P	12/23/2024	11:37	LB134066
	Barium	90.4	100	90	80 - 120	P	12/23/2024	11:37	LB134066
	Beryllium	5.75	6.0	96	80 - 120	P	12/23/2024	11:37	LB134066
	Cadmium	5.89	6.0	98	80 - 120	P	12/23/2024	11:37	LB134066
	Chromium	9.87	10.0	99	80 - 120	P	12/23/2024	11:37	LB134066
	Copper	21.8	20.0	109	80 - 120	P	12/23/2024	11:37	LB134066
	Lead	12.0	12.0	100	80 - 120	P	12/23/2024	11:37	LB134066
	Manganese	19.5	20.0	98	80 - 120	P	12/23/2024	11:37	LB134066
	Nickel	39.3	40.0	98	80 - 120	P	12/23/2024	11:37	LB134066
	Selenium	19.0	20.0	95	80 - 120	P	12/23/2024	11:37	LB134066
	Silver	10.8	10.0	108	80 - 120	P	12/23/2024	11:37	LB134066
	Zinc	45.7	40.0	114	80 - 120	P	12/23/2024	11:37	LB134066

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: Tetra Tech NUS, Inc. **SDG No.:** P5316

Contract: TETR06 **Lab Code:** CHEM **Case No.:** P5316 **SAS No.:** P5316

Initial Calibration Source: EPA

Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
CCV01	Arsenic	5110	5000	102	90 - 110	P	12/23/2024	12:31	LB134066
	Barium	9710	10000	97	90 - 110	P	12/23/2024	12:31	LB134066
	Beryllium	232	250	93	90 - 110	P	12/23/2024	12:31	LB134066
	Cadmium	2450	2500	98	90 - 110	P	12/23/2024	12:31	LB134066
	Chromium	998	1000	100	90 - 110	P	12/23/2024	12:31	LB134066
	Copper	1280	1250	102	90 - 110	P	12/23/2024	12:31	LB134066
	Lead	4890	5000	98	90 - 110	P	12/23/2024	12:31	LB134066
	Manganese	2380	2500	95	90 - 110	P	12/23/2024	12:31	LB134066
	Nickel	2470	2500	99	90 - 110	P	12/23/2024	12:31	LB134066
	Selenium	5190	5000	104	90 - 110	P	12/23/2024	12:31	LB134066
	Silver	1250	1250	100	90 - 110	P	12/23/2024	12:31	LB134066
	Zinc	2450	2500	98	90 - 110	P	12/23/2024	12:31	LB134066
CCV02	Arsenic	5360	5000	107	90 - 110	P	12/23/2024	13:43	LB134066
	Barium	10100	10000	101	90 - 110	P	12/23/2024	13:43	LB134066
	Beryllium	230	250	92	90 - 110	P	12/23/2024	13:43	LB134066
	Cadmium	2500	2500	100	90 - 110	P	12/23/2024	13:43	LB134066
	Chromium	1000	1000	100	90 - 110	P	12/23/2024	13:43	LB134066
	Copper	1340	1250	107	90 - 110	P	12/23/2024	13:43	LB134066
	Lead	4990	5000	100	90 - 110	P	12/23/2024	13:43	LB134066
	Manganese	2410	2500	96	90 - 110	P	12/23/2024	13:43	LB134066
	Nickel	2530	2500	101	90 - 110	P	12/23/2024	13:43	LB134066
	Selenium	5480	5000	110	90 - 110	P	12/23/2024	13:43	LB134066
	Silver	1270	1250	102	90 - 110	P	12/23/2024	13:43	LB134066
	Zinc	2450	2500	98	90 - 110	P	12/23/2024	13:43	LB134066
CCV03	Arsenic	5340	5000	107	90 - 110	P	12/23/2024	14:07	LB134066
	Barium	10200	10000	102	90 - 110	P	12/23/2024	14:07	LB134066
	Beryllium	230	250	92	90 - 110	P	12/23/2024	14:07	LB134066
	Cadmium	2470	2500	99	90 - 110	P	12/23/2024	14:07	LB134066
	Chromium	1000	1000	100	90 - 110	P	12/23/2024	14:07	LB134066
	Copper	1340	1250	107	90 - 110	P	12/23/2024	14:07	LB134066
	Lead	4940	5000	99	90 - 110	P	12/23/2024	14:07	LB134066
	Manganese	2400	2500	96	90 - 110	P	12/23/2024	14:07	LB134066
	Nickel	2510	2500	100	90 - 110	P	12/23/2024	14:07	LB134066
	Selenium	5480	5000	110	90 - 110	P	12/23/2024	14:07	LB134066

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: Tetra Tech NUS, Inc. **SDG No.:** P5316

Contract: TETR06 **Lab Code:** CHEM **Case No.:** P5316 **SAS No.:** P5316

Initial Calibration Source: EPA

Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
CCV03	Silver	1270	1250	102	90 - 110	P	12/23/2024	14:07	LB134066
	Zinc	2450	2500	98	90 - 110	P	12/23/2024	14:07	LB134066
CCV04	Arsenic	5210	5000	104	90 - 110	P	12/23/2024	15:06	LB134066
	Barium	9720	10000	97	90 - 110	P	12/23/2024	15:06	LB134066
	Beryllium	226	250	90	90 - 110	P	12/23/2024	15:06	LB134066
	Cadmium	2450	2500	98	90 - 110	P	12/23/2024	15:06	LB134066
	Chromium	998	1000	100	90 - 110	P	12/23/2024	15:06	LB134066
	Copper	1300	1250	104	90 - 110	P	12/23/2024	15:06	LB134066
	Lead	4870	5000	97	90 - 110	P	12/23/2024	15:06	LB134066
	Manganese	2330	2500	93	90 - 110	P	12/23/2024	15:06	LB134066
	Nickel	2480	2500	99	90 - 110	P	12/23/2024	15:06	LB134066
	Selenium	5340	5000	107	90 - 110	P	12/23/2024	15:06	LB134066
	Silver	1250	1250	100	90 - 110	P	12/23/2024	15:06	LB134066
CCV05	Zinc	2350	2500	94	90 - 110	P	12/23/2024	15:06	LB134066
	Arsenic	5380	5000	108	90 - 110	P	12/23/2024	16:06	LB134066
	Barium	9910	10000	99	90 - 110	P	12/23/2024	16:06	LB134066
	Beryllium	227	250	91	90 - 110	P	12/23/2024	16:06	LB134066
	Cadmium	2470	2500	99	90 - 110	P	12/23/2024	16:06	LB134066
	Chromium	1010	1000	101	90 - 110	P	12/23/2024	16:06	LB134066
	Copper	1340	1250	108	90 - 110	P	12/23/2024	16:06	LB134066
	Lead	4920	5000	98	90 - 110	P	12/23/2024	16:06	LB134066
	Manganese	2360	2500	94	90 - 110	P	12/23/2024	16:06	LB134066
	Nickel	2510	2500	100	90 - 110	P	12/23/2024	16:06	LB134066
	Selenium	5500	5000	110	90 - 110	P	12/23/2024	16:06	LB134066
CCV06	Silver	1270	1250	102	90 - 110	P	12/23/2024	16:06	LB134066
	Zinc	2350	2500	94	90 - 110	P	12/23/2024	16:06	LB134066
	Arsenic	5350	5000	107	90 - 110	P	12/23/2024	16:54	LB134066
	Barium	10100	10000	101	90 - 110	P	12/23/2024	16:54	LB134066
	Beryllium	243	250	97	90 - 110	P	12/23/2024	16:54	LB134066
	Cadmium	2540	2500	102	90 - 110	P	12/23/2024	16:54	LB134066
	Chromium	1040	1000	104	90 - 110	P	12/23/2024	16:54	LB134066
	Copper	1350	1250	108	90 - 110	P	12/23/2024	16:54	LB134066
	Lead	5030	5000	100	90 - 110	P	12/23/2024	16:54	LB134066
	Manganese	2430	2500	97	90 - 110	P	12/23/2024	16:54	LB134066

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: Tetra Tech NUS, Inc. SDG No.: P5316

Contract: TETR06 Lab Code: CHEM Case No.: P5316 SAS No.: P5316

Initial Calibration Source: EPA

Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
CCV06	Nickel	2560	2500	102	90 - 110	P	12/23/2024	16:54	LB134066
	Selenium	5460	5000	109	90 - 110	P	12/23/2024	16:54	LB134066
	Silver	1280	1250	103	90 - 110	P	12/23/2024	16:54	LB134066
	Zinc	2340	2500	94	90 - 110	P	12/23/2024	16:54	LB134066