



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Metals

- 2b -

CRDL STANDARD FOR AA & ICP

Client: Tetra Tech NUS, Inc. **SDG No.:** P5316
Contract: TETR06 **Lab Code:** CHEM **Case No.:** P5316 **SAS No.:** P5316
Initial Calibration Source: _____
Continuing Calibration Source: _____

Sample ID	Analyte	Result ug/L	True Value ug/L	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
CRA	Mercury	0.19	0.2	97	40 - 160	CV	12/20/2024	14:22	LB134050
CRI01	Arsenic	21.1	20.0	106	40 - 160	P	12/20/2024	16:26	LB134054
	Barium	87.8	100	88	40 - 160	P	12/20/2024	16:26	LB134054
	Beryllium	5.71	6.0	95	40 - 160	P	12/20/2024	16:26	LB134054
	Cadmium	5.88	6.0	98	40 - 160	P	12/20/2024	16:26	LB134054
	Chromium	10.2	10.0	102	40 - 160	P	12/20/2024	16:26	LB134054
	Copper	21.8	20.0	109	40 - 160	P	12/20/2024	16:26	LB134054
	Lead	11.1	12.0	93	40 - 160	P	12/20/2024	16:26	LB134054
	Manganese	18.9	20.0	95	40 - 160	P	12/20/2024	16:26	LB134054
	Nickel	39.0	40.0	98	40 - 160	P	12/20/2024	16:26	LB134054
	Selenium	18.3	20.0	92	40 - 160	P	12/20/2024	16:26	LB134054
	Silver	10.3	10.0	103	40 - 160	P	12/20/2024	16:26	LB134054
	Zinc	38.8	40.0	97	40 - 160	P	12/20/2024	16:26	LB134054
CRI01	Arsenic	19.2	20.0	96	40 - 160	P	12/23/2024	11:51	LB134066
	Barium	92.0	100	92	40 - 160	P	12/23/2024	11:51	LB134066
	Beryllium	5.71	6.0	95	40 - 160	P	12/23/2024	11:51	LB134066
	Cadmium	5.81	6.0	97	40 - 160	P	12/23/2024	11:51	LB134066
	Chromium	9.74	10.0	97	40 - 160	P	12/23/2024	11:51	LB134066
	Copper	21.9	20.0	110	40 - 160	P	12/23/2024	11:51	LB134066
	Lead	11.0	12.0	92	40 - 160	P	12/23/2024	11:51	LB134066
	Manganese	19.6	20.0	98	40 - 160	P	12/23/2024	11:51	LB134066
	Nickel	39.3	40.0	98	40 - 160	P	12/23/2024	11:51	LB134066
	Selenium	17.1	20.0	86	40 - 160	P	12/23/2024	11:51	LB134066
	Silver	11.0	10.0	110	40 - 160	P	12/23/2024	11:51	LB134066
	Zinc	41.4	40.0	104	40 - 160	P	12/23/2024	11:51	LB134066