

Metals

- 6 -

DUPLICATE SAMPLE SUMMARY

Client: Tetra Tech NUS, Inc. **Level:** LOW **SDG No.:** P5316
Contract: TETR06 **Lab Code:** CHEM **Case No.:** P5316 **SAS No.:** P5316
Matrix: Solid **Sample ID:** P5355-01 **Client ID:** RBR251688DUP
Percent Solids for Sample: 81.7 **Duplicate ID** P5355-01DUP **Percent Solids for Spike Sample:** 81.7

Analyte	Units	Acceptance Limit	Sample Result	C	Duplicate Result	C	RPD	Qual	M
Arsenic	mg/Kg	20	4.65		4.32		7		P
Barium	mg/Kg	20	332		308		8		P
Beryllium	mg/Kg	20	0.99		0.94		5		P
Cadmium	mg/Kg	20	3.59		3.46		4		P
Chromium	mg/Kg	20	19.6		18.6		5		P
Copper	mg/Kg	20	37.2		35.0		6		P
Lead	mg/Kg	20	95.0		89.7		6		P
Manganese	mg/Kg	20	2230		2080		7		P
Nickel	mg/Kg	20	52.2		49.5		5		P
Selenium	mg/Kg	20	1.08	U	1.01	U			P
Silver	mg/Kg	20	0.99		1.00		1		P
Zinc	mg/Kg	20	127		117		8		P

Metals

- 6 -

DUPLICATE SAMPLE SUMMARY

Client: Tetra Tech NUS, Inc. **Level:** LOW **SDG No.:** P5316
Contract: TETR06 **Lab Code:** CHEM **Case No.:** P5316 **SAS No.:** P5316
Matrix: Solid **Sample ID:** P5355-01MS **Client ID:** RBR251688MSD
Percent Solids for Sample: 81.7 **Duplicate ID** P5355-01MSD **Percent Solids for Spike Sample:** 81.7

Analyte	Units	Acceptance Limit	Sample Result	C	Duplicate Result	C	RPD	Qual	M
Arsenic	mg/Kg	20	40.3		37.5		7		P
Barium	mg/Kg	20	351		333		5		P
Beryllium	mg/Kg	20	7.89		7.46		6		P
Cadmium	mg/Kg	20	14.6		13.7		6		P
Chromium	mg/Kg	20	36.4		34.3		6		P
Copper	mg/Kg	20	51.2		47.9		7		P
Lead	mg/Kg	20	151		141		7		P
Manganese	mg/Kg	20	2320		2200		5		P
Nickel	mg/Kg	20	82.2		77.0		7		P
Selenium	mg/Kg	20	81.0		75.6		7		P
Silver	mg/Kg	20	4.31		4.06		6		P
Zinc	mg/Kg	20	137		128		7		P

Metals

- 6 -

DUPLICATE SAMPLE SUMMARY

Client: Tetra Tech NUS, Inc. **Level:** LOW **SDG No.:** P5316
Contract: TETR06 **Lab Code:** CHEM **Case No.:** P5316 **SAS No.:** P5316
Matrix: Solid **Sample ID:** P5365-01 **Client ID:** TAPFTA-SB01I-4.5-121924-00-T1DUP
Percent Solids for Sample: 90.1 **Duplicate ID** P5365-01DUP **Percent Solids for Spike Sample:** 90.1

Analyte	Units	Acceptance Limit	Sample Result	C	Duplicate Result	C	RPD	Qual	M
Mercury	mg/Kg	20	0.0080	J	0.0080	J	0		CV

Metals

- 6 -

DUPLICATE SAMPLE SUMMARY

Client: Tetra Tech NUS, Inc. **Level:** LOW **SDG No.:** P5316
Contract: TETR06 **Lab Code:** CHEM **Case No.:** P5316 **SAS No.:** P5316
Matrix: Solid **Sample ID:** P5365-01MS **Client ID:** TAPFTA-SB01I-4.5-121924-00-T1MSD
Percent Solids for Sample: 90.1 **Duplicate ID** P5365-01MSD **Percent Solids for Spike Sample:** 90.1

Analyte	Units	Acceptance Limit	Sample Result	C	Duplicate Result	C	RPD	Qual	M
Mercury	mg/Kg	20	0.29		0.30		3		CV