

Metals

- 7 -

LABORATORY CONTROL SAMPLE SUMMARY

Client: Tetra Tech NUS, Inc. **SDG No.:** P5316
Contract: TETR06 **Lab Code:** CHEM **Case No.:** P5316 **SAS No.:** P5316

Analyte	Units	True Value	Result	C	% Recovery	Acceptance Limits	M
PB165786BS							
Arsenic	mg/Kg	35.4	35.9		101	82 - 111	P
Barium	mg/Kg	8.8	7.76		88	83 - 113	P
Beryllium	mg/Kg	8.8	8.43		96	83 - 113	P
Cadmium	mg/Kg	8.8	8.61		98	82 - 113	P
Chromium	mg/Kg	17.7	18.2		103	85 - 113	P
Copper	mg/Kg	13.3	13.7		103	81 - 117	P
Lead	mg/Kg	44.2	43.1		98	81 - 112	P
Manganese	mg/Kg	8.8	8.70		99	84 - 114	P
Nickel	mg/Kg	22.1	21.9		99	83 - 113	P
Selenium	mg/Kg	88.5	88.4		100	78 - 111	P
Silver	mg/Kg	3.3	3.35		102	82 - 112	P
Zinc	mg/Kg	8.8	8.90		101	82 - 113	P

Metals

- 7 -

LABORATORY CONTROL SAMPLE SUMMARY

Client: Tetra Tech NUS, Inc.

SDG No.: P5316

Contract: TETR06

Lab Code: CHEM

Case No.: P5316

SAS No.: P5316

Analyte	Units	True Value	Result	C	% Recovery	Acceptance Limits	M
PB165798BS Mercury	mg/Kg	0.25	0.26		106	80 - 124	CV