

### CALIBRATION VERIFICATION SUMMARY

Contract: TETR06

Lab Code: CHEM Case No.: P5316 SAS No.: P5316 SDG NO.: P5316

Continuing Calib Date: 12/18/2024 Initial Calibration Date(s): 12/06/2024 12/06/2024

Continuing Calib Time: 14:46 Initial Calibration Time(s): 14:19 22:34

GC Column: ZB-MR1 ID: 0.32 (mm)

| COMPOUND             |     | CCAL<br>RT | AVG<br>RT | RT WINDOW<br>FROM   TO |      | DIFF<br>RT |
|----------------------|-----|------------|-----------|------------------------|------|------------|
| Aroclor-1016-1       | (1) | 4.80       | 4.81      | 4.71                   | 4.91 | 0.01       |
| Aroclor-1016-2       | (2) | 4.82       | 4.83      | 4.73                   | 4.93 | 0.01       |
| Aroclor-1016-3       | (3) | 4.88       | 4.89      | 4.79                   | 4.99 | 0.01       |
| Aroclor-1016-4       | (4) | 5.00       | 5.01      | 4.91                   | 5.11 | 0.01       |
| Aroclor-1016-5       | (5) | 5.26       | 5.27      | 5.17                   | 5.37 | 0.01       |
| Aroclor-1260-1       | (1) | 6.30       | 6.31      | 6.21                   | 6.41 | 0.01       |
| Aroclor-1260-2       | (2) | 6.49       | 6.50      | 6.40                   | 6.60 | 0.01       |
| Aroclor-1260-3       | (3) | 6.86       | 6.87      | 6.77                   | 6.97 | 0.01       |
| Aroclor-1260-4       | (4) | 7.12       | 7.13      | 7.03                   | 7.23 | 0.01       |
| Aroclor-1260-5       | (5) | 7.36       | 7.37      | 7.27                   | 7.47 | 0.01       |
| Tetrachloro-m-xylene |     | 3.71       | 3.71      | 3.61                   | 3.81 | 0.00       |
| Decachlorobiphenyl   |     | 8.78       | 8.79      | 8.69                   | 8.89 | 0.02       |

### CALIBRATION VERIFICATION SUMMARY

Contract: TETR06

Lab Code: CHEM Case No.: P5316 SAS No.: P5316 SDG NO.: P5316

Continuing Calib Date: 12/18/2024 Initial Calibration Date(s): 12/06/2024 12/06/2024

Continuing Calib Time: 14:46 Initial Calibration Time(s): 14:19 22:34

GC Column: ZB-MR2 ID: 0.32 (mm)

| COMPOUND             |     | CCAL<br>RT | AVG<br>RT | RT WINDOW |      | DIFF<br>RT |
|----------------------|-----|------------|-----------|-----------|------|------------|
|                      |     |            |           | FROM      | TO   |            |
| Aroclor-1016-1       | (1) | 4.79       | 4.80      | 4.70      | 4.90 | 0.01       |
| Aroclor-1016-2       | (2) | 4.81       | 4.82      | 4.72      | 4.92 | 0.01       |
| Aroclor-1016-3       | (3) | 4.99       | 4.99      | 4.89      | 5.09 | 0.00       |
| Aroclor-1016-4       | (4) | 5.03       | 5.04      | 4.94      | 5.14 | 0.01       |
| Aroclor-1016-5       | (5) | 5.24       | 5.25      | 5.15      | 5.35 | 0.01       |
| Aroclor-1260-1       | (1) | 6.28       | 6.29      | 6.19      | 6.39 | 0.01       |
| Aroclor-1260-2       | (2) | 6.46       | 6.47      | 6.37      | 6.57 | 0.01       |
| Aroclor-1260-3       | (3) | 6.62       | 6.63      | 6.53      | 6.73 | 0.01       |
| Aroclor-1260-4       | (4) | 7.09       | 7.10      | 7.00      | 7.20 | 0.01       |
| Aroclor-1260-5       | (5) | 7.33       | 7.34      | 7.24      | 7.44 | 0.01       |
| Tetrachloro-m-xylene |     | 3.70       | 3.71      | 3.61      | 3.81 | 0.01       |
| Decachlorobiphenyl   |     | 8.72       | 8.74      | 8.64      | 8.84 | 0.02       |