

CALIBRATION VERIFICATION SUMMARY

Contract: TETR06

Lab Code: CHEM Case No.: P5316 SAS No.: P5316 SDG NO.: P5316

GC Column: ZB-MR1 ID: 0.32 (mm) Initi. Calib. Date(s): 12/06/2024 12/06/2024

Client Sample No.: CCAL02 Date Analyzed: 12/18/2024

Lab Sample No.: AR1660CCC500 Data File : PO108635.D Time Analyzed: 20:04

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	4.803	4.709	4.909	498.640	500.000	-0.3
Aroclor-1016-2	4.823	4.729	4.929	505.570	500.000	1.1
Aroclor-1016-3	4.879	4.785	4.985	503.330	500.000	0.7
Aroclor-1016-4	5.000	4.907	5.107	503.810	500.000	0.8
Aroclor-1016-5	5.258	5.165	5.365	505.700	500.000	1.1
Aroclor-1260-1	6.301	6.210	6.410	488.870	500.000	-2.2
Aroclor-1260-2	6.490	6.398	6.598	489.350	500.000	-2.1
Aroclor-1260-3	6.859	6.769	6.969	496.430	500.000	-0.7
Aroclor-1260-4	7.119	7.029	7.229	491.890	500.000	-1.6
Aroclor-1260-5	7.360	7.270	7.470	495.310	500.000	-0.9
Decachlorobiphenyl	8.775	8.691	8.891	46.280	50.000	-7.4
Tetrachloro-m-xylene	3.706	3.610	3.810	50.760	50.000	1.5

CALIBRATION VERIFICATION SUMMARY

Contract: TETR06

Lab Code: CHEM Case No.: P5316 SAS No.: P5316 SDG NO.: P5316

GC Column: ZB-MR2 ID: 0.32 (mm) Initi. Calib. Date(s): 12/06/2024 12/06/2024

Client Sample No.: CCAL02 Date Analyzed: 12/18/2024

Lab Sample No.: AR1660CCC500 Data File : PO108635.D Time Analyzed: 20:04

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	4.791	4.698	4.898	552.620	500.000	10.5
Aroclor-1016-2	4.811	4.718	4.918	552.950	500.000	10.6
Aroclor-1016-3	4.986	4.894	5.094	541.900	500.000	8.4
Aroclor-1016-4	5.028	4.935	5.135	532.940	500.000	6.6
Aroclor-1016-5	5.242	5.150	5.350	555.680	500.000	11.1
Aroclor-1260-1	6.277	6.186	6.386	553.460	500.000	10.7
Aroclor-1260-2	6.463	6.373	6.573	552.480	500.000	10.5
Aroclor-1260-3	6.617	6.527	6.727	553.540	500.000	10.7
Aroclor-1260-4	7.089	7.000	7.200	574.670	500.000	14.9
Aroclor-1260-5	7.329	7.239	7.439	580.680	500.000	16.1
Decachlorobiphenyl	8.725	8.641	8.841	54.120	50.000	8.2
Tetrachloro-m-xylene	3.704	3.608	3.808	55.170	50.000	10.3