

CALIBRATION VERIFICATION SUMMARY

Contract: TETR06

Lab Code: CHEM Case No.: P5316 SAS No.: P5316 SDG NO.: P5316

GC Column: ZB-MR1 ID: 0.32 (mm) Initi. Calib. Date(s): 12/06/2024 12/06/2024

Client Sample No.: CCAL03 Date Analyzed: 12/19/2024

Lab Sample No.: AR1660CCC500 Data File : PO108649.D Time Analyzed: 01:03

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
Aroclor-1016-1	4.803	4.709	4.909	506.530	500.000	1.3
Aroclor-1016-2	4.822	4.729	4.929	505.100	500.000	1.0
Aroclor-1016-3	4.878	4.785	4.985	509.000	500.000	1.8
Aroclor-1016-4	4.999	4.907	5.107	508.070	500.000	1.6
Aroclor-1016-5	5.257	5.165	5.365	507.310	500.000	1.5
Aroclor-1260-1	6.300	6.210	6.410	493.950	500.000	-1.2
Aroclor-1260-2	6.489	6.398	6.598	493.160	500.000	-1.4
Aroclor-1260-3	6.858	6.769	6.969	500.940	500.000	0.2
Aroclor-1260-4	7.118	7.029	7.229	498.360	500.000	-0.3
Aroclor-1260-5	7.359	7.270	7.470	503.010	500.000	0.6
Decachlorobiphenyl	8.773	8.691	8.891	46.710	50.000	-6.6
Tetrachloro-m-xylene	3.706	3.610	3.810	51.090	50.000	2.2

CALIBRATION VERIFICATION SUMMARY

Contract: TETR06

Lab Code: CHEM Case No.: P5316 SAS No.: P5316 SDG NO.: P5316

GC Column: ZB-MR2 ID: 0.32 (mm) Initi. Calib. Date(s): 12/06/2024 12/06/2024

Client Sample No.: CCAL03 Date Analyzed: 12/19/2024

Lab Sample No.: AR1660CCC500 Data File : PO108649.D Time Analyzed: 01:03

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	4.791	4.698	4.898	562.100	500.000	12.4
Aroclor-1016-2	4.810	4.718	4.918	562.390	500.000	12.5
Aroclor-1016-3	4.986	4.894	5.094	555.600	500.000	11.1
Aroclor-1016-4	5.028	4.935	5.135	547.020	500.000	9.4
Aroclor-1016-5	5.241	5.150	5.350	566.660	500.000	13.3
Aroclor-1260-1	6.276	6.186	6.386	562.810	500.000	12.6
Aroclor-1260-2	6.462	6.373	6.573	562.140	500.000	12.4
Aroclor-1260-3	6.617	6.527	6.727	563.250	500.000	12.7
Aroclor-1260-4	7.089	7.000	7.200	582.400	500.000	16.5
Aroclor-1260-5	7.328	7.239	7.439	587.540	500.000	17.5
Decachlorobiphenyl	8.725	8.641	8.841	55.060	50.000	10.1
Tetrachloro-m-xylene	3.704	3.608	3.808	55.980	50.000	12.0