

Matrix Spike/Matrix Spike Duplicate Summary

SW-846

SDG No.: P5316

Client: Tetra Tech NUS, Inc.

Analytical Method: 8082A

DataFile : PO108632.D

Lab Sample ID:	Parameter	Spike	Sample		Units	Rec	Rec		RPD		Limits	
			Result	Result			Qual	RPD	Qual	Low	High	RPD
Client Sample ID:	OU4-VSL-07-121224MS											
P5306-01MS	AR1016	183.3	0	239	ug/kg	130				47	134	
	AR1260	183.3	0	242	ug/kg	132				53	140	