

### Matrix Spike/Matrix Spike Duplicate Summary

SW-846

SDG No.: P5316

Client: Tetra Tech NUS, Inc.

Analytical Method: 8082A

DataFile : PO108633.D

| Lab Sample ID:    | Parameter            | Spike | Sample |        |       | Rec |      | RPD |      | Limits |      |     |
|-------------------|----------------------|-------|--------|--------|-------|-----|------|-----|------|--------|------|-----|
|                   |                      |       | Result | Result | Units | Rec | Qual | RPD | Qual | Low    | High | RPD |
| Client Sample ID: | OU4-VSL-07-121224MSD |       |        |        |       |     |      |     |      |        |      |     |
| P5306-01MSD       | AR1016               | 183.2 | 0      | 229    | ug/kg | 125 |      | 4   |      | 47     | 134  | 20  |
|                   | AR1260               | 183.2 | 0      | 220    | ug/kg | 120 |      | 10  |      | 53     | 140  | 20  |