

CALIBRATION VERIFICATION SUMMARY

Contract: TETR06

Lab Code: CHEM Case No.: P5316 SAS No.: P5316 SDG NO.: P5316

GC Column: ZB-MR1 ID: 0.32 (mm) Initi. Calib. Date(s): 11/25/2024 11/25/2024

Client Sample No.: CCAL01 Date Analyzed: 12/18/2024

Lab Sample No.: PSTDCCC050 Data File : PL093415.D Time Analyzed: 14:24

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
4,4'-DDD	6.712	6.608	6.808	54.320	50.000	8.6
4,4'-DDE	6.195	6.090	6.290	52.890	50.000	5.8
4,4'-DDT	7.026	6.922	7.122	54.440	50.000	8.9
Aldrin	5.259	5.155	5.355	52.760	50.000	5.5
alpha-BHC	3.997	3.892	4.092	54.060	50.000	8.1
alpha-Chlordane	6.021	5.916	6.116	52.540	50.000	5.1
beta-BHC	4.527	4.423	4.623	52.320	50.000	4.6
Decachlorobiphenyl	9.058	8.954	9.154	57.880	50.000	15.8
delta-BHC	4.774	4.670	4.870	52.050	50.000	4.1
Dieldrin	6.347	6.242	6.442	52.610	50.000	5.2
Endosulfan I	6.071	5.967	6.167	52.440	50.000	4.9
Endosulfan II	6.796	6.692	6.892	53.100	50.000	6.2
Endosulfan sulfate	7.161	7.056	7.256	53.010	50.000	6.0
Endrin	6.575	6.472	6.672	55.380	50.000	10.8
gamma-BHC (Lindane)	4.329	4.225	4.425	53.920	50.000	7.8
Heptachlor	4.918	4.813	5.013	52.780	50.000	5.6
Tetrachloro-m-xylene	3.541	3.436	3.636	53.140	50.000	6.3

CALIBRATION VERIFICATION SUMMARY

Contract: TETR06

Lab Code: CHEM Case No.: P5316 SAS No.: P5316 SDG NO.: P5316

GC Column: ZB-MR2 ID: 0.32 (mm) Initi. Calib. Date(s): 11/25/2024 11/25/2024

Client Sample No.: CCAL01 Date Analyzed: 12/18/2024

Lab Sample No.: PSTDCCC050 Data File : PL093415.D Time Analyzed: 14:24

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
4,4'-DDD	5.787	5.685	5.885	56.990	50.000	14.0
4,4'-DDE	5.232	5.130	5.330	56.250	50.000	12.5
4,4'-DDT	6.037	5.935	6.135	57.720	50.000	15.4
Aldrin	4.227	4.125	4.325	56.670	50.000	13.3
alpha-BHC	3.279	3.176	3.376	56.990	50.000	14.0
alpha-Chlordane	5.043	4.941	5.141	56.400	50.000	12.8
beta-BHC	3.909	3.807	4.007	55.080	50.000	10.2
Decachlorobiphenyl	7.914	7.812	8.012	57.470	50.000	14.9
delta-BHC	4.138	4.035	4.235	55.650	50.000	11.3
Dieldrin	5.364	5.262	5.462	56.950	50.000	13.9
Endosulfan I	5.099	4.997	5.197	50.910	50.000	1.8
Endosulfan II	5.934	5.832	6.032	57.220	50.000	14.4
Endosulfan sulfate	6.337	6.234	6.434	56.230	50.000	12.5
Endrin	5.639	5.537	5.737	58.340	50.000	16.7
gamma-BHC (Lindane)	3.609	3.506	3.706	57.010	50.000	14.0
Heptachlor	3.947	3.845	4.045	56.740	50.000	13.5
Tetrachloro-m-xylene	2.776	2.674	2.874	55.470	50.000	10.9