

CALIBRATION VERIFICATION SUMMARY

Contract: TETR06

Lab Code: CHEM Case No.: P5316 SAS No.: P5316 SDG NO.: P5316

GC Column: ZB-MR1 ID: 0.32 (mm) Initi. Calib. Date(s): 11/25/2024 11/25/2024

Client Sample No.: CCAL02 Date Analyzed: 12/18/2024

Lab Sample No.: PSTDCCC050 Data File : PL093427.D Time Analyzed: 18:18

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
4,4'-DDD	6.711	6.608	6.808	49.090	50.000	-1.8
4,4'-DDE	6.194	6.090	6.290	47.160	50.000	-5.7
4,4'-DDT	7.025	6.922	7.122	48.250	50.000	-3.5
Aldrin	5.258	5.155	5.355	46.780	50.000	-6.4
alpha-BHC	3.995	3.892	4.092	47.510	50.000	-5.0
alpha-Chlordane	6.020	5.916	6.116	46.660	50.000	-6.7
beta-BHC	4.526	4.423	4.623	46.620	50.000	-6.8
Decachlorobiphenyl	9.057	8.954	9.154	52.070	50.000	4.1
delta-BHC	4.773	4.670	4.870	45.820	50.000	-8.4
Dieldrin	6.346	6.242	6.442	46.760	50.000	-6.5
Endosulfan I	6.071	5.967	6.167	46.550	50.000	-6.9
Endosulfan II	6.795	6.692	6.892	47.450	50.000	-5.1
Endosulfan sulfate	7.159	7.056	7.256	47.290	50.000	-5.4
Endrin	6.574	6.472	6.672	49.110	50.000	-1.8
gamma-BHC (Lindane)	4.328	4.225	4.425	47.560	50.000	-4.9
Heptachlor	4.916	4.813	5.013	46.670	50.000	-6.7
Tetrachloro-m-xylene	3.540	3.436	3.636	46.880	50.000	-6.2

CALIBRATION VERIFICATION SUMMARY

Contract: TETR06

Lab Code: CHEM Case No.: P5316 SAS No.: P5316 SDG NO.: P5316

GC Column: ZB-MR2 ID: 0.32 (mm) Initi. Calib. Date(s): 11/25/2024 11/25/2024

Client Sample No.: CCAL02 Date Analyzed: 12/18/2024

Lab Sample No.: PSTDCCC050 Data File : PL093427.D Time Analyzed: 18:18

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
4,4'-DDD	5.788	5.685	5.885	51.970	50.000	3.9
4,4'-DDE	5.232	5.130	5.330	50.760	50.000	1.5
4,4'-DDT	6.038	5.935	6.135	51.110	50.000	2.2
Aldrin	4.227	4.125	4.325	50.780	50.000	1.6
alpha-BHC	3.278	3.176	3.376	50.980	50.000	2.0
alpha-Chlordane	5.044	4.941	5.141	50.880	50.000	1.8
beta-BHC	3.908	3.807	4.007	49.650	50.000	-0.7
Decachlorobiphenyl	7.913	7.812	8.012	53.120	50.000	6.2
delta-BHC	4.137	4.035	4.235	49.850	50.000	-0.3
Dieldrin	5.364	5.262	5.462	51.040	50.000	2.1
Endosulfan I	5.099	4.997	5.197	47.490	50.000	-5.0
Endosulfan II	5.935	5.832	6.032	51.340	50.000	2.7
Endosulfan sulfate	6.337	6.234	6.434	51.090	50.000	2.2
Endrin	5.639	5.537	5.737	52.190	50.000	4.4
gamma-BHC (Lindane)	3.608	3.506	3.706	51.000	50.000	2.0
Heptachlor	3.947	3.845	4.045	50.770	50.000	1.5
Tetrachloro-m-xylene	2.776	2.674	2.874	49.910	50.000	-0.2