

PESTICIDE CALIBRATION VERIFICATION SUMMARY

Contract: TETR06

Lab Code: CHEM Case No.: P5316 SAS No.: P5316 SDG NO.: P5316

GC Column: ZB-MR2 ID: 0.32 (mm) Initi. Calib. Date(s): 11/25/2024 11/25/2024

Client Sample No. (PEM): PEM - PL093231.D Date Analyzed: 11/25/2024

Lab Sample No.(PEM): PEM Time Analyzed: 11:05

PEM COMPOUND	RT	RT WINDOW		CALC	NOM	%D
		FROM	TO	AMOUNT(ng)	AMOUNT(ng)	
Decachlorobiphenyl	9.054	8.950	9.150	20.040	20.000	0.2
Tetrachloro-m-xylene	3.536	3.490	3.590	20.160	20.000	0.8
alpha-BHC	3.991	3.940	4.040	10.350	10.000	3.5
beta-BHC	4.522	4.470	4.570	10.650	10.000	6.5
gamma-BHC (Lindane)	4.324	4.270	4.370	10.190	10.000	1.9
Endrin	6.572	6.500	6.640	45.410	50.000	-9.2
4,4'-DDT	7.022	6.950	7.090	89.610	100.000	-10.4
Methoxychlor	7.497	7.430	7.570	212.340	250.000	-15.1

GC Column: ZB-MR1 ID: 0.32 (mm) Initi. Calib. Date(s): 11/25/2024 11/25/2024

Client Sample No. (PEM): PEM - PL093231.D Date Analyzed: 11/25/2024

Lab Sample No.(PEM): PEM Time Analyzed: 11:05

PEM COMPOUND	RT	RT WINDOW		CALC	NOM	%D
		FROM	TO	AMOUNT(ng)	AMOUNT(ng)	
Decachlorobiphenyl	7.912	7.810	8.010	19.260	20.000	-3.7
Tetrachloro-m-xylene	2.774	2.720	2.820	19.400	20.000	-3.0
alpha-BHC	3.276	3.230	3.330	9.230	10.000	-7.7
beta-BHC	3.906	3.860	3.960	9.990	10.000	-0.1
gamma-BHC (Lindane)	3.607	3.560	3.660	8.790	10.000	-12.1
Endrin	5.637	5.570	5.710	46.140	50.000	-7.7
4,4'-DDT	6.036	5.970	6.110	100.270	100.000	0.3
Methoxychlor	6.610	6.540	6.680	224.790	250.000	-10.1