

Instrument ID: ECD_L

Daily Analysis Runlog For Sequence/QC Batch ID # PL112524

Review By	yogesh	Review On	11/26/2024 8:18:30 AM
Supervise By	Ankita	Supervise On	11/26/2024 4:22:56 PM
SubDirectory	PL112524	HP Acquire Method	HP Processing Method PL112524
STD. NAME	STD REF.#		
Tune/Reschk	PP23793,PP23517		
Initial Calibration Stds	PP23673,PP23674,PP23675,PP23676,PP23677,PP23678,PP23679,PP23680,PP23681,PP23682,PP23683		
CCC	PP23686,PP23690,PP23695		
Internal Standard/PEM			
ICV/I.BLK	PP23687,PP23693,PP23698		
Surrogate Standard			
MS/MSD Standard			
LCS Standard			

Sr#	SampleId	Data File Name	Date-Time	Operator	Status
1	HEXANE	PL093229.D	25 Nov 2024 10:39	AR\AJ	Ok
2	I.BLK	PL093230.D	25 Nov 2024 10:52	AR\AJ	Ok
3	PEM	PL093231.D	25 Nov 2024 11:05	AR\AJ	Ok,M
4	RESCHK	PL093232.D	25 Nov 2024 11:18	AR\AJ	Ok
5	PSTDICC100	PL093233.D	25 Nov 2024 11:32	AR\AJ	Ok
6	PSTDICC075	PL093234.D	25 Nov 2024 11:45	AR\AJ	Ok
7	PSTDICC050	PL093235.D	25 Nov 2024 11:58	AR\AJ	Ok
8	PSTDICC025	PL093236.D	25 Nov 2024 12:11	AR\AJ	Ok
9	PSTDICC005	PL093237.D	25 Nov 2024 12:25	AR\AJ	Ok,M
10	PCHLORICC1000	PL093238.D	25 Nov 2024 12:38	AR\AJ	Ok
11	PCHLORICC750	PL093239.D	25 Nov 2024 12:51	AR\AJ	Ok
12	PCHLORICC500	PL093240.D	25 Nov 2024 13:04	AR\AJ	Ok
13	PCHLORICC250	PL093241.D	25 Nov 2024 13:18	AR\AJ	Ok
14	PCHLORICC050	PL093242.D	25 Nov 2024 13:31	AR\AJ	Ok,M
15	PTOXICC1000	PL093243.D	25 Nov 2024 13:44	AR\AJ	Ok
16	PTOXICC750	PL093244.D	25 Nov 2024 13:57	AR\AJ	Ok
17	PTOXICC500	PL093245.D	25 Nov 2024 14:11	AR\AJ	Ok
18	PTOXICC250	PL093246.D	25 Nov 2024 14:24	AR\AJ	Ok
19	PTOXICC100	PL093247.D	25 Nov 2024 14:37	AR\AJ	Ok,M
20	PSTDICV050	PL093248.D	25 Nov 2024 14:50	AR\AJ	Ok
21	PCHLORICV500	PL093249.D	25 Nov 2024 15:03	AR\AJ	Ok,M

Instrument ID: ECD_L

Daily Analysis Runlog For Sequence/QC Batch ID # PL112524

Review By	yogesh	Review On	11/26/2024 8:18:30 AM
Supervise By	Ankita	Supervise On	11/26/2024 4:22:56 PM
SubDirectory	PL112524	HP Acquire Method	HP Processing Method PL112524
STD. NAME	STD REF.#		
Tune/Reschk	PP23793,PP23517		
Initial Calibration Stds	PP23673,PP23674,PP23675,PP23676,PP23677,PP23678,PP23679,PP23680,PP23681,PP23682,PP23683		
CCC	PP23686,PP23690,PP23695		
Internal Standard/PEM			
ICV/I.BLK	PP23687,PP23693,PP23698		
Surrogate Standard			
MS/MSD Standard			
LCS Standard			

22	PTOXICV500	PL093250.D	25 Nov 2024 15:16	AR\AJ	Ok
23	I.BLK	PL093251.D	25 Nov 2024 15:30	AR\AJ	Ok
24	PEM	PL093252.D	25 Nov 2024 15:43	AR\AJ	Ok
25	PSTDCCC050	PL093253.D	25 Nov 2024 15:56	AR\AJ	Ok
26	PB165184BS	PL093254.D	25 Nov 2024 16:22	AR\AJ	Ok,M
27	PB165202BL	PL093255.D	25 Nov 2024 16:37	AR\AJ	Ok
28	PB165202BS	PL093256.D	25 Nov 2024 16:50	AR\AJ	Ok,M
29	PB165202BSD	PL093257.D	25 Nov 2024 17:03	AR\AJ	Ok,M
30	P4947-01	PL093258.D	25 Nov 2024 17:16	AR\AJ	Ok
31	I.BLK	PL093259.D	25 Nov 2024 17:29	AR\AJ	Ok
32	PSTDCCC050	PL093260.D	25 Nov 2024 17:42	AR\AJ	Ok

M : Manual Integration

Instrument ID: ECD_L

Daily Analysis Runlog For Sequence/QC Batch ID # PL121824

Review By	Abdul	Review On	12/19/2024 9:44:00 AM
Supervise By	Ankita	Supervise On	12/19/2024 10:29:06 AM
SubDirectory	PL121824	HP Acquire Method	HP Processing Method pl112524 8081
STD. NAME	STD REF.#		
Tune/Reschk	PP23793,PP23517		
Initial Calibration Stds	PP23673,PP23674,PP23675,PP23676,PP23677,PP23678,PP23679,PP23680,PP23681,PP23682,PP23683		
CCC	PP23686,PP23690,PP23695		
Internal Standard/PEM			
ICV/I.BLK	PP23687,PP23693,PP23698		
Surrogate Standard			
MS/MSD Standard			
LCS Standard			

Sr#	SampleId	Data File Name	Date-Time	Operator	Status
1	HEXANE	PL093406.D	18 Dec 2024 10:22	AR\AJ	Ok
2	I.BLK	PL093407.D	18 Dec 2024 10:35	AR\AJ	Ok
3	PEM	PL093408.D	18 Dec 2024 10:48	AR\AJ	Ok,M
4	PSTDCCC050	PL093409.D	18 Dec 2024 11:02	AR\AJ	Ok,M
5	PP24080	PL093410.D	18 Dec 2024 11:37	AR\AJ	Ok,M
6	PP24081	PL093411.D	18 Dec 2024 12:28	AR\AJ	Ok,M
7	PP24088	PL093412.D	18 Dec 2024 13:03	AR\AJ	Ok,M
8	PP24091	PL093413.D	18 Dec 2024 13:17	AR\AJ	Ok
9	I.BLK	PL093414.D	18 Dec 2024 14:10	AR\AJ	Ok
10	PSTDCCC050	PL093415.D	18 Dec 2024 14:24	AR\AJ	Ok,M
11	PB165704BL	PL093416.D	18 Dec 2024 15:49	AR\AJ	Ok
12	PB165704BS	PL093417.D	18 Dec 2024 16:03	AR\AJ	Ok
13	P5312-01	PL093418.D	18 Dec 2024 16:18	AR\AJ	Ok,M
14	P5312-03	PL093419.D	18 Dec 2024 16:32	AR\AJ	Ok,M
15	P5316-01	PL093420.D	18 Dec 2024 16:45	AR\AJ	Ok,M
16	P5306-01	PL093421.D	18 Dec 2024 16:58	AR\AJ	Ok,M
17	P5306-01MS	PL093422.D	18 Dec 2024 17:11	AR\AJ	Ok,M
18	P5306-01MSD	PL093423.D	18 Dec 2024 17:24	AR\AJ	Ok,M
19	P5306-03	PL093424.D	18 Dec 2024 17:38	AR\AJ	Ok,M
20	P5306-05	PL093425.D	18 Dec 2024 17:51	AR\AJ	Ok,M
21	I.BLK	PL093426.D	18 Dec 2024 18:04	AR\AJ	Ok

Instrument ID: ECD_L

Daily Analysis Runlog For Sequence/QC Batch ID # PL121824

Review By	Abdul	Review On	12/19/2024 9:44:00 AM
Supervise By	Ankita	Supervise On	12/19/2024 10:29:06 AM
SubDirectory	PL121824	HP Acquire Method	HP Processing Method pl112524 8081
STD. NAME	STD REF.#		
Tune/Reschk	PP23793,PP23517		
Initial Calibration Stds	PP23673,PP23674,PP23675,PP23676,PP23677,PP23678,PP23679,PP23680,PP23681,PP23682,PP23683		
CCC	PP23686,PP23690,PP23695		
Internal Standard/PEM			
ICV/I.BLK	PP23687,PP23693,PP23698		
Surrogate Standard			
MS/MSD Standard			
LCS Standard			

22	PSTDCCC050	PL093427.D	18 Dec 2024 18:18	AR\AJ	Ok,M
----	------------	------------	-------------------	-------	------

M : Manual Integration