

### Matrix Spike/Matrix Spike Duplicate Summary

SW-846

SDG No.: P5316

Client: Tetra Tech NUS, Inc.

Analytical Method: 8081B

DataFile : PL093422.D

| Lab Sample ID:    | Parameter           | Spike | Sample<br>Result | Result | Units | Rec<br>Qual | RPD<br>Qual | Low | Limits<br>High | RPD |
|-------------------|---------------------|-------|------------------|--------|-------|-------------|-------------|-----|----------------|-----|
| Client Sample ID: | OU4-VSL-07-121224MS |       |                  |        |       |             |             |     |                |     |
| P5306-01MS        | alpha-BHC           | 18.31 | 0                | 18.3   | ug/kg | 100         |             | 45  | 137            |     |
|                   | beta-BHC            | 18.31 | 0                | 18.5   | ug/kg | 101         |             | 50  | 136            |     |
|                   | delta-BHC           | 18.31 | 0                | 17.3   | ug/kg | 94          |             | 47  | 139            |     |
|                   | gamma-BHC (Lindane) | 18.31 | 0                | 18.0   | ug/kg | 98          |             | 49  | 135            |     |
|                   | Heptachlor          | 18.31 | 0                | 19.1   | ug/kg | 104         |             | 47  | 136            |     |
|                   | Aldrin              | 18.31 | 0                | 18.1   | ug/kg | 99          |             | 45  | 136            |     |
|                   | Endosulfan I        | 18.31 | 0                | 19.0   | ug/kg | 104         |             | 53  | 132            |     |
|                   | Dieldrin            | 18.31 | 0                | 18.9   | ug/kg | 103         |             | 56  | 136            |     |
|                   | 4,4'-DDE            | 18.31 | 0                | 19.0   | ug/kg | 104         |             | 56  | 134            |     |
|                   | Endrin              | 18.31 | 0                | 19.8   | ug/kg | 108         |             | 57  | 140            |     |
|                   | Endosulfan II       | 18.31 | 0                | 19.2   | ug/kg | 105         |             | 53  | 134            |     |
|                   | 4,4'-DDD            | 18.31 | 0                | 19.2   | ug/kg | 105         |             | 56  | 139            |     |
|                   | Endosulfan sulfate  | 18.31 | 0                | 19.1   | ug/kg | 104         |             | 55  | 136            |     |
|                   | 4,4'-DDT            | 18.31 | 0                | 20.1   | ug/kg | 110         |             | 50  | 141            |     |
|                   | alpha-Chlordane     | 18.31 | 0                | 19.2   | ug/kg | 105         |             | 54  | 133            |     |