

Matrix Spike/Matrix Spike Duplicate Summary

SW-846

SDG No.: P5316

Client: Tetra Tech NUS, Inc.

Analytical Method: 8081B

DataFile : PL093423.D

Lab Sample ID:	Parameter	Spike	Sample Result	Result	Units	Rec Qual	RPD Qual	Low	Limits High	RPD
Client Sample ID:	OU4-VSL-07-121224MSD									
P5306-01MSD	alpha-BHC	18.34	0	18.5	ug/kg	101	1	45	137	20
	beta-BHC	18.34	0	18.6	ug/kg	101	0	50	136	20
	delta-BHC	18.34	0	17.4	ug/kg	95	1	47	139	20
	gamma-BHC (Lindane)	18.34	0	18.2	ug/kg	99	1	49	135	20
	Heptachlor	18.34	0	19.2	ug/kg	105	1	47	136	20
	Aldrin	18.34	0	18.2	ug/kg	99	0	45	136	20
	Endosulfan I	18.34	0	19.1	ug/kg	104	0	53	132	20
	Dieldrin	18.34	0	19.1	ug/kg	104	1	56	136	20
	4,4'-DDE	18.34	0	19.2	ug/kg	105	1	56	134	20
	Endrin	18.34	0	20.1	ug/kg	110	2	57	140	20
	Endosulfan II	18.34	0	19.3	ug/kg	105	0	53	134	20
	4,4'-DDD	18.34	0	19.4	ug/kg	106	1	56	139	20
	Endosulfan sulfate	18.34	0	19.2	ug/kg	105	1	55	136	20
	4,4'-DDT	18.34	0	20.2	ug/kg	110	0	50	141	20
	alpha-Chlordane	18.34	0	19.2	ug/kg	105	0	54	133	20