

Surrogate Summary

SW-846

SDG No.: P5316

Client: Tetra Tech NUS, Inc.

Analytical Method: 8270E

Lab Sample ID	Client ID	Parameter	Spike (PPM)	Result (PPM)	Recovery (%)	Qual	Limits (%)	
							Low	High
P5306-09MS	OU4-VSL-11-121224MS	2-Fluorophenol	150	131	88		35	115
		Phenol-d6	150	134	90		34	127
		Nitrobenzene-d5	100	89.0	89		37	122
		2-Fluorobiphenyl	100	84.0	84		44	115
		2,4,6-Tribromophenol	150	122	82		39	132
		Terphenyl-d14	100	91.0	91		54	127
P5306-09MSD	OU4-VSL-11-121224MSD	2-Fluorophenol	150	142	95		35	115
		Phenol-d6	150	140	94		34	127
		Nitrobenzene-d5	100	89.7	90		37	122
		2-Fluorobiphenyl	100	86.8	87		44	115
		2,4,6-Tribromophenol	150	134	90		39	132
		Terphenyl-d14	100	92.3	92		54	127
P5316-01	TT-304-IDWSO-20241217-1	2-Fluorophenol	150	92.6	62		35	115
		Phenol-d6	150	97.5	65		34	127
		Nitrobenzene-d5	100	59.0	59		37	122
		2-Fluorobiphenyl	100	56.2	56		44	115
		2,4,6-Tribromophenol	150	86.1	57		39	132
		Terphenyl-d14	100	48.0	48	*	54	127
PB165705BL	PB165705BL	2-Fluorophenol	150	150	100		35	115
		Phenol-d6	150	153	102		34	127
		Nitrobenzene-d5	100	107	107		37	122
		2-Fluorobiphenyl	100	103	103		44	115
		2,4,6-Tribromophenol	150	146	97		39	132
		Terphenyl-d14	100	110	110		54	127
PB165705BS	PB165705BS	2-Fluorophenol	150	143	96		35	115
		Phenol-d6	150	136	91		34	127
		Nitrobenzene-d5	100	91.6	92		37	122
		2-Fluorobiphenyl	100	105	105		44	115
		2,4,6-Tribromophenol	150	135	90		39	132
		Terphenyl-d14	100	121	121		54	127