

Matrix Spike/Matrix Spike Duplicate Summary

SW-846

SDG No.: P5316

Client: Tetra Tech NUS, Inc.

Analytical Method: SW8270E

Parameter	Spike	Sample Result	Result	Units	Rec	Rec Qual	RPD	RPD Qual	Low	Limits High	RPD
Lab Sample ID:	P5306-09MS	Client Sample ID:	OU4-VSL-11-121224MS					DataFile:	BF140948.D		
Phenol	1800	0	2000	ug/Kg	111				34	121	
2-Methylphenol	1800	0	1900	ug/Kg	106				32	122	
3+4-Methylphenols	1800	0	2000	ug/Kg	111				34	119	
Naphthalene	1800	0	1800	ug/Kg	100				35	123	
Acenaphthylene	1800	0	2000	ug/Kg	111				32	132	
Acenaphthene	1800	0	1900	ug/Kg	106				40	123	
Fluorene	1800	0	1800	ug/Kg	100				43	125	
Pentachlorophenol	3500	0	2800	ug/Kg	80				25	133	
Anthracene	1800	0	1900	ug/Kg	106				47	123	
Fluoranthene	1800	0	1700	ug/Kg	94				50	127	
Pyrene	1800	0	2000	ug/Kg	111				47	127	
Benzo(a)anthracene	1800	0	1900	ug/Kg	106				49	126	
Chrysene	1800	0	1800	ug/Kg	100				50	124	
Benzo(b)fluoranthene	1800	0	1700	ug/Kg	94				45	132	
Benzo(k)fluoranthene	1800	0	2000	ug/Kg	111				47	132	
Benzo(a)pyrene	1800	0	2000	ug/Kg	111				45	129	
Indeno(1,2,3-cd)pyrene	1800	0	2400	ug/Kg	133				45	133	
Dibenz(a,h)anthracene	1800	0	2300	ug/Kg	128				45	134	
Benzo(g,h,i)perylene	1800	0	2200	ug/Kg	122				43	134	