

VOLATILE CONTINUING CALIBRATION CHECK

Lab Name: CHEMTECH Contract: TETR06

Lab Code: CHEM Case No.: P5316 SAS No.: P5316 SDG No.: P5316

Instrument ID: MSVOA_Y Calibration Date/Time: 12/20/2024 19:35

Lab File ID: VY020682.D Init. Calib. Date(s): 12/17/2024 12/17/2024

Heated Purge: (Y/N) Y Init. Calib. Time(s): 10:01 14:51

GC Column: RXI-624 ID: 0.25 (mm)

COMPOUND	RRF	RRF050	MIN RRF	%D	MAX%D
Vinyl Chloride	0.431	0.262		-39.21	50
1,1-Dichloroethene	0.627	0.500		-20.25	50
Acetone	0.125	0.087		-30.4	50
Methyl tert-butyl Ether	1.618	1.685		4.14	50
Methylene Chloride	0.681	0.659		-3.23	50
trans-1,2-Dichloroethene	0.703	0.713		1.42	50
1,1-Dichloroethane	1.306	1.366	0.1	4.59	50
2-Butanone	0.204	0.202		-0.98	50
Carbon Tetrachloride	0.711	0.692		-2.67	50
cis-1,2-Dichloroethene	0.809	0.836		3.34	50
Chloroform	1.528	1.423		-6.87	50
1,1,1-Trichloroethane	1.200	1.204		0.33	50
Benzene	1.883	1.905		1.17	50
1,2-Dichloroethane	0.500	0.515		3	50
Trichloroethene	0.463	0.463		0	50
Toluene	1.170	1.198		2.39	50
Tetrachloroethene	0.470	0.443		-5.74	50
Chlorobenzene	1.398	1.409	0.3	0.79	50
Ethyl Benzene	2.579	2.542		-1.43	50
m/p-Xylenes	0.959	0.944		-1.56	50
o-Xylene	0.895	0.891		-0.45	50
n-propylbenzene	5.257	4.958		-5.69	50
tert-Butylbenzene	3.206	3.020		-5.8	50
1,2,4-Trimethylbenzene	3.473	3.335		-3.97	50
sec-Butylbenzene	4.662	4.327		-7.19	50
1,3-Dichlorobenzene	1.876	1.803		-3.89	50
1,4-Dichlorobenzene	1.858	1.784		-3.98	50
n-Butylbenzene	3.643	3.417		-6.2	50
1,2-Dichlorobenzene	1.626	1.586		-2.46	50
1,2-Dichloroethane-d4	0.548	0.472		-13.87	50
Dibromofluoromethane	0.352	0.311		-11.65	50
Toluene-d8	1.233	1.049		-14.92	50
4-Bromofluorobenzene	0.482	0.406		-15.77	50
1,4-Dioxane	2.668	2.853	0.05	6.93	50

All other compounds must meet a minimum RRF of 0.010.
RRF of 1,4-Dioxane = Value should be divide by 1000.