

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH Contract: TETR06
Lab Code: CHEM Case No.: P5316 SAS No.: P5316 SDG NO.: P5316
Lab File ID: VY020664.D Date Analyzed: 12/20/2024
Instrument ID: MSVOA_Y Time Analyzed: 09:09
GC Column: RXI-624 ID: 0.25 (mm) Heated Purge: (Y/N) Y

	IS1 AREA #	RT #	IS2 AREA #	RT #	IS3 AREA #	RT #
12 HOUR STD	162547	7.71	240673	8.62	212145	11.42
UPPER LIMIT	325094	8.213	481346	9.116	424290	11.92
LOWER LIMIT	81273.5	7.213	120337	8.116	106073	10.92
EPA SAMPLE NO.						
TT-304-IDWSO-20241217-4	21219 *	7.71	29944 *	8.62	18784 *	11.41
TT-304-IDWSO-20241217-4RE	164449	7.71	291183	8.62	241631	11.41
VY1220SBL01	159056	7.71	246351	8.62	208495	11.41
VY1220SBS01	157442	7.71	237485	8.62	207752	11.42
VY1220SBSD01	154597	7.71	238791	8.62	207851	11.41

IS1 = Pentafluorobenzene
IS2 = 1,4-Difluorobenzene
IS3 = Chlorobenzene-d5

AREA UPPER LIMIT = +100% of internal standard area
AREA LOWER LIMIT = -50% of internal standard area
RT UPPER LIMIT = +0.50 minutes of internal standard RT
RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.
* Values outside of QC limits.