

Instrument ID: MSVOA_Y

Daily Analysis Runlog For Sequence/QC Batch ID # VY121724

Review By	Maresh Dadoda	Review On	12/18/2024 8:38:34 AM
Supervise By	Semsettin Yesilyurt	Supervise On	12/18/2024 8:51:04 AM
SubDirectory	VY121724	HP Acquire Method	HP Processing Method 82y121724s.m

STD. NAME	STD REF.#
Tune/Reschk	VP132182
Initial Calibration Stds	VP132183,VP132184,VP132185,VP132186,VP132187,VP132188
CCC	
Internal Standard/PEM	VP131783
ICV/I.BLK	VP132189
Surrogate Standard	
MS/MSD Standard	
LCS Standard	

Sr#	Sampleld	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	BFB	BFB	VY020611.D	17 Dec 2024 09:00		SY/MD	Ok
2	VSTDICC005	VSTDICC005	VY020612.D	17 Dec 2024 09:38		SY/MD	Ok
3	VSTDICC010	VSTDICC010	VY020613.D	17 Dec 2024 10:01	Com #16,30 on LR	SY/MD	Ok,M
4	VSTDICC020	VSTDICC020	VY020614.D	17 Dec 2024 10:23	Com #11 on QR	SY/MD	Ok,M
5	VSTDICCC050	VSTDICCC050	VY020615.D	17 Dec 2024 10:51	Com #13 fail for method	SY/MD	Ok,M
6	VSTDICC100	VSTDICC100	VY020616.D	17 Dec 2024 11:14		SY/MD	Ok,M
7	VSTDICC150	VSTDICC150	VY020617.D	17 Dec 2024 11:37		SY/MD	Ok,M
8	VIBLK	VIBLK	VY020618.D	17 Dec 2024 12:24		SY/MD	Ok
9	VSTDICC005	VSTDICC005	VY020619.D	17 Dec 2024 14:51	%D fail for com #11	SY/MD	Ok,M
10	VSTDICV050	ICVVY121724	VY020620.D	17 Dec 2024 17:41	ICV fail for com #11,16	SY/MD	Ok,M

M : Manual Integration

Instrument ID: MSVOA_Y

Daily Analysis Runlog For Sequence/QC Batch ID # VY122024

Review By	Semsettin Yesilyurt	Review On	12/23/2024 7:25:27 AM
Supervise By	Maresh Dadoda	Supervise On	12/24/2024 12:28:45 AM
SubDirectory	VY122024	HP Acquire Method	MSVOA_Y HP Processing Method 82y12022024s.m
STD. NAME	STD REF.#		
Tune/Reschk Initial Calibration Stds	VP132264		
CCC	VP132265,VP132266		
Internal Standard/PEM	VP131783		
ICV/I.BLK			
Surrogate Standard			
MS/MSD Standard			
LCS Standard			

Sr#	Sampleld	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	BFB	BFB	VY020663.D	20 Dec 2024 08:37		SY/MD	Ok
2	VSTDCCC050	VSTDCCC050	VY020664.D	20 Dec 2024 09:09		SY/MD	Ok,M
3	VY1220SBL01	VY1220SBL01	VY020665.D	20 Dec 2024 11:24		SY/MD	Ok
4	VY1220SBS01	VY1220SBS01	VY020666.D	20 Dec 2024 13:12		SY/MD	Ok,M
5	VY1220SBSD01	VY1220SBSD01	VY020667.D	20 Dec 2024 13:34		SY/MD	Ok,M
6	P5279-03	ROCKAWAY-PARK	VY020668.D	20 Dec 2024 14:10	vial-B	SY/MD	Ok
7	P5316-01	TT-304-IDWSO-202412	VY020669.D	20 Dec 2024 14:34	vial-A Surrogate Fail	SY/MD	ReRun
8	P5316-02	TT-304-IDWSO-202412	VY020670.D	20 Dec 2024 14:57	vial-A Surrogate Fail	SY/MD	ReRun
9	P5316-03	TT-304-IDWSO-202412	VY020671.D	20 Dec 2024 15:21	vial-A Surrogate Fail	SY/MD	ReRun
10	P5316-04	TT-304-IDWSO-202412	VY020672.D	20 Dec 2024 15:44	vial-A Internal Standard and Surrogate Fail	SY/MD	ReRun
11	P5343-03	VNJ281	VY020673.D	20 Dec 2024 16:07	vial-A Not purged	SY/MD	Not Ok
12	P5343-01	VNJ210	VY020674.D	20 Dec 2024 16:31	vial-B Not purged	SY/MD	Not Ok
13	P5316-04RE	TT-304-IDWSO-202412	VY020675.D	20 Dec 2024 16:52	vial-B Surrogate Fail	SY/MD	Confirms
14	P5306-01	OU4-VSL-07-121224	VY020676.D	20 Dec 2024 17:16	vial-B NOT PURGE	SY/MD	Not Ok
15	P5306-03	OU4-VSL-08-121224	VY020677.D	20 Dec 2024 17:39	vial-B Surrogate Fail	SY/MD	Ok
16	P5306-05	OU4-VSL-09-121224	VY020678.D	20 Dec 2024 18:02	vial-B	SY/MD	Ok
17	P5306-13RE	OU4-VSL-13-121224RE	VY020679.D	20 Dec 2024 18:26	vial-B Surrogate Fail;Internal Standard Fail	SY/MD	Confirms

Instrument ID: MSVOA_Y

Daily Analysis Runlog For Sequence/QC Batch ID # VY122024

Review By	Semsettin Yesilyurt	Review On	12/23/2024 7:25:27 AM		
Supervise By	Mahesh Dadoda	Supervise On	12/24/2024 12:28:45 AM		
SubDirectory	VY122024	HP Acquire Method	MSVOA_Y	HP Processing Method	82y12022024s.m

STD. NAME	STD REF.#
Tune/Reschk Initial Calibration Stds	VP132264
CCC Internal Standard/PEM ICV/I.BLK Surrogate Standard MS/MSD Standard LCS Standard	VP132265,VP132266 VP131783

18	P5306-07	OU4-VSL-10-121224	VY020680.D	20 Dec 2024 18:49	vial-B	SY/MD	Ok
19	P5306-09RE	OU4-VSL-11-121224RE	VY020681.D	20 Dec 2024 19:13	vial-B Surrogate Fail;Internal Standard Fail	SY/MD	Confirms
20	VSTDCCC050	VSTDCCC050EC	VY020682.D	20 Dec 2024 19:35		SY/MD	Ok,M

M : Manual Integration

Instrument ID: MSVOA_Y

Daily Analysis Runlog For Sequence/QC Batch ID # VY122624

Review By	Mahesh Dadoda	Review On	12/27/2024 9:56:45 AM				
Supervise By	Semsettin Yesilyurt	Supervise On	12/27/2024 10:00:03 AM				
SubDirectory	VY122624	HP Acquire Method	MSVOA_Y	HP Processing Method	82Y122624S.M		
STD. NAME		STD REF.#					
Tune/Reschk		VP132321					
Initial Calibration Stds		VP132326,VP132327,VP132328,VP132329,VP132330,VP132331					
CCC		VP132333,VP132334					
Internal Standard/PEM							
ICV/I.BLK		VP132332					
Surrogate Standard							
MS/MSD Standard							
LCS Standard							

Sr#	SampleId	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	BFB	BFB	VY020705.D	26 Dec 2024 08:29		SY/MD	Ok
2	VSTDICC005	VSTDICC005	VY020706.D	26 Dec 2024 09:31	Fail for acroline	SY/MD	Ok,M
3	VSTDICC010	VSTDICC010	VY020707.D	26 Dec 2024 09:53	LR-11,50	SY/MD	Ok,M
4	VSTDICC020	VSTDICC020	VY020708.D	26 Dec 2024 10:16		SY/MD	Ok,M
5	VSTDICCC050	VSTDICCC050	VY020709.D	26 Dec 2024 10:39		SY/MD	Ok,M
6	VSTDICC150	VSTDICC150	VY020710.D	26 Dec 2024 11:01		SY/MD	Ok,M
7	VSTDICC100	VSTDICC100	VY020711.D	26 Dec 2024 11:51		SY/MD	Ok,M
8	VIBLK	VIBLK	VY020712.D	26 Dec 2024 12:13		SY/MD	Ok
9	VSTDICV050	ICVVY122624	VY020713.D	26 Dec 2024 12:51		SY/MD	Ok,M
10	VY1226SBL01	VY1226SBL01	VY020714.D	26 Dec 2024 13:56		SY/MD	Ok
11	VY1226SBS01	VY1226SBS01	VY020715.D	26 Dec 2024 14:32		SY/MD	Ok,M
12	VY1226SBSD01	VY1226SBSD01	VY020716.D	26 Dec 2024 14:54		SY/MD	Ok,M
13	P5383-01	OK-02-12232024	VY020717.D	26 Dec 2024 15:18	Not purged,wrong purge	SY/MD	Not Ok
14	P5316-01	TT-304-IDWSO-202412	VY020718.D	26 Dec 2024 15:41	vial B	SY/MD	Ok
15	P5383-01	OK-02-12232024	VY020719.D	26 Dec 2024 16:04	ISTD Fail.vial B	SY/MD	ReRun
16	P5316-02	TT-304-IDWSO-202412	VY020720.D	26 Dec 2024 16:28	vial B	SY/MD	Ok
17	P5383-01	OK-02-12232024	VY020721.D	26 Dec 2024 16:51	vial A	SY/MD	Ok
18	P5316-03	TT-304-IDWSO-202412	VY020722.D	26 Dec 2024 17:15	vial B	SY/MD	Ok

Instrument ID: MSVOA_Y

Daily Analysis Runlog For Sequence/QC Batch ID # VY122624

Review By	Mahesh Dadoda	Review On	12/27/2024 9:56:45 AM		
Supervise By	Semsettin Yesilyurt	Supervise On	12/27/2024 10:00:03 AM		
SubDirectory	VY122624	HP Acquire Method	MSVOA_Y	HP Processing Method	82Y122624S.M
STD. NAME		STD REF.#			
Tune/Reschk		VP132321			
Initial Calibration Stds		VP132326,VP132327,VP132328,VP132329,VP132330,VP132331			
CCC		VP132333,VP132334			
Internal Standard/PEM					
ICV/I.BLK		VP132332			
Surrogate Standard					
MS/MSD Standard					
LCS Standard					

19	P5342-01RE	CHRT26634RE	VY020723.D	26 Dec 2024 17:38	Internal Standard Fail ;Surrogate fail,vial B	SY/MD	Confirms
20	P5342-03RE	HT2651RE	VY020724.D	26 Dec 2024 18:01	Internal Standard Fail,surrogate fail,vial B	SY/MD	Confirms
21	P5342-05RE	RB21198RE	VY020725.D	26 Dec 2024 18:25	Internal Standard Fail,vial B	SY/MD	Confirms
22	P5318-01	AU-06-122024	VY020726.D	26 Dec 2024 18:48	Internal Standard Fail,vial A	SY/MD	ReRun
23	P5382-01	COMP-1	VY020727.D	26 Dec 2024 19:12	vial A	SY/MD	Ok
24	VSTDCCC050	VSTDCCC050EC	VY020728.D	26 Dec 2024 19:34		SY/MD	Ok,M

M : Manual Integration